


LB107316

INSTRUMENT ID: CV1

7471B

μ Abs.:

47381



Linear

A= 0.0000e+000

B= 2.1108e-004

C= -9.7175e-002

Rho= 0.9998521

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.055	-0.055	199	0.000	199				
0.2	0.200	0.170	-0.030	1267	0.0 %	1267				
2.5	2.500	2.554	0.054	12560	0.0 %	12560				
5.0	5.000	5.058	0.058	24423	0.0 %	24423				
7.5	7.500	7.569	0.069	36318	0.0 %	36318				
10.0	10.000	9.904	-0.096	47381	0.0 %	47381				

LB107316 INSTRUMENT ID :CV1

Method: 7471B		Operator: Admin		Date of Analysis: 23 Jan 2020 14:52:56							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type		
0.0 - 1	SD.0	199	-	0.0000	7471B	PPB	23 Jan 2020 14:54:36	S	Std		
0.2 - 1	SD.2	1267	-	0.2000	7471B	PPB	23 Jan 2020 14:56:53	S	Std		
2.5 - 1	SD.5	12560	-	2.5000	7471B	PPB	23 Jan 2020 14:59:10	S	Std		
5.0 - 1	SD.8	24423	-	5.0000	7471B	PPB	23 Jan 2020 15:01:27	S	Std		
7.5 - 1	SD.5	36318	-	7.5000	7471B	PPB	23 Jan 2020 15:03:43	S	Std		
10.0 - 1	SD.0	47381	-	10.0000	7471B	PPB	23 Jan 2020 15:06:00	S	Std		
ICV83 - 1	ICV83	18915	3.8954	-	7471B	PPB	23 Jan 2020 15:08:54	U	SMPL		
ICB83 - 1	ICB83	201	-0.0547	-	7471B	PPB	23 Jan 2020 15:11:09	U	SMPL		
CCV100 - 1 CCV01	CCV100 CCV01	24426	5.0587	-	7471B	PPB	23 Jan 2020 15:13:25	U	SMPL		
CCB100 - 1 CCB01	CCB100 CCB01	195	-0.0560	-	7471B	PPB	23 Jan 2020 15:15:40	U	SMPL		
CRA - 1	CRA	1438	0.2064	-	7471B	PPB	23 Jan 2020 15:21:39	U	SMPL		
HighStd - 1	HighStd	47501	9.9294	-	7471B	PPB	23 Jan 2020 15:23:53	U	SMPL		
ChkStd - 1	ChkStd	36862	7.6837	-	7471B	PPB	23 Jan 2020 15:26:08	U	SMPL		
PB126311BL - 1	PBS	143	-0.0670	-	7471B	PPB	23 Jan 2020 15:28:25	U	SMPL		
PB126311BS - 1	LCSS	19589	4.0377	-	7471B	PPB	23 Jan 2020 15:30:41	U	SMPL		
L1004-13 - 1	BU-03-012020	1726	0.2671	-	7471B	PPB	23 Jan 2020 15:32:57	U	SMPL		
L1004-13DUP - 1	BU-03-012020DUP	1656	0.2524	-	7471B	PPB	23 Jan 2020 15:35:14	U	SMPL		
L1004-13MS - 1	BU-03-012020MS	23757	4.9175	-	7471B	PPB	23 Jan 2020 15:37:31	U	SMPL		
L1004-13MSD - 1	BU-03-012020MSD	23737	4.9132	-	7471B	PPB	23 Jan 2020 15:39:48	U	SMPL		
L1210-06 - 1	Q012-WC1	3682	0.6800	-	7471B	PPB	23 Jan 2020 15:42:03	U	SMPL		
CCV101 - 1 CCV02	CCV101 CCV02	24087	4.9871	-	7471B	PPB	23 Jan 2020 15:44:18	U	SMPL		
CCB101 - 1 CCB02	CCB101 CCB02	145	-0.0666	-	7471B	PPB	23 Jan 2020 15:46:33	U	SMPL		
L1211-01 - 1	B1-0-15	1590	0.2384	-	7471B	PPB	23 Jan 2020 15:48:49	U	SMPL		
L1211-03 - 1	B1-15-30	1725	0.2669	-	7471B	PPB	23 Jan 2020 15:51:04	U	SMPL		
L1213-01 - 1	TP-1	5171	0.9943	-	7471B	PPB	23 Jan 2020 15:53:19	U	SMPL		
L1213-03 - 1	TP-2	4704	0.8957	-	7471B	PPB	23 Jan 2020 15:55:35	U	SMPL		
L1215-01 - 1	RBR-200055	4601	0.8740	-	7471B	PPB	23 Jan 2020 15:57:50	U	SMPL		
L1215-02 - 1	RBR-200031	5513	1.0665	-	7471B	PPB	23 Jan 2020 16:00:06	U	SMPL		
L1219-01 - 1	ST-CONCRETE	708	0.0523	-	7471B	PPB	23 Jan 2020 16:02:22	U	SMPL		
L1230-01 - 1	STOCK-PILE-A	1748	0.2718	-	7471B	PPB	23 Jan 2020 16:04:39	U	SMPL		
L1230-03 - 1	STOCK-PILE-B	1888	0.3013	-	7471B	PPB	23 Jan 2020 16:06:56	U	SMPL		
L1004-13LX5 - 1		317	-0.0303	-	7471B	PPB	23 Jan 2020 16:09:11	U	SMPL		
CCV102 - 1 CCV03	CCV102 CCV03	25949	5.3802	-	7471B	PPB	23 Jan 2020 16:15:55	U	SMPL		
CCB102 - 1 CCB03	CCB102 CCB03	20	-0.0930	-	7471B	PPB	23 Jan 2020 16:18:11	U	SMPL		

01/22/20