


LIB 107345

7470A

Instrument ID = CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.028	-0.028	221	0.000	221				
0.2	0.200	0.173	-0.027	993	0.0 %	993				
2.5	2.500	2.552	0.052	10163	0.0 %	10163				
5.0	5.000	5.030	0.030	19714	0.0 %	19714				
7.5	7.500	7.503	0.003	29247	0.0 %	29247				
10.0	10.000	9.969	-0.031	38751	0.0 %	38751				

LB107345 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Jan 2020 17:13:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD-0	221	-	0.00007470A	PPB		24 Jan 2020 17:33:31	S	Std
0.2 - 1	SD-2	993	-	0.20007470A	PPB		24 Jan 2020 17:35:38	S	Std
2.5 - 1	SD-5	10163	-	2.50007470A	PPB		24 Jan 2020 17:37:45	S	Std
5.0 - 1	SD-10	19714	-	5.00007470A	PPB		24 Jan 2020 17:39:53	S	Std
7.5 - 1	SD-15	29247	-	7.50007470A	PPB		24 Jan 2020 17:42:00	S	Std
10.0 - 1	SD-20	38751	-	10.00007470A	PPB		24 Jan 2020 17:44:06	S	Std
ICV88 - 1	ICV88	16230	4.1261	-7470A	PPB		24 Jan 2020 17:48:21	U	SMPL
ICB88 - 1	ICB88	146	-0.0470	-7470A	PPB		24 Jan 2020 17:50:26	U	SMPL
CCV14 - 1	CCV14	20101	5.1305	-7470A	PPB		24 Jan 2020 17:52:32	U	SMPL
CCB14 - 1	CCB14	156	-0.0444	-7470A	PPB		24 Jan 2020 17:54:37	U	SMPL
CRA - 1	CRA	1036	0.1839	-7470A	PPB		24 Jan 2020 18:00:22	U	SMPL
HighStd - 1	HighStd	39384	10.1336	-7470A	PPB		24 Jan 2020 18:02:27	U	SMPL
ChkStd - 1	ChkStd	28679	7.3561	-7470A	PPB		24 Jan 2020 18:04:33	U	SMPL
PB126350BL - 1	PBW	146	-0.0470	-7470A	PPB		24 Jan 2020 18:06:40	U	SMPL
PB126350BS - 1	LCSW	16337	4.1539	-7470A	PPB		24 Jan 2020 18:08:46	U	SMPL
L1007-10 - 1	JC-03-012320-A	136	-0.0496	-7470A	PPB		24 Jan 2020 18:10:53	U	SMPL
L1007-10DUP - 1	JC-03-012320-ADUP	132	-0.0506	-7470A	PPB		24 Jan 2020 18:13:01	U	SMPL
L1007-10MS - 1	JC-03-012320-AMS	15980	4.0612	-7470A	PPB		24 Jan 2020 18:15:08	U	SMPL
L1007-10MSD - 1	JC-03-012320-AMSD	16277	4.1383	-7470A	PPB		24 Jan 2020 18:17:16	U	SMPL
L1007-18 - 1	JC-03-012320-B	135	-0.0499	-7470A	PPB		24 Jan 2020 18:19:21	U	SMPL
CCV15 - 1	CCV15	19961	5.0941	-7470A	PPB		24 Jan 2020 18:21:27	U	SMPL
CCB15 - 1	CCB15	143	-0.0478	-7470A	PPB		24 Jan 2020 18:23:33	U	SMPL
L1007-26 - 1	JC-03-012320-C	119	-0.0540	-7470A	PPB		24 Jan 2020 18:25:38	U	SMPL
L1016-14 - 1	SU-04-012320	126	-0.0522	-7470A	PPB		24 Jan 2020 18:27:44	U	SMPL
L1234-03 - 1	NS-TP	138	-0.0491	-7470A	PPB		24 Jan 2020 18:29:50	U	SMPL
L1236-16 - 1	40218-S	120	-0.0538	-7470A	PPB		24 Jan 2020 18:31:56	U	SMPL
L1240-03 - 1	SB-05-COMP	130	-0.0512	-7470A	PPB		24 Jan 2020 18:34:02	U	SMPL
L1244-01 - 1	CTY-WC-S-04	137	-0.0493	-7470A	PPB		24 Jan 2020 18:36:09	U	SMPL
L1244-02 - 1	CTY-WC-S-BC	180	-0.0382	-7470A	PPB		24 Jan 2020 18:38:16	U	SMPL
L1247-06 - 1	TP-4-5-COMP	134	-0.0501	-7470A	PPB		24 Jan 2020 18:40:24	U	SMPL
L1247-14 - 1	TP-7-6-COMP	128	-0.0517	-7470A	PPB		24 Jan 2020 18:42:32	U	SMPL
L1247-22 - 1	TP-10-9-COMP	129	-0.0514	-7470A	PPB		24 Jan 2020 18:44:38	U	SMPL
CCV16 - 1	CCV16	19999	5.1040	-7470A	PPB		24 Jan 2020 18:46:44	U	SMPL
CCB16 - 1	CCB16	130	-0.0512	-7470A	PPB		24 Jan 2020 18:48:50	U	SMPL
L1247-30 - 1	TP-13-15-COMP	130	-0.0512	-7470A	PPB		24 Jan 2020 18:50:56	U	SMPL
L1247-38 - 1	TP-19-18-COMP	134	-0.0501	-7470A	PPB		24 Jan 2020 18:53:02	U	SMPL
PB126351BL - 1	PBW	110	-0.0564	-7470A	PPB		24 Jan 2020 18:55:09	U	SMPL
PB126351BS - 1	LCSW	16154	4.1064	-7470A	PPB		24 Jan 2020 18:57:15	U	SMPL
L1219-02 - 1	ST-CONCRETE	165	-0.0421	-7470A	PPB		24 Jan 2020 18:59:22	U	SMPL
L1230-02 - 1	STOCK-PILE-A	144	-0.0475	-7470A	PPB		24 Jan 2020 19:01:29	U	SMPL
L1230-04 - 1	STOCK-PILE-B	145	-0.0473	-7470A	PPB		24 Jan 2020 19:03:35	U	SMPL
L1230-04DUP - 1	STOCK-PILE-BDUP	133	-0.0504	-7470A	PPB		24 Jan 2020 19:05:43	U	SMPL
L1230-04MS - 1	STOCK-PILE-BMS	17793	4.5316	-7470A	PPB		24 Jan 2020 19:07:50	U	SMPL
L1230-04MSD - 1	STOCK-PILE-BMSD	17561	4.4714	-7470A	PPB		24 Jan 2020 19:09:56	U	SMPL
CCV17 - 1	CCV17	20349	5.1948	-7470A	PPB		24 Jan 2020 19:12:03	U	SMPL
CCB17 - 1	CCB17	95	-0.0602	-7470A	PPB		24 Jan 2020 19:14:09	U	SMPL
PB126353BL - 1	PBW	93	-0.0608	-7470A	PPB		24 Jan 2020 19:16:15	U	SMPL
PB126353BS - 1	LCSW	15771	4.0070	-7470A	PPB		24 Jan 2020 19:18:21	U	SMPL
L1173-03 - 1	0052-DECANT-H2O-COMP-4	270	-0.0148	-7470A	PPB		24 Jan 2020 19:20:28	U	SMPL
L1173-04 - 1	0051-DECANT-H2O-COMP-3	404	0.0199	-7470A	PPB		24 Jan 2020 19:22:34	U	SMPL
L1173-05 - 1	0049-DECANT-H2O-COMP-1	150	-0.0460	-7470A	PPB		24 Jan 2020 19:24:41	U	SMPL
L1173-06 - 1	0048-SITE-WATER	129	-0.0514	-7470A	PPB		24 Jan 2020 19:26:48	U	SMPL
L1173-06DUP - 1	0048-SITE-WATERDUP	120	-0.0538	-7470A	PPB		24 Jan 2020 19:28:54	U	SMPL
L1173-06MS - 1	0048-SITE-WATERMS	19007	4.8466	-7470A	PPB		24 Jan 2020 19:31:02	U	SMPL
L1173-06MSD - 1	0048-SITE-WATERMSD	0	-0.0849	-7470A	PPB		24 Jan 2020 19:33:09	U	SMPL
L1236-01 - 1	COMP-1	353	0.0067	-7470A	PPB		24 Jan 2020 19:38:32	U	SMPL
CCV18 - 1	CCV18	19697	5.0256	-7470A	PPB		24 Jan 2020 19:40:38	U	SMPL
CCB18 - 1	CCB18	67	-0.0675	-7470A	PPB		24 Jan 2020 19:42:45	U	SMPL
L1236-07 - 1	COMP-2	25	-0.0784	-7470A	PPB		24 Jan 2020 19:44:51	U	SMPL
L1245-06 - 1	20200047-FIELD-BLANK	26	-0.0781	-7470A	PPB		24 Jan 2020 19:46:57	U	SMPL
PB126350TB - 1	PB126302TB	87	-0.0623	-7470A	PPB		24 Jan 2020 19:49:03	U	SMPL
PB126351TB - 1	PB126281TB	82	-0.0636	-7470A	PPB		24 Jan 2020 19:51:09	U	SMPL
L1007-10LX5 - 1		94	-0.0605	-7470A	PPB		24 Jan 2020 19:53:16	U	SMPL
L1230-04LX5 - 1		92	-0.0610	-7470A	PPB		24 Jan 2020 19:55:23	U	SMPL
L1173-06LX5 - 1		136	-0.0496	-7470A	PPB		24 Jan 2020 19:57:30	U	SMPL
CCV19 - 1	CCV19	20000	5.1043	-7470A	PPB		24 Jan 2020 19:59:38	U	SMPL
CCB19 - 1	CCB19	100	-0.0589	-7470A	PPB		24 Jan 2020 20:01:46	U	SMPL