


LB107367

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.081	-0.081	86	0.000	86				
0.2	0.200	0.147	-0.053	1385	0.0 %	1385				
2.5	2.500	2.548	0.048	15072	0.0 %	15072				
5.0	5.000	5.193	0.193	30143	0.0 %	30143				
7.5	7.500	7.531	0.031	43468	0.0 %	43468				
10.0	10.000	9.863	-0.137	56760	0.0 %	56760				

LB107367 INSTRUMENT ID:CV1

Method: 7471B		Operator: Admin		Date of Analysis: 27 Jan 2020 14:30:25					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD.0	86	-	0.0000	7471B	PPB	27 Jan 2020 14:42:07	S	Std
0.2 - 1	SD.2	1385	-	0.2000	7471B	PPB	27 Jan 2020 14:44:24	S	Std
2.5 - 1	SD.5	15072	-	2.5000	7471B	PPB	27 Jan 2020 14:46:41	S	Std
5.0 - 1	SD.10	30143	-	5.0000	7471B	PPB	27 Jan 2020 14:48:57	S	Std
7.5 - 1	SD.15	43468	-	7.5000	7471B	PPB	27 Jan 2020 14:51:13	S	Std
10.0 - 1	SD.20	56760	-	10.0000	7471B	PPB	27 Jan 2020 15:02:50	S	Std
ICV89 - 1	ICV89	22917	3.9248	-	7471B	PPB	27 Jan 2020 15:06:44	U	SMPL
ICB89 - 1	ICB89	5	-0.0954	-	7471B	PPB	27 Jan 2020 15:08:59	U	SMPL
CCV20 - 1	CCV20	29052	5.0012	-	7471B	PPB	27 Jan 2020 15:11:14	U	SMPL
CCB20 - 1	CCB20	-1	-0.0964	-	7471B	PPB	27 Jan 2020 15:13:29	U	SMPL
CRA - 1	CRA	1415	0.1520	-	7471B	PPB	27 Jan 2020 15:15:43	U	SMPL
HighStd - 1	HighStd	56491	9.8156	-	7471B	PPB	27 Jan 2020 15:17:58	U	SMPL
ChkStd - 1	ChkStd	39403	6.8174	-	7471B	PPB	27 Jan 2020 15:20:13	U	SMPL
PB126354BL - 1	PBS	-23	-0.1003	-	7471B	PPB	27 Jan 2020 15:22:29	U	SMPL
PB126354BS - 1	LCSS	23135	3.9630	-	7471B	PPB	27 Jan 2020 15:24:45	U	SMPL
L1007-01 - 1	JC-03-012320-A	6445	1.0346	-	7471B	PPB	27 Jan 2020 15:27:01	U	SMPL
L1007-01DUP - 1	JC-03-012320-ADUP	6756	1.0892	-	7471B	PPB	27 Jan 2020 15:29:18	U	SMPL
L1007-01MS - 1	JC-03-012320-AMS	28130	4.8394	-	7471B	PPB	27 Jan 2020 15:31:34	U	SMPL
L1007-01MSD - 1	JC-03-012320-AMSD	29633	5.1031	-	7471B	PPB	27 Jan 2020 15:33:52	U	SMPL
L1007-11 - 1	JC-03-012320-B	7782	1.2692	-	7471B	PPB	27 Jan 2020 15:36:07	U	SMPL
CCV21 - 1	CCV21	29542	5.0872	-	7471B	PPB	27 Jan 2020 15:38:22	U	SMPL
CCB21 - 1	CCB21	11	-0.0943	-	7471B	PPB	27 Jan 2020 15:40:37	U	SMPL
L1007-19 - 1	JC-03-012320-C	7050	1.1407	-	7471B	PPB	27 Jan 2020 15:42:52	U	SMPL
L1010-10 - 1	ok-01-012420	1128	0.1017	-	7471B	PPB	27 Jan 2020 15:45:07	U	SMPL
L1017-07 - 1	TR-04-012420	452	-0.0169	-	7471B	PPB	27 Jan 2020 15:47:22	U	SMPL
L1234-01 - 1	NS-TP	15995	2.7102	-	7471B	PPB	27 Jan 2020 15:49:38	U	SMPL
L1237-01 - 1	PATERSON-UST	1161	0.1075	-	7471B	PPB	27 Jan 2020 15:51:54	U	SMPL
L1238-01 - 1	CLIFTON-POLES	4103	0.6237	-	7471B	PPB	27 Jan 2020 15:54:10	U	SMPL
L1243-07 - 1	CTY-SB-04-0-5	25529	4.3831	-	7471B	PPB	27 Jan 2020 15:56:26	U	SMPL
L1243-08 - 1	CTY-SB-BC-0-3.5	41589	7.2009	-	7471B	PPB	27 Jan 2020 15:58:43	U	SMPL
L1246-01 - 1	SP-1	965	0.0731	-	7471B	PPB	27 Jan 2020 16:01:00	U	SMPL
CCV22 - 1	CCV22	699	0.0264	-	7471B	PPB	27 Jan 2020 16:03:15	U	SMPL
CCB22 - 1	CCB22	30495	5.2544	-	7471B	PPB	27 Jan 2020 16:05:30	U	SMPL
L1246-03 - 1	SP-2	25	-0.0919	-	7471B	PPB	27 Jan 2020 16:07:46	U	SMPL
L1246-05 - 1	SP-3	2172	0.2849	-	7471B	PPB	27 Jan 2020 16:12:16	U	SMPL
L1247-05 - 1	TP-4-5-COMP	2493	0.3412	-	7471B	PPB	27 Jan 2020 16:14:31	U	SMPL
L1247-13 - 1	TP-7-6-COMP	1247	0.1226	-	7471B	PPB	27 Jan 2020 16:16:46	U	SMPL
L1247-21 - 1	TP-10-9-COMP	1317	0.1348	-	7471B	PPB	27 Jan 2020 16:19:02	U	SMPL
L1247-29 - 1	TP-13-15-COMP	818	0.0473	-	7471B	PPB	27 Jan 2020 16:21:18	U	SMPL
L1247-37 - 1	TP-19-18-COMP	828	0.0490	-	7471B	PPB	27 Jan 2020 16:23:34	U	SMPL
L1242-01 - 1	20200055-COMPOSITE-1	32105	5.5369	-	7471B	PPB	27 Jan 2020 16:25:50	U	SMPL
L1242-01DUP - 1	20200055-COMPOSITE-1DUP	32389	5.5867	-	7471B	PPB	27 Jan 2020 16:28:06	U	SMPL
CCV23 - 1	CCV23	30565	5.2667	-	7471B	PPB	27 Jan 2020 16:32:39	U	SMPL
CCB23 - 1	CCB23	20	-0.0927	-	7471B	PPB	27 Jan 2020 16:34:56	U	SMPL
L1242-01MSD - 1	20200055-COMPOSITE-1MSD	57503	9.9932	-	7471B	PPB	27 Jan 2020 16:37:11	U	SMPL
L1242-02 - 1	20200056-COMPOSITE-2	32625	5.6281	-	7471B	PPB	27 Jan 2020 16:39:27	U	SMPL
L1242-03 - 1	20200057-COMPOSITE-3	29180	5.0237	-	7471B	PPB	27 Jan 2020 16:41:42	U	SMPL
L1242-04 - 1	20200058-COMPOSITE-4	29468	5.0742	-	7471B	PPB	27 Jan 2020 16:43:58	U	SMPL
L1242-05 - 1	20200059-COMPOSITE-5	41274	7.1457	-	7471B	PPB	27 Jan 2020 16:46:13	U	SMPL
L1242-06 - 1	20200054-FIELD-DUP-COMP	26728	4.5934	-	7471B	PPB	27 Jan 2020 16:48:29	U	SMPL
L1242-07 - 1	20200060-COMPOSITE-B6	27318	4.6970	-	7471B	PPB	27 Jan 2020 16:50:45	U	SMPL
L1242-08 - 1	20200061-COMPOSITE-B7	34395	5.9387	-	7471B	PPB	27 Jan 2020 16:53:01	U	SMPL
L1242-09 - 1	20200062-COMPOSITE-B8	45038	7.8061	-	7471B	PPB	27 Jan 2020 16:55:17	U	SMPL
L1242-10 - 1	20200063-AMENDED-COMP-1	24513	4.2048	-	7471B	PPB	27 Jan 2020 16:57:35	U	SMPL
CCV24 - 1	CCV24	27668	4.7584	-	7471B	PPB	27 Jan 2020 16:59:50	U	SMPL
CCB24 - 1	CCB24	49	-0.0876	-	7471B	PPB	27 Jan 2020 17:02:05	U	SMPL
L1242-11 - 1	20200064-AMENDED-COMP-2	22937	3.9283	-	7471B	PPB	27 Jan 2020 17:04:21	U	SMPL
L1242-12 - 1	20200065-AMENDED-COMP-3	16185	2.7436	-	7471B	PPB	27 Jan 2020 17:06:36	U	SMPL
L1242-13 - 1	20200066-AMENDED-COMP-4	29617	5.1003	-	7471B	PPB	27 Jan 2020 17:08:51	U	SMPL
L1242-14 - 1	20200067-AMENDED-COMP-5	26885	4.6210	-	7471B	PPB	27 Jan 2020 17:11:07	U	SMPL
PB126389BL - 1	PBS	39	-0.0894	-	7471B	PPB	27 Jan 2020 17:13:22	U	SMPL
PB126389BS - 1	LCSS	24907	4.2739	-	7471B	PPB	27 Jan 2020 17:15:37	U	SMPL
L1007-01LX5 - 1		1313	0.1341	-	7471B	PPB	27 Jan 2020 17:17:53	U	SMPL
L1242-01LX5 - 1		6473	1.0395	-	7471B	PPB	27 Jan 2020 17:20:10	U	SMPL
L1242-01MS - 1	20200055-COMPOSITE-1MS	54887	9.5342	-	7471B	PPB	27 Jan 2020 17:22:27	U	SMPL
CCV25 - 1	CCV25	29328	5.0496	-	7471B	PPB	27 Jan 2020 17:24:44	U	SMPL
CCB25 - 1	CCB25	35	-0.0901	-	7471B	PPB	27 Jan 2020 17:27:00	U	SMPL