

LB107609

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

Instrument ID : Konelab

2/7/2020 8:50

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.652	0.0	0.102	
ICB1	0.336	0.0	0.003	
CCV1	243.585	0.0	0.253	
CCB1	0.403	0.0	0.003	
PB126628BL	0.000	0.0	0.002	
L1365-08	0.002	0.0	0.002	
L1365-08DUP	0.019	0.0	0.002	
L1394-09	0.061	0.0	0.002	
CCV2	242.249	0.0	0.252	
CCB2	0.396	0.0	0.003	

N	10
Mean	58.370
SD	101.8110
CV%	174.42

Aquakem v. 7.2AQ1

AK

Results from time period:

Thu Feb 06 16:55:08 2020

Thu Feb 06 16:55:17 2020

Sample Id	Sam/Ctr/c	Test short Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive Cl P	0.3472	µg/l	2/6/2020 8:54:27	
5.OPPBCN	A	Reactive Cl P	4.7562	µg/l	2/6/2020 8:54:28	
10PPBCN	A	Reactive Cl P	9.5506	µg/l	2/6/2020 8:54:29	
50PPBCN	A	Reactive Cl P	49.9391	µg/l	2/6/2020 8:54:30	
100PPBCN	A	Reactive Cl P	99.9994	µg/l	2/6/2020 8:54:31	
250PPBCN	A	Reactive Cl P	250.7798	µg/l	2/6/2020 8:54:32	
500PPBCN	A	Reactive Cl P	499.6277	µg/l	2/6/2020 8:54:33	
ICV1	S	Reactive Cl P	96.6515	µg/l	2/6/2020 16:55:08	
ICB1	S	Reactive Cl P	0.336	µg/l	2/6/2020 16:55:09	
CCV1	S	Reactive Cl P	243.5854	µg/l	2/6/2020 16:55:10	
CCB1	S	Reactive Cl P	0.4035	µg/l	2/6/2020 16:55:11	
PB126628BL	S	Reactive Cl P	0.0004	µg/l	2/6/2020 16:55:12	
L1365-08	S	Reactive Cl P	0.0022	µg/l	2/6/2020 16:55:13	
L1365-08DUP	S	Reactive Cl P	0.0189	µg/l	2/6/2020 16:55:14	
L1394-09	S	Reactive Cl P	0.0609	µg/l	2/6/2020 16:55:15	
CCV2	S	Reactive Cl P	242.2488	µg/l	2/6/2020 16:55:16	
CCB2	S	Reactive Cl P	0.3964	µg/l	2/6/2020 16:55:17	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

Instrument ID : Konelab

2/6/2020 16:01


Test Total CN

Accepted 2/6/2020 8:59

Factor 971.7
Bias 0.002

Coeff. of det. 0.999995

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.3472	0.0000	
2	5.0PPBCN	0.007	4.7562	5.0000	
3	10PPBCN	0.012	9.5506	10.0000	
4	50PPBCN	0.054	49.9391	50.0000	
5	100PPBCN	0.105	99.9994	100.0000	
6	250PPBCN	0.260	250.7798	250.0000	
7	500PPBCN	0.517	499.6277	500.0000	