

LB107628

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AR

Instrument ID : Konelab

2/7/2020 13:10

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.027	0.0	0.102	
ICB1	-0.007	0.0	0.003	
CCV1	244.164	0.0	0.251	
CCB1	-0.444	0.0	0.002	
PB126668BL	-0.430	0.0	0.002	
PB126668BS	98.097	0.0	0.103	
MIDPB126668	246.408	0.0	0.253	
L1327-02	3.750	0.0	0.007	
L1425-01MS	37.779	0.0	0.041	
L1425-01MSD	38.247	0.0	0.042	
L1425-01	0.153	0.0	0.003	
L1425-01DUP	0.151	0.0	0.003	
L1434-01	0.022	0.0	0.003	
CCV2	261.306	0.0	0.268	
CCB2	-0.547	0.0	0.002	

N	15
Mean	68.378
SD	100.1507
CV%	146.47

AK

Aquakem v. 7.2AQ1

Results from time period:

Fri Feb 07 11:51:33 2020

Fri Feb 07 13:07:27 2020

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.631	µg/l	2/7/2020 11:51:33	
5.OPPBCN	A	Total CN	P	4.201	µg/l	2/7/2020 11:51:34	
10PPBCN	A	Total CN	P	9.2921	µg/l	2/7/2020 11:51:35	
50PPBCN	A	Total CN	P	50.0669	µg/l	2/7/2020 11:51:36	
100PPBCN	A	Total CN	P	100.7766	µg/l	2/7/2020 11:51:37	
250PPBCN	A	Total CN	P	252.8686	µg/l	2/7/2020 11:51:38	
500PPBCN	A	Total CN	P	498.4258	µg/l	2/7/2020 11:51:39	
ICV1	S	Total CN	P	97.0265	µg/l	2/7/2020 13:03:05	
ICB1	S	Total CN	P	-0.0073	µg/l	2/7/2020 13:03:06	
CCV1	S	Total CN	P	244.1643	µg/l	2/7/2020 13:03:07	
CCB1	S	Total CN	P	-0.4444	µg/l	2/7/2020 13:03:08	
PB126668BL	S	Total CN	P	-0.4302	µg/l	2/7/2020 13:03:09	
PB126668BS	S	Total CN	P	98.0969	µg/l	2/7/2020 13:03:10	
MIDPB126668	S	Total CN	P	246.4077	µg/l	2/7/2020 13:03:11	
L1327-02	S	Total CN	P	3.7498	µg/l	2/7/2020 13:03:12	
L1425-01MS	S	Total CN	P	37.7786	µg/l	2/7/2020 13:03:13	
L1425-01MSD	S	Total CN	P	38.2469	µg/l	2/7/2020 13:03:14	
L1425-01	S	Total CN	P	0.1525	µg/l	2/7/2020 13:03:15	
L1425-01DUP	S	Total CN	P	0.1505	µg/l	2/7/2020 13:07:24	
L1434-01	S	Total CN	P	0.0218	µg/l	2/7/2020 13:07:25	
CCV2	S	Total CN	P	261.3064	µg/l	2/7/2020 13:07:26	
CCB2	S	Total CN	P	-0.5475	µg/l	2/7/2020 13:07:27	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : AK

Instrument ID : Konelab

2/7/2020 12:00

Test Total CN


Accepted 2/7/2020 12:00

Factor 983.8

Bias 0.003

Coeff. of det. 0.999937

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.002	-0.6310	0.0000	
2	5.0PPBCN	0.007	4.2010	5.0000	
3	10PPBCN	0.012	9.2921	10.0000	
4	50PPBCN	0.054	50.0669	50.0000	
5	100PPBCN	0.105	100.7766	100.0000	
6	250PPBCN	0.260	252.8686	250.0000	
7	500PPBCN	0.509	498.4258	500.0000	