

TitriSoft Report

Ident	Amount	DateTime	Alkalinity	Final pH (sample)	Initial pH (sample)	offset	slope
LB108439BLW	100	25/03/2020 16:26	0.00	4.39	4.79	6.93	102.8
LB108439BSW	100	25/03/2020 16:28	46.24	4.44	8.50	6.93	102.8
L2018-01	50	25/03/2020 16:34	81.68	4.43	5.63	6.93	102.8
L2018-01DUP	50	25/03/2020 16:39	79.20	4.39	5.69	6.93	102.8
L2018-02	50	25/03/2020 16:43	247.20	4.42	6.30	6.93	102.8
L2018-03	50	25/03/2020 16:48	276.88	4.45	6.50	6.93	102.8
L2018-04	50	25/03/2020 16:54	431.68	4.45	6.63	6.93	102.8

3/25/2020 AM