

IC-2

M300.0

4/20/2020
AM


Clear table

ident	concentra tion F-	concentrati on CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-04-13_	4/13/20 15:26		AM/AP
STD2	0.2410	0.3470	0.2420	1.1240	0.2900	0.4930	1.7210	2020-04-13_	4/13/20 16:00		AM/AP
STD3	0.4190	0.6230	0.4020	2.0950	0.5210	0.9820	3.0600	2020-04-13_	4/13/20 16:34		AM/AP
STD4	0.8080	1.1820	1.4240	3.9430	0.9840	1.9570	5.9500	2020-04-13_	4/13/20 17:08		AM/AP
STD5	1.9230	2.9400	3.1140	9.8040	2.4490	5.0850	14.7260	2020-04-13_	4/13/20 17:43		AM/AP
STD6	3.9330	5.9330	5.9810	19.8200	4.9400	10.0510	29.7380	2020-04-13_	4/13/20 18:17		AM/AP
STD7	5.0760	7.5740	7.4380	25.2130	6.3160	12.4330	37.8060	2020-04-13_	4/13/20 18:52		AM/AP
ICV	1.9530	2.9390	3.1230	9.8040	2.4650	5.2060	14.5490	2020-04-13_	4/13/20 19:27		AM/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-04-13_	4/13/20 20:09		AM/AP
CCV	1.9560	2.9500	3.0180	9.5190	2.4740	4.2940	16.3550	2020-04-20_	4/20/20 10:21		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-04-20_	4/20/20 10:53		AM/AP
LB108810BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-04-20_	4/20/20 11:26		AM/AP
LB108810BSW	1.9200	2.9510	3.0300	9.5490	2.3930	4.3970	16.3780	2020-04-20_	4/20/20 11:59		AM/AP
L2354-01	0.0000	0.2490	0.0000	0.0000	0.0970	0.0000	0.6130	2020-04-20_	4/20/20 13:29		AM/AP
L2354-01DUP	0.0000	0.2460	0.0000	0.0000	0.1040	0.0000	0.6050	2020-04-20_	4/20/20 14:02		AM/AP
L2354-01MS	1.9160	3.0660	2.9970	9.4170	2.4750	4.3840	16.3950	2020-04-20_	4/20/20 14:35		AM/AP
L2354-01MSD	1.8930	3.0660	3.0200	9.4840	2.5700	4.4630	16.5280	2020-04-20_	4/20/20 15:07		AM/AP
CCV	1.8850	2.9650	3.0640	9.5780	2.3950	4.5430	16.3460	2020-04-20_	4/20/20 15:40		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-04-20_	4/20/20 16:13		AM/AP

CALIBRATION OF COMPONENT F-

Method: AnionsIC2-041320.mtw
 Equation: $Q = 0.0799175 \cdot A + 2.57571$
 RSD: 3.271 %
 Correlation coefficient: 0.999548

IC-2
 M300.0
 4/14/2020
 AM




K3 = 0 K2 = 0 K1 = 0.0799175 K0 = 2.57571
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1		0	0	0	0		0
2	0.8585		28	0.2	20		6.085
3	2.057		72.7	0.4	20		6.085
4	4.508		169.9	0.8	20		6.085
5	12.95		449	2	20		6.085
6	28.99		952.1	4	20		6.085
7	36.08		1238	5	20		6.085

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-041320.mtw
 Equation: $Q = 0.127058 \cdot A + 2.70244$
 RSD: 2.088 %
 Correlation coefficient: 0.999816




K3 = 0 K2 = 0 K1 = 0.127058 K0 = 2.70244
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	2.27	33.42	0.3	20	9.528		
3	4.803	76.85	0.6	20	9.528		
4	9.789	164.7	1.2	20	9.528		
5	25.74	441.5	3	20	9.528		
6	52.89	912.6	6	20	9.528		
7	67.09	1171	7.5	20	9.528		

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-041320.mtw
 Equation: $Q = 0.0537393 \cdot A + 2.15169$
 RSD: 5.338 %
 Correlation coefficient: 0.998795




K3 = 0 K2 = 0 K1 = 0.0537393 K0 = 2.15169
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	3.139	50	0.3	20	11.62		
3	6.229	109.7	0.6	20	11.62		
4	12.92	489.8	1.2	20	11.62		
5	28.77	1119	3	20	11.62		
6	51.12	2186	6	20	11.62		
7	61.95	2728	7.5	20	11.62		

1.□

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-041320.mtw
 Equation: $Q = 0.284412 \cdot A + 4.78117$
 RSD: 1.833 %
 Correlation coefficient: 0.999858




K3 = 0 K2 = 0 K1 = 0.284412 K0 = 4.78117
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	2.344	62.25	1	20	15.26	
3	4.866	130.5	2	20	15.26	
4	9.803	260.5	4	20	15.26	
5	25.9	672.6	10	20	15.26	
6	54.33	1377	20	20	15.26	
7	69.67	1756	25	20	15.26	

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-041320.mtw
 Equation: $Q = 0.0503991 \cdot A + 1.1825$
 RSD: 2.190 %
 Correlation coefficient: 0.999797




K3 = 0 K2 = 0 K1 = 0.0503991 K0 = 1.1825
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	2.846	91.73	0.25	20	17.81		
3	5.741	183.1	0.5	20	17.81		
4	11.42	367.2	1	20	17.81		
5	29.34	948.5	2.5	20	17.81		
6	60.18	1937	5	20	17.81		
7	76.76	2483	6.25	20	17.81		

□□□

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-041320.mtw
 Equation: $Q = 0.0865819 \cdot A + 3.20646$
 RSD: 1.246 %
 Correlation coefficient: 0.999934




K3 = 0 K2 = 0 K1 = 0.0865819 K0 = 3.20646
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		
2	1.747	76.79	0.5	20	23.96		
3	4.032	189.7	1	20	23.96		
4	8.619	415.1	2	20	23.96		
5	24.26	1138	5	20	23.96		
6	51.49	2285	10	20	23.96		
7	65.1	2835	12.5	20	23.96		

1.□

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-041320.mtw
 Equation: $Q = 0.192218 \cdot A + 7.31348$
 RSD: 1.743 %
 Correlation coefficient: 0.999872



K3 = 0 K2 = 0 K1 = 0.192218 K0 = 7.31348
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		
2	3.713	141	1.5	20	29.59		
3	7.436	280.3	3	20	29.59		
4	14.93	581	6	20	29.59		
5	38.9	1494	15	20	29.59		
6	80.8	3056	30	20	29.59		
7	103.7	3896	37.5	20	29.59		

Report date: 4/14/2020 9:52:15 AM
Printed by: wet


Ident: STD1
Analysis from: 4/13/2020 3:26:13 PM
File: _2020-04-13_

Last save: 4/14/2020 9:52:09 AM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 24977

Last save: 4/13/2020 6:49:

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/14/2020 9:52:22 AM
Printed by: wet


Ident: STD2
Analysis from: 4/13/2020 4:00:21 PM
File: _2020-04-13_

Last save: 4/14/2020 9:52:09 AM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 24978

Last save: 4/13/2020 6:49:

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.08	0.448	0.86	4.74	28.005	5.30
2	9.53	0.201	2.27	12.54	33.420	6.32
3	11.62	0.247	3.14	17.36	50.001	9.46
4	15.26	0.347	2.34	12.94	62.252	11.78
5	17.81	0.415	2.85	15.72	91.727	17.35
6	23.96	0.516	1.75	9.64	76.794	14.53
7	29.59	0.538	3.71	20.50	140.993	26.67
7	31.50	0.387	16.91	93.42	483.192	91.40

This report has been created by IC Net
METROHM LTD

Report date: 4/14/2020 9:52:30 AM
Printed by: wet


Ident: STD3
Analysis from: 4/13/2020 4:34:29 PM
File: _2020-04-13_

Last save: 4/14/2020 9:52:09 AM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 24979

Last save: 4/13/2020 6:49:

SAMPLE:
: AM/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.08	0.445	2.05	5.72	72.699	6.79
2	9.53	0.210	4.80	13.36	76.848	7.18
3	11.61	0.271	6.23	17.32	109.729	10.25
4	15.25	0.345	4.86	13.53	130.494	12.18
5	17.77	0.410	5.74	15.96	183.133	17.10
6	23.99	0.522	4.03	11.21	189.708	17.71
7	29.62	0.535	7.43	20.67	280.319	26.17
7	31.50	0.391	35.16	97.77	1042.929	97.38

This report has been created by IC Net
METROHM LTD

Report date: 4/14/2020 9:52:37 AM
Printed by: wet


Ident: STD4
Analysis from: 4/13/2020 5:08:37 PM
File: _2020-04-13_

Last save: 4/14/2020 9:52:09 AM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 24980

Last save: 4/13/2020 6:49:

SAMPLE:
: AM/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.10	0.468	4.51	6.26	169.898	6.94
2	9.54	0.216	9.79	13.60	164.739	6.73
3	11.59	0.351	12.92	17.94	489.766	20.01
4	15.23	0.339	9.80	13.62	260.495	10.64
5	17.73	0.411	11.42	15.86	367.155	15.00
6	24.03	0.528	8.62	11.97	415.078	16.95
7	29.66	0.535	14.93	20.73	580.993	23.73
7	32.00	0.407	71.98	100.00	2448.124	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/14/2020 8:47:58 AM
Printed by: wet


Ident: STD5
Analysis from: 4/13/2020 5:43:15 PM
File: _2020-04-13_

Last save: 4/13/2020 8:05:36 PM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 24981

Last save: 4/13/2020 6:49:

SAMPLE:
: AM/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.09	0.446	12.95	6.97	449.000	7.17
2	9.52	0.218	25.74	13.85	441.512	7.05
3	11.51	0.435	28.77	15.48	1118.733	17.87
4	15.18	0.329	25.90	13.93	672.643	10.74
5	17.63	0.411	29.34	15.79	948.539	15.15
6	24.04	0.513	24.26	13.05	1137.548	18.17
7	29.64	0.529	38.90	20.93	1494.171	23.86
7	32.00	0.411	185.86	100.00	6262.145	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/14/2020 8:48:09 AM
Printed by: wet


Ident: STD6
Analysis from: 4/13/2020 6:17:53 PM
File: _2020-04-13_

Last save: 4/13/2020 8:05:36 PM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 24982

Last save: 4/13/2020 6:49:

SAMPLE:
: AM/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.10	0.427	28.98	7.63	952.098	7.49
2	9.50	0.219	52.89	13.93	912.641	7.18
3	11.44	0.540	51.12	13.46	2185.844	17.20
4	15.15	0.319	54.33	14.30	1376.954	10.84
5	17.55	0.407	60.18	15.84	1936.778	15.24
6	24.06	0.489	51.49	13.56	2284.702	17.98
7	29.64	0.522	80.80	21.27	3056.173	24.05
7	32.00	0.418	379.77	100.00	12705.189	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/14/2020 8:48:19 AM
Printed by: wet


Ident: STD7
Analysis from: 4/13/2020 6:52:30 PM
File: _2020-04-13_

Last save: 4/13/2020 8:06:24 PM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 24983

Last save: 4/13/2020 6:49:

SAMPLE:
: AM/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.12	0.456	36.08	7.51	1238.123	7.69
2	9.51	0.220	67.09	13.97	1171.020	7.27
3	11.42	0.577	61.95	12.90	2728.002	16.94
4	15.17	0.316	69.66	14.50	1756.166	10.90
5	17.55	0.407	76.76	15.98	2482.801	15.41
6	24.11	0.484	65.10	13.55	2834.817	17.60
7	29.68	0.521	103.72	21.59	3895.579	24.19
7	32.00	0.426	480.36	100.00	16106.509	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/14/2020 8:48:29 AM
Printed by: wet


Ident: ICV
Analysis from: 4/13/2020 7:27:08 PM
File: _2020-04-13_

Last save: 4/13/2020 8:06:24 PM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 24984

Last save: 4/13/2020 6:49:

SAMPLE:
: AM/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.10	0.458	12.86	6.97	456.417	7.26
2	9.53	0.219	25.41	13.77	441.391	7.02
3	11.54	0.439	28.30	15.34	1122.257	17.85
4	15.22	0.330	25.63	13.89	672.597	10.70
5	17.68	0.413	29.02	15.73	954.611	15.18
6	24.08	0.515	24.59	13.33	1165.637	18.54
7	29.71	0.531	38.70	20.97	1475.745	23.47
7	32.00	0.415	184.51	100.00	6288.654	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/14/2020 8:48:38 AM
Printed by: wet


Ident: ICB
Analysis from: 4/13/2020 8:09:20 PM
File: _2020-04-13_

Last save: 4/13/2020 9:18:59 PM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 24985

Last save: 4/13/2020 8:07:

SAMPLE:
:
Vial number: AM/AP
Volume: 10
Dilution: 20.0 μ L
Amount: 1.00
1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/21/2020 9:25:05 AM
Printed by: wet


Ident: CCV
Analysis from: 4/20/2020 10:21:11 AM
File: _2020-04-20_

Last save: 4/20/2020 12:39:18 PM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 25057

Last save: 4/16/2020 10:57

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.03	0.502	11.80	7.26	457.339	7.36
2	9.35	0.242	22.98	14.13	443.112	7.13
3	11.33	0.742	24.47	15.05	1083.269	17.44
4	14.89	0.373	22.05	13.56	652.542	10.50
5	17.30	0.477	24.91	15.32	958.142	15.42
6	22.65	0.525	19.57	12.03	954.781	15.37
7	27.39	0.525	36.81	22.64	1663.704	26.78
7	30.00	0.484	162.59	99.99	6212.889	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/21/2020 9:25:26 AM
Printed by: wet


Ident: CCB
Analysis from: 4/20/2020 10:53:49 AM
File: _2020-04-20_

Last save: 4/20/2020 11:49:25 AM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 25058

Last save: 4/16/2020 10:57

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/21/2020 9:25:46 AM
Printed by: wet


Ident: LB108810BLW
Analysis from: 4/20/2020 11:26:27 AM
File: _2020-04-20_

Last save: 4/21/2020 9:25:45 AM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 25059

Last save: 4/16/2020 10:57

SAMPLE:
: AM/AP
Vial number: 4
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/21/2020 9:25:58 AM
Printed by: wet


Ident: LB108810BSW
Analysis from: 4/20/2020 11:59:05 AM
File: _2020-04-20_

Last save: 4/21/2020 9:24:39 AM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 25060

Last save: 4/16/2020 10:57

SAMPLE:
: AM/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.96	0.474	12.21	7.43	448.371	7.23
2	9.24	0.246	22.91	13.93	443.312	7.14
3	11.18	0.623	25.55	15.53	1087.722	17.53
4	14.68	0.372	22.30	13.56	654.651	10.55
5	17.05	0.470	25.15	15.29	926.335	14.93
6	22.57	0.529	20.15	12.25	978.755	15.77
7	27.31	0.532	36.20	22.01	1666.106	26.85
7	30.00	0.464	164.46	100.00	6205.252	100.00

Report date: 4/21/2020 9:26:13 AM
Printed by: wet


Ident: L2354-01
Analysis from: 4/20/2020 1:29:51 PM
File: _2020-04-20_

Last save: 4/20/2020 5:27:20 PM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 25062

Last save: 4/16/2020 10:57

SAMPLE:
: AM/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	9.24	0.217	1.06	48.74	17.962	30.66
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	17.29	0.418	0.56	25.96	14.900	25.43
6	0.00	0.000	0.00	0.00	0.000	0.00
7	27.32	0.560	0.55	25.11	25.724	43.91
7	30.00	0.171	2.17	99.81	58.586	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/21/2020 9:26:22 AM
Printed by: wet


Ident: L2354-01DUP
Analysis from: 4/20/2020 2:02:29 PM
File: _2020-04-20_

Last save: 4/20/2020 5:27:20 PM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 25063

Last save: 4/16/2020 10:57

SAMPLE:
: AM/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	9.25	0.223	1.01	49.36	17.383	29.04
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	17.32	0.443	0.51	24.97	17.612	29.43
6	0.00	0.000	0.00	0.00	0.000	0.00
7	27.35	0.568	0.52	25.46	24.855	41.53
7	30.00	0.176	2.05	99.79	59.850	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/21/2020 9:26:32 AM
Printed by: wet


Ident: L2354-01MS
Analysis from: 4/20/2020 2:35:07 PM
File: _2020-04-20_

Last save: 4/20/2020 5:27:57 PM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 25064

Last save: 4/16/2020 10:57

SAMPLE:
: AM/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.97	0.487	11.92	7.24	447.142	7.18
2	9.25	0.246	23.82	14.47	461.328	7.40
3	11.20	0.645	24.78	15.05	1075.388	17.26
4	14.71	0.371	22.02	13.37	645.416	10.36
5	17.08	0.468	26.04	15.81	958.869	15.39
6	22.61	0.535	19.78	12.01	975.687	15.66
7	27.36	0.533	36.29	22.04	1667.816	26.76
7	30.00	0.469	164.64	100.00	6231.647	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/21/2020 9:26:45 AM
Printed by: wet


Ident: L2354-01MSD
Analysis from: 4/20/2020 3:07:45 PM
File: _2020-04-20_

Last save: 4/20/2020 5:27:57 PM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 25065

Last save: 4/16/2020 10:57

SAMPLE:
: AM/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.95	0.487	11.66	6.98	441.592	7.00
2	9.22	0.243	24.08	14.42	461.303	7.31
3	11.17	0.595	25.38	15.20	1083.815	17.18
4	14.66	0.369	22.32	13.37	650.093	10.30
5	17.01	0.467	27.14	16.25	996.209	15.79
6	22.57	0.540	19.92	11.93	994.010	15.76
7	27.31	0.533	36.49	21.85	1681.688	26.66
7	30.00	0.462	167.00	100.00	6308.710	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/21/2020 9:26:55 AM
Printed by: wet


Ident: CCV
Analysis from: 4/20/2020 3:40:26 PM
File: _2020-04-20_

Last save: 4/20/2020 5:27:57 PM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 25066

Last save: 4/16/2020 10:57

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.95	0.480	11.75	7.14	439.477	7.04
2	9.22	0.246	23.07	14.01	445.515	7.14
3	11.16	0.611	25.52	15.50	1100.217	17.62
4	14.64	0.370	22.49	13.66	656.686	10.52
5	16.99	0.467	25.35	15.40	927.026	14.85
6	22.57	0.538	20.44	12.42	1012.402	16.21
7	27.33	0.534	35.98	21.86	1662.717	26.63
7	30.00	0.464	164.60	99.99	6244.040	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/21/2020 9:27:09 AM
Printed by: wet


Ident: CCB
Analysis from: 4/20/2020 4:13:03 PM
File: _2020-04-20_

Last save: 4/20/2020 5:28:45 PM

Method: AnionsIC2-041320.mtw
Run operator: wet
Analysis number: 25067

Last save: 4/16/2020 10:57

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD