

Clear table


IC-2
300.0
5/05/2020
AM

ident	concentra tion F-	concentra on CL-	concentra on NO2	concentra on BR-	concentra on NO3	concentra on HPO4	concentra on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-04_	5/4/20 17:07		AM/AP
STD2	0.2550	0.3710	0.4210	1.2620	0.3100	0.8170	1.9720	2020-05-04_	5/4/20 17:37		AM/AP
STD3	0.4160	0.6160	0.6430	2.0900	0.5220	1.2100	3.1800	2020-05-04_	5/4/20 22:44		AM/AP
STD4	0.7560	1.2020	1.1760	3.9660	1.0130	2.1950	5.9480	2020-05-04_	5/4/20 23:14		AM/AP
STD5	1.9550	2.8800	2.8060	9.5650	2.3730	4.0110	14.1750	2020-05-04_	5/4/20 23:45		AM/AP
STD6	3.9920	5.9530	5.9460	19.8480	4.9580	10.1380	29.7610	2020-05-05_	5/5/20 0:16		AM/AP
STD7	5.0270	7.5770	7.6070	25.2690	6.3240	12.6290	37.9640	2020-05-05_	5/5/20 0:46		AM/AP
ICV	1.8830	2.8570	2.8040	9.5470	2.3900	5.1040	14.1970	2020-05-05_	5/5/20 1:17		AM/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-05_	5/5/20 1:47		AM/AP
CCV	1.8750	2.8890	2.8310	9.6190	2.3990	5.0940	14.2370	2020-05-05_	5/5/20 2:18		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-05_	5/5/20 2:49		AM/AP
LB108995BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-05_	5/5/20 3:19		AM/AP
LB108995BSW	1.8770	2.8800	2.8230	9.5990	2.3900	5.0890	14.1740	2020-05-05_	5/5/20 3:50		AM/AP
L2485-01	0.0000	0.1290	0.0000	0.0000	0.1610	0.0000	0.6630	2020-05-05_	5/5/20 4:21		AM/AP
L2485-01DUP	0.0000	0.1250	0.0000	0.0000	0.1620	0.0000	0.6660	2020-05-05_	5/5/20 4:51		AM/AP
L2485-01MS	1.9160	3.0490	2.8630	9.7710	2.5240	5.0160	14.6290	2020-05-05_	5/5/20 5:22		AM/AP
L2485-01MSD	1.9070	2.9340	2.7840	9.5680	2.4600	4.9280	14.3200	2020-05-05_	5/5/20 5:53		AM/AP
CCV	1.9390	2.9030	2.8410	9.6680	2.4140	5.1800	14.2430	2020-05-05_	5/5/20 6:23		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-05_	5/5/20 6:54		AM/AP

CALIBRATION OF COMPONENT F-

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.0799146 \cdot A + 1.62322$
 RSD: 2.162 %
 Correlation coefficient: 0.999802

IC-2
 M300.0
 5/05/2020
 A.M




K3 = 0 K2 = 0 K1 = 0.0799146 K0 = 1.62322
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1		0	0	0	0		0
2	1.127	43.45	0.2	20	5.75		
3	2.153	83.77	0.4	20	5.75		
4	4.324	168.9	0.8	20	5.75		
5	18.49	469	2	20	5.75		
6	27.93	978.7	4	20	5.75		
7	34.32	1238	5	20	5.75		

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.124004 \cdot A + 1.56524$
 RSD: 2.684 %
 Correlation coefficient: 0.999696




K3 = 0 K2 = 0 K1 = 0.124004 K0 = 1.56524
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0		0	0	0	
2	3.173		47.2	0.3	20	8.834
3	5.534		86.75	0.6	20	8.834
4	11.59		181.3	1.2	20	8.834
5	29		451.9	3	20	8.834
6	61.82		947.6	6	20	8.834
7	79.33		1209	7.5	20	8.834

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.0574611 \cdot A + 3.53486$
 RSD: 4.235 %
 Correlation coefficient: 0.999242




K3 = 0 K2 = 0 K1 = 0.0574611 K0 = 3.53486
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		
2	4.596	85.1	0.3	20	10.73		
3	8.527	162.4	0.6	20	10.73		
4	17.64	347.8	1.2	20	10.73		
5	43.8	915.3	3	20	10.73		
6	88.99	2008	6	20	10.73		
7	111	2586	7.5	20	10.73		

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.278566 \cdot A + 7.42532$
 RSD: 2.915 %
 Correlation coefficient: 0.999641




K3 = 0 K2 = 0 K1 = 0.278566 K0 = 7.42532
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	2.93	63.97	1	20			14
3	5.491	123.4	2	20			14
4	11.56	258.1	4	20			14
5	29.86	660.1	10	20			14
6	64.53	1398	20	20			14
7	83.39	1788	25	20			14

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.0492074 \cdot A + 1.5542$
 RSD: 3.213 %
 Correlation coefficient: 0.999564




K3 = 0 K2 = 0 K1 = 0.0492074 K0 = 1.5542
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1		0	0	0	0		
2	3.653	94.48	0.25	20	16.26		
3	6.565	180.7	0.5	20	16.26		
4	13.88	380.1	1	20	16.26		
5	33.89	932.9	2.5	20	16.26		
6	70.99	1984	5	20	16.26		
7	90.43	2539	6.25	20	16.26		

EW

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.0725012 \cdot A + 8.64363$
 RSD: 10.585 %
 Correlation coefficient: 0.995254




K3 = 0 K2 = 0 K1 = 0.0725012 K0 = 8.64363
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		
2	2.618	106.3	0.5	20	21.2		
3	5.061	214.5	1	20	21.2		
4	11.37	486.3	2	20	21.2		
5	19.79	987.3	5	20	21.2		
6	66.24	2677	10	20	21.2		
7	84.38	3365	12.5	20	21.2		

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.163467 \cdot A + 11.0703$
 RSD: 3.551 %
 Correlation coefficient: 0.999467



K3 = 0 K2 = 0 K1 = 0.163467 K0 = 11.0703
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0		0	0	0		
2	5.05	173.6	1.5	20	24.57		
3	9.227	321.3	3	20	24.57		
4	19.16	660	6	20	24.57		
5	48.9	1667	15	20	24.57		
6	105.7	3574	30	20	24.57		
7	136	4577	37.5	20	24.57		

Report date: 5/5/2020 11:07:26 AM
Printed by: wet


Ident: STD1
Analysis from: 5/4/2020 5:07:20 PM
File: _2020-05-04_

Last save: 5/5/2020 8:57:48 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25286

Last save: 5/5/2020 12:44:

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:07:36 AM
Printed by: wet


Ident: STD2
Analysis from: 5/4/2020 5:37:56 PM
File: _2020-05-04_

Last save: 5/5/2020 8:57:48 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25287

Last save: 5/5/2020 12:44:

SAMPLE:
:
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.79	0.651	1.12	4.86	43.450	7.08
2	8.76	0.201	3.17	13.71	47.200	7.69
3	10.57	0.248	4.60	19.86	85.100	13.86
4	13.64	0.299	2.93	12.65	63.970	10.42
5	15.75	0.354	3.65	15.77	94.482	15.39
6	21.23	0.472	2.62	11.31	106.275	17.31
7	24.62	0.441	5.05	21.81	173.605	28.27
7	28.00	0.381	23.14	99.97	614.082	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:07:43 AM
Printed by: wet


Ident: STD3
Analysis from: 5/4/2020 10:44:08 PM
File: _2020-05-04_

Last save: 5/5/2020 8:57:48 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25291

Last save: 5/5/2020 12:44:

SAMPLE:
:
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.532	2.15	5.06	83.771	7.14
2	8.79	0.200	5.53	13.00	86.746	7.40
3	10.61	0.243	8.53	20.04	162.449	13.85
4	13.71	0.292	5.49	12.90	123.426	10.52
5	15.84	0.351	6.56	15.42	180.651	15.40
6	21.22	0.481	5.06	11.89	214.523	18.29
7	24.61	0.443	9.23	21.68	321.316	27.40
7	28.00	0.363	42.56	99.99	1172.883	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:07:49 AM
Printed by: wet


Ident: STD4
Analysis from: 5/4/2020 11:14:46 PM
File: _2020-05-04_

Last save: 5/5/2020 8:57:48 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25292

Last save: 5/5/2020 12:44:

SAMPLE:
: AM/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.497	4.32	4.83	168.896	6.80
2	8.79	0.199	11.59	12.94	181.283	7.30
3	10.60	0.246	17.64	19.70	347.758	14.01
4	13.69	0.288	11.56	12.92	258.058	10.40
5	15.80	0.351	13.88	15.51	380.070	15.31
6	21.19	0.482	11.37	12.70	486.269	19.59
7	24.57	0.438	19.16	21.40	660.032	26.59
7	28.00	0.358	89.53	100.00	2482.366	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:07:55 AM
Printed by: wet


Ident: STD5
Analysis from: 5/4/2020 11:45:24 PM
File: _2020-05-04_

Last save: 5/5/2020 8:57:48 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25293

Last save: 5/5/2020 12:44:

SAMPLE:
: AM/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.74	0.308	18.49	8.26	468.959	7.71
2	8.79	0.198	29.00	12.96	451.931	7.43
3	10.58	0.260	43.80	19.58	915.278	15.05
4	13.69	0.283	29.86	13.35	660.055	10.85
5	15.77	0.355	33.89	15.15	932.926	15.34
6	21.19	0.550	19.79	8.85	987.266	16.23
7	24.56	0.433	48.90	21.86	1666.522	27.40
7	28.00	0.341	223.72	100.00	6082.938	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:08:02 AM
Printed by: wet


Ident: STD6
Analysis from: 5/5/2020 12:16:02 AM
File: _2020-05-05_

Last save: 5/5/2020 8:57:48 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25294

Last save: 5/5/2020 12:44:

SAMPLE:
:
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.83	0.464	27.92	5.74	978.680	7.21
2	8.80	0.195	61.82	12.71	947.582	6.98
3	10.57	0.276	88.99	18.30	2008.139	14.80
4	13.69	0.276	64.53	13.27	1398.379	10.31
5	15.73	0.359	70.99	14.60	1983.638	14.62
6	21.14	0.460	66.24	13.62	2677.320	19.73
7	24.55	0.429	105.74	21.75	3573.508	26.34
7	28.00	0.351	486.23	100.00	13567.246	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:08:08 AM
Printed by: wet


Ident: STD7
Analysis from: 5/5/2020 12:46:40 AM
File: _2020-05-05_

Last save: 5/5/2020 8:59:35 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25295

Last save: 5/5/2020 8:59:1

SAMPLE:
:
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.83	0.475	34.32	5.55	1237.684	7.15
2	8.79	0.194	79.33	12.82	1209.415	6.99
3	10.56	0.286	111.03	17.94	2586.114	14.95
4	13.69	0.272	83.39	13.47	1787.544	10.33
5	15.71	0.359	90.43	14.61	2538.591	14.67
6	21.09	0.454	84.38	13.63	3364.608	19.45
7	24.50	0.427	136.00	21.98	4577.146	26.46
7	28.00	0.353	618.88	100.00	17301.101	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:08:16 AM
Printed by: wet


Ident: ICV
Analysis from: 5/5/2020 1:17:18 AM
File: _2020-05-05_

Last save: 5/5/2020 8:59:35 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25296

Last save: 5/5/2020 8:59:1

SAMPLE:
:
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.520	11.41	5.11	450.912	7.08
2	8.81	0.198	28.65	12.82	448.187	7.04
3	10.62	0.263	42.88	19.18	914.481	14.36
4	13.76	0.287	29.25	13.09	658.795	10.34
5	15.85	0.360	33.14	14.82	939.930	14.75
6	21.15	0.479	30.12	13.48	1288.850	20.23
7	24.54	0.437	48.09	21.51	1669.231	26.20
7	28.00	0.363	223.54	100.00	6370.385	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:08:23 AM
Printed by: wet


Ident: ICB
Analysis from: 5/5/2020 1:47:56 AM
File: _2020-05-05_

Last save: 5/5/2020 9:00:02 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25297

Last save: 5/5/2020 8:59:1

SAMPLE:
:
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:11:49 AM
Printed by: wet


Ident: CCV
Analysis from: 5/5/2020 2:18:35 AM
File: _2020-05-05_

Last save: 5/5/2020 8:59:36 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25298

Last save: 5/5/2020 8:59:1

SAMPLE:
: AM/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.484	11.88	5.31	448.913	7.02
2	8.83	0.199	28.64	12.81	453.300	7.09
3	10.65	0.260	43.20	19.33	923.946	14.45
4	13.80	0.289	29.04	12.99	663.918	10.38
5	15.92	0.364	32.80	14.68	943.499	14.76
6	21.14	0.477	30.16	13.50	1286.009	20.11
7	24.52	0.439	47.78	21.38	1674.163	26.18
7	28.00	0.359	223.50	100.00	6393.747	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:12:06 AM
Printed by: wet


Ident: CCB
Analysis from: 5/5/2020 2:49:13 AM
File: _2020-05-05_

Last save: 5/5/2020 8:59:36 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25299

Last save: 5/5/2020 8:59:1

SAMPLE:
:
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:12:13 AM
Printed by: wet


Ident: LB108995BLW
Analysis from: 5/5/2020 3:19:51 AM
File: _2020-05-05_

Last save: 5/5/2020 8:59:36 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25300

Last save: 5/5/2020 8:59:1

SAMPLE:
: AM/AP
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:12:22 AM
 Printed by: wet


Ident: LB108995BSW
 Analysis from: 5/5/2020 3:50:28 AM
 File: _2020-05-05_

Last save: 5/5/2020 8:59:36 AM

Method: AnionsIC2-050420.mtw
 Run operator: wet
 Analysis number: 25301

Last save: 5/5/2020 8:59:1

SAMPLE:
 :
 Vial number: 14
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.83	0.483	11.87	5.41	449.360	7.05
2	8.84	0.202	28.25	12.88	451.871	7.09
3	10.67	0.265	42.51	19.38	920.920	14.44
4	13.84	0.292	28.56	13.02	662.530	10.39
5	15.96	0.368	32.23	14.70	939.861	14.74
6	21.13	0.483	29.44	13.43	1284.526	20.15
7	24.52	0.444	46.44	21.18	1666.396	26.14
7	28.00	0.362	219.30	100.00	6375.464	100.00

This report has been created by IC Net
 METROHM LTD

Report date: 5/5/2020 11:12:29 AM
Printed by: wet


Ident: L2485-01
Analysis from: 5/5/2020 4:21:06 AM
File: _2020-05-05_

Last save: 5/5/2020 9:01:27 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25302

Last save: 5/5/2020 8:59:1

SAMPLE:
: AM/AP
Vial number: 15
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.85	0.192	0.57	28.39	8.185	14.77
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	16.17	0.367	1.11	55.68	33.861	61.09
6	0.00	0.000	0.00	0.00	0.000	0.00
7	24.58	0.513	0.31	15.65	13.381	24.14
7	28.00	0.153	1.99	99.72	55.427	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:12:34 AM
Printed by: wet


Ident: L2485-01DUP
Analysis from: 5/5/2020 4:51:44 AM
File: _2020-05-05_

Last save: 5/5/2020 8:59:36 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25303

Last save: 5/5/2020 8:59:1

SAMPLE:
: AM/AP
Vial number: 16
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.85	0.196	0.52	26.81	7.582	13.61
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	16.17	0.373	1.11	57.15	34.415	61.76
6	0.00	0.000	0.00	0.00	0.000	0.00
7	24.59	0.521	0.31	15.82	13.729	24.64
7	28.00	0.156	1.93	99.78	55.725	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:12:39 AM
 Printed by: wet


Ident: L2485-01MS
 Analysis from: 5/5/2020 5:22:22 AM
 File: _2020-05-05_

Last save: 5/5/2020 8:59:36 AM

Method: AnionsIC2-050420.mtw
 Run operator: wet
 Analysis number: 25304

Last save: 5/5/2020 8:59:1

SAMPLE:
 : AM/AP
 Vial number: 17
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.83	0.487	12.00	5.41	459.231	7.03
2	8.84	0.204	29.65	13.37	479.092	7.34
3	10.66	0.269	42.51	19.17	934.887	14.32
4	13.83	0.293	28.80	12.99	674.856	10.34
5	15.94	0.371	33.69	15.19	994.346	15.23
6	21.12	0.492	28.05	12.65	1264.506	19.37
7	24.50	0.449	47.01	21.20	1722.101	26.38
7	28.00	0.366	221.71	100.00	6529.019	100.00

This report has been created by IC Net
 METROHM LTD

Report date: 5/5/2020 11:12:45 AM
Printed by: wet


Ident: L2485-01MSD
Analysis from: 5/5/2020 5:53:00 AM
File: _2020-05-05_

Last save: 5/5/2020 8:59:36 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25305

Last save: 5/5/2020 8:59:1

SAMPLE:
:
Vial number: 18
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.470	12.35	5.76	456.989	7.17
2	8.84	0.206	28.24	13.16	460.582	7.22
3	10.67	0.276	40.38	18.81	907.329	14.23
4	13.84	0.295	28.01	13.05	660.310	10.35
5	15.95	0.373	32.73	15.25	968.196	15.18
6	21.13	0.495	27.36	12.75	1240.072	19.44
7	24.51	0.451	45.56	21.23	1684.284	26.41
7	28.00	0.367	214.62	100.00	6377.762	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:12:51 AM
 Printed by: wet


Ident: CCV
 Analysis from: 5/5/2020 6:23:42 AM
 File: _2020-05-05_

Last save: 5/5/2020 8:59:36 AM

Method: AnionsIC2-050420.mtw
 Run operator: wet
 Analysis number: 25306

Last save: 5/5/2020 8:59:1

SAMPLE:
 : AM/AP
 Vial number: 11
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.84	0.483	12.42	5.77	464.841	7.21
2	8.85	0.208	27.78	12.91	455.622	7.06
3	10.67	0.273	41.72	19.39	927.290	14.38
4	13.85	0.297	28.04	13.03	667.441	10.35
5	15.97	0.376	31.65	14.71	949.422	14.72
6	21.13	0.499	28.66	13.32	1309.785	20.31
7	24.51	0.454	44.87	20.86	1674.909	25.97
7	28.00	0.370	215.14	100.00	6449.310	100.00

This report has been created by IC Net
 METROHM LTD

Report date: 5/5/2020 11:12:59 AM
Printed by: wet


Ident: CCB
Analysis from: 5/5/2020 6:54:19 AM
File: _2020-05-05_

Last save: 5/5/2020 8:59:36 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25307

Last save: 5/5/2020 8:59:1

SAMPLE:
:
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD