

LB109060

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by :

Instrument ID : Konelab

5/7/2020 16:48

Test: Phenolics- mg/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.068	0.0	0.106	
ICB1	0.002	0.0	0.005	
CCV1	1.076	0.0	0.107	
CCB1	0.005	0.0	0.006	
PB128609BL	0.005	0.0	0.006	
PB128609BS	1.064	0.0	0.106	
L2182-12	0.026	0.0	0.008	
CCV2	1.060	0.0	0.105	
CCB2	0.002	0.0	0.005	

N	9
Mean	0.479
SD	0.5580
CV%	116.54

Aquakem v. 7.2AQ1

Results from time period:

Thu May 07 15:52:20 2020

Thu May 07 16:47:31 2020

h

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0527	mg/l	5/7/2020 15:52:20	
0.1PPM	A	Phenolics-	P	0.1034	mg/l	5/7/2020 15:52:21	
0.25PPM	A	Phenolics-	P	0.253	mg/l	5/7/2020 15:52:22	
0.50PPM	A	Phenolics-	P	0.5017	mg/l	5/7/2020 15:52:23	
1.0PPM	A	Phenolics-	P	0.9806	mg/l	5/7/2020 15:52:24	
2.0PPM	A	Phenolics-	P	2.0087	mg/l	5/7/2020 15:52:25	
ICV1	S	Phenolics-	P	1.0684	mg/l	5/7/2020 16:41:09	
ICB1	S	Phenolics-	P	0.0025	mg/l	5/7/2020 16:41:10	
CCV1	S	Phenolics-	P	1.0759	mg/l	5/7/2020 16:41:13	
CCB1	S	Phenolics-	P	0.0054	mg/l	5/7/2020 16:41:14	
PB128609BL	S	Phenolics-	P	0.0053	mg/l	5/7/2020 16:41:17	
PB128609BS	S	Phenolics-	P	1.0641	mg/l	5/7/2020 16:41:18	
L2182-12	S	Phenolics-	P	0.0261	mg/l	5/7/2020 16:47:27	
CCV2	S	Phenolics-	P	1.0596	mg/l	5/7/2020 16:47:29	
CCB2	S	Phenolics-	P	0.0021	mg/l	5/7/2020 16:47:30	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : DA

Instrument ID : Konelab

5/7/2020 15:53


Test Phenolics-

Accepted 5/7/2020 15:53

Factor 10.58
Bias 0.005

Coeff. of det. 0.999826

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0527	0.0500	
2	Phenol-2	0.015	0.1034	0.1000	
3	Phenol-2	0.029	0.2530	0.2500	
4	Phenol-2	0.052	0.5017	0.5000	
5	Phenol-2	0.098	0.9806	1.0000	
6	Phenol-2	0.195	2.0087	2.0000	