

Clear table


ID-2
300.0
5/18/2020
AM

ident	concentra tion F-	concentrati on CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-04_	5/4/20 17:07		AM/AP
STD2	0.2550	0.3710	0.4210	1.2620	0.3100	0.8170	1.9720	2020-05-04_	5/4/20 17:37		AM/AP
STD3	0.4160	0.6160	0.6430	2.0900	0.5220	1.2100	3.1800	2020-05-04_	5/4/20 22:44		AM/AP
STD4	0.7560	1.2020	1.1760	3.9660	1.0130	2.1950	5.9480	2020-05-04_	5/4/20 23:14		AM/AP
STD5	1.9550	2.8800	2.8060	9.5650	2.3730	4.0110	14.1750	2020-05-04_	5/4/20 23:45		AM/AP
STD6	3.9920	5.9530	5.9460	19.8480	4.9580	10.1380	29.7610	2020-05-05_	5/5/20 0:16		AM/AP
STD7	5.0270	7.5770	7.6070	25.2690	6.3240	12.6290	37.9640	2020-05-05_	5/5/20 0:46		AM/AP
ICV	1.8830	2.8570	2.8040	9.5470	2.3900	5.1040	14.1970	2020-05-05_	5/5/20 1:17		AM/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-05_	5/5/20 1:47		AM/AP
CCV	1.8960	2.9920	3.0770	9.9650	2.5090	6.2080	15.0830	2020-05-15_	5/15/20 9:38		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-15_	5/15/20 10:07		AM/AP
LB109192BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-15_	5/15/20 10:37		AM/AP
LB109192BSW	1.8970	2.9800	3.1670	9.9460	2.4850	6.0170	15.0120	2020-05-15_	5/15/20 11:06		AM/AP
L2676-01	0.3200	12.9750	0.0000	0.0000	0.0000	0.0000	16.2240	2020-05-15_	5/15/20 12:05		AM/AP
L2676-02	0.2210	86.0020	0.0000	0.0000	0.0000	0.0000	42.0270	2020-05-15_	5/15/20 12:35		AM/AP
L2676-03	0.0000	9.7990	0.0000	0.0000	0.0000	0.0000	55.9530	2020-05-15_	5/15/20 13:04		AM/AP
L2676-03DUP	0.0000	9.8120	0.0000	0.0000	0.0000	0.0000	55.8850	2020-05-15_	5/15/20 13:33		AM/AP
CCV	2.0250	2.9860	3.0490	10.0280	2.5040	6.6610	14.4280	2020-05-15_	5/15/20 16:03		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-15_	5/15/20 16:43		AM/AP
L2676-03MS	2.4640	11.9540	2.6730	9.6460	2.4260	10.5330	65.6850	2020-05-15_	5/15/20 17:13		AM/AP
L2676-03MSD	2.4930	11.9680	2.7130	9.6920	2.4290	10.2680	65.3450	2020-05-15_	5/15/20 17:43		AM/AP
CCV	1.9220	2.8720	2.9660	9.3530	2.3850	5.3100	13.5250	2020-05-15_	5/15/20 22:10		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-15_	5/15/20 22:40		AM/AP
CCV	1.9640	3.0150	2.9370	9.9030	2.4880	5.6510	13.8930	2020-05-16_	5/16/20 14:18		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-16_	5/16/20 14:49		AM/AP
L2649-04DLX4	0.0000	4.0340	0.0000	0.0000	0.0000	0.0000	8.7820	2020-05-16_	5/16/20 15:20		AM/AP
L2676-01DLX2	0.1490	5.2840	0.0000	0.0000	0.0000	0.0000	6.5210	2020-05-16_	5/16/20 15:51		AM/AP
L2676-02DLX2	0.0000	37.3090	0.0000	0.0000	0.0000	0.0000	19.6340	2020-05-16_	5/16/20 16:21		AM/AP
L2676-02DLX2X20	0.0000	3.3630	0.0000	0.0000	0.0000	0.0000	2.3270	2020-05-16_	5/16/20 16:52		AM/AP
L2676-03DLX2	0.0000	4.6010	0.0000	0.0000	0.0000	0.0000	25.6690	2020-05-16_	5/16/20 17:23		AM/AP
L2649-05DLX2	0.0000	2.1620	0.0000	0.0000	0.0000	0.0000	26.2800	2020-05-16_	5/16/20 18:55		AM/AP
L2649-01DLX10	0.0000	5.6150	0.0000	0.0000	0.0000	0.0000	7.3540	2020-05-16_	5/16/20 19:25		AM/AP
L2649-08DLX5	0.0000	4.7170	0.0000	0.0000	0.0000	0.0000	9.9320	2020-05-16_	5/16/20 19:56		AM/AP
CCV	2.0080	3.0100	2.9780	9.9010	2.4990	5.9020	13.8190	2020-05-16_	5/16/20 20:27		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-16_	5/16/20 20:57		AM/AP
L2649-02DLX4	0.0000	4.5510	0.0000	0.0000	0.0000	0.0000	11.3470	2020-05-16_	5/16/20 21:28		AM/AP
L2649-03DLX4	0.0000	4.1900	0.0000	0.0000	0.0000	0.0000	10.5280	2020-05-16_	5/16/20 21:59		AM/AP
L2649-06DLX20	0.0000	4.8150	0.0000	0.0000	0.0000	0.0000	1.5280	2020-05-16_	5/16/20 22:29		AM/AP
CCV	1.9220	2.9460	2.9010	9.6640	2.4410	5.5400	13.5760	2020-05-16_	5/16/20 23:31		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-05-17_	5/17/20 0:01		AM/AP

CALIBRATION OF COMPONENT F-

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.0799146 \cdot A + 1.62322$
 RSD: 2.162 %
 Correlation coefficient: 0.999802

IC-2
 M300.0
 5/05/2020
 A.M




K3 = 0 K2 = 0 K1 = 0.0799146 K0 = 1.62322
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	1.127	43.45	0.2	20	5.75		
3	2.153	83.77	0.4	20	5.75		
4	4.324	168.9	0.8	20	5.75		
5	18.49	469	2	20	5.75		
6	27.93	978.7	4	20	5.75		
7	34.32	1238	5	20	5.75		

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.124004 \cdot A + 1.56524$
 RSD: 2.684 %
 Correlation coefficient: 0.999696




K3 = 0 K2 = 0 K1 = 0.124004 K0 = 1.56524
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0		0	0	0		
2	3.173		47.2	0.3	20		8.834
3	5.534		86.75	0.6	20		8.834
4	11.59		181.3	1.2	20		8.834
5	29		451.9	3	20		8.834
6	61.82		947.6	6	20		8.834
7	79.33		1209	7.5	20		8.834

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.0574611 \cdot A + 3.53486$
 RSD: 4.235 %
 Correlation coefficient: 0.999242




K3 = 0 K2 = 0 K1 = 0.0574611 K0 = 3.53486
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	4.596	85.1	0.3	20	10.73	10.73
3	8.527	162.4	0.6	20	10.73	10.73
4	17.64	347.8	1.2	20	10.73	10.73
5	43.8	915.3	3	20	10.73	10.73
6	88.99	2008	6	20	10.73	10.73
7	111	2586	7.5	20	10.73	10.73

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.278566 \cdot A + 7.42532$
 RSD: 2.915 %
 Correlation coefficient: 0.999641




K3 = 0 K2 = 0 K1 = 0.278566 K0 = 7.42532
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	2.93	63.97	1	20			14
3	5.491	123.4	2	20			14
4	11.56	258.1	4	20			14
5	29.86	660.1	10	20			14
6	64.53	1398	20	20			14
7	83.39	1788	25	20			14

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.0492074 \cdot A + 1.5542$
 RSD: 3.213 %
 Correlation coefficient: 0.999564




K3 = 0 K2 = 0 K1 = 0.0492074 K0 = 1.5542
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		
2	3.653	94.48	0.25	20	16.26		
3	6.565	180.7	0.5	20	16.26		
4	13.88	380.1	1	20	16.26		
5	33.89	932.9	2.5	20	16.26		
6	70.99	1984	5	20	16.26		
7	90.43	2539	6.25	20	16.26		

EW

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.0725012 \cdot A + 8.64363$
 RSD: 10.585 %
 Correlation coefficient: 0.995254




K3 = 0 K2 = 0 K1 = 0.0725012 K0 = 8.64363
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		
2	2.618	106.3	0.5	20	21.2		
3	5.061	214.5	1	20	21.2		
4	11.37	486.3	2	20	21.2		
5	19.79	987.3	5	20	21.2		
6	66.24	2677	10	20	21.2		
7	84.38	3365	12.5	20	21.2		

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-050420.mtw
 Equation: $Q = 0.163467 \cdot A + 11.0703$
 RSD: 3.551 %
 Correlation coefficient: 0.999467



K3 = 0 K2 = 0 K1 = 0.163467 K0 = 11.0703
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0		0	0	0		
2	5.05	173.6	1.5	20	24.57		
3	9.227	321.3	3	20	24.57		
4	19.16	660	6	20	24.57		
5	48.9	1667	15	20	24.57		
6	105.7	3574	30	20	24.57		
7	136	4577	37.5	20	24.57		

Report date: 5/5/2020 11:07:26 AM
Printed by: wet


Ident: STD1
Analysis from: 5/4/2020 5:07:20 PM
File: _2020-05-04_

Last save: 5/5/2020 8:57:48 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25286

Last save: 5/5/2020 12:44:

SAMPLE:
:
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:07:36 AM
Printed by: wet


Ident: STD2
Analysis from: 5/4/2020 5:37:56 PM
File: _2020-05-04_

Last save: 5/5/2020 8:57:48 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25287

Last save: 5/5/2020 12:44:

SAMPLE:
:
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.79	0.651	1.12	4.86	43.450	7.08
2	8.76	0.201	3.17	13.71	47.200	7.69
3	10.57	0.248	4.60	19.86	85.100	13.86
4	13.64	0.299	2.93	12.65	63.970	10.42
5	15.75	0.354	3.65	15.77	94.482	15.39
6	21.23	0.472	2.62	11.31	106.275	17.31
7	24.62	0.441	5.05	21.81	173.605	28.27
7	28.00	0.381	23.14	99.97	614.082	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:07:43 AM
Printed by: wet


Ident: STD3
Analysis from: 5/4/2020 10:44:08 PM
File: _2020-05-04_

Last save: 5/5/2020 8:57:48 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25291

Last save: 5/5/2020 12:44:

SAMPLE:
: AM/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.532	2.15	5.06	83.771	7.14
2	8.79	0.200	5.53	13.00	86.746	7.40
3	10.61	0.243	8.53	20.04	162.449	13.85
4	13.71	0.292	5.49	12.90	123.426	10.52
5	15.84	0.351	6.56	15.42	180.651	15.40
6	21.22	0.481	5.06	11.89	214.523	18.29
7	24.61	0.443	9.23	21.68	321.316	27.40
7	28.00	0.363	42.56	99.99	1172.883	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:07:49 AM
Printed by: wet


Ident: STD4
Analysis from: 5/4/2020 11:14:46 PM
File: _2020-05-04_

Last save: 5/5/2020 8:57:48 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25292

Last save: 5/5/2020 12:44:

SAMPLE:
: AM/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.497	4.32	4.83	168.896	6.80
2	8.79	0.199	11.59	12.94	181.283	7.30
3	10.60	0.246	17.64	19.70	347.758	14.01
4	13.69	0.288	11.56	12.92	258.058	10.40
5	15.80	0.351	13.88	15.51	380.070	15.31
6	21.19	0.482	11.37	12.70	486.269	19.59
7	24.57	0.438	19.16	21.40	660.032	26.59
7	28.00	0.358	89.53	100.00	2482.366	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:07:55 AM
Printed by: wet


Ident: STD5
Analysis from: 5/4/2020 11:45:24 PM
File: _2020-05-04_

Last save: 5/5/2020 8:57:48 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25293

Last save: 5/5/2020 12:44:

SAMPLE:
: AM/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.74	0.308	18.49	8.26	468.959	7.71
2	8.79	0.198	29.00	12.96	451.931	7.43
3	10.58	0.260	43.80	19.58	915.278	15.05
4	13.69	0.283	29.86	13.35	660.055	10.85
5	15.77	0.355	33.89	15.15	932.926	15.34
6	21.19	0.550	19.79	8.85	987.266	16.23
7	24.56	0.433	48.90	21.86	1666.522	27.40
7	28.00	0.341	223.72	100.00	6082.938	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:08:02 AM
Printed by: wet


Ident: STD6
Analysis from: 5/5/2020 12:16:02 AM
File: _2020-05-05_

Last save: 5/5/2020 8:57:48 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25294

Last save: 5/5/2020 12:44:

SAMPLE:
:
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.83	0.464	27.92	5.74	978.680	7.21
2	8.80	0.195	61.82	12.71	947.582	6.98
3	10.57	0.276	88.99	18.30	2008.139	14.80
4	13.69	0.276	64.53	13.27	1398.379	10.31
5	15.73	0.359	70.99	14.60	1983.638	14.62
6	21.14	0.460	66.24	13.62	2677.320	19.73
7	24.55	0.429	105.74	21.75	3573.508	26.34
7	28.00	0.351	486.23	100.00	13567.246	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:08:08 AM
Printed by: wet


Ident: STD7
Analysis from: 5/5/2020 12:46:40 AM
File: _2020-05-05_

Last save: 5/5/2020 8:59:35 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25295

Last save: 5/5/2020 8:59:1

SAMPLE:
:
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.83	0.475	34.32	5.55	1237.684	7.15
2	8.79	0.194	79.33	12.82	1209.415	6.99
3	10.56	0.286	111.03	17.94	2586.114	14.95
4	13.69	0.272	83.39	13.47	1787.544	10.33
5	15.71	0.359	90.43	14.61	2538.591	14.67
6	21.09	0.454	84.38	13.63	3364.608	19.45
7	24.50	0.427	136.00	21.98	4577.146	26.46
7	28.00	0.353	618.88	100.00	17301.101	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:08:16 AM
Printed by: wet


Ident: ICV
Analysis from: 5/5/2020 1:17:18 AM
File: _2020-05-05_

Last save: 5/5/2020 8:59:35 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25296

Last save: 5/5/2020 8:59:1

SAMPLE:
:
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.520	11.41	5.11	450.912	7.08
2	8.81	0.198	28.65	12.82	448.187	7.04
3	10.62	0.263	42.88	19.18	914.481	14.36
4	13.76	0.287	29.25	13.09	658.795	10.34
5	15.85	0.360	33.14	14.82	939.930	14.75
6	21.15	0.479	30.12	13.48	1288.850	20.23
7	24.54	0.437	48.09	21.51	1669.231	26.20
7	28.00	0.363	223.54	100.00	6370.385	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/5/2020 11:08:23 AM
Printed by: wet


Ident: ICB
Analysis from: 5/5/2020 1:47:56 AM
File: _2020-05-05_

Last save: 5/5/2020 9:00:02 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25297

Last save: 5/5/2020 8:59:1

SAMPLE:
:
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/15/2020 2:53:52 PM
Printed by: wet


Ident: CCV
Analysis from: 5/15/2020 9:38:05 AM
File: _2020-05-15_

Last save: 5/15/2020 10:04:53 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25606

Last save: 5/15/2020 10:03

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.543	10.83	5.59	454.168	6.51
2	8.80	0.291	24.44	12.62	469.998	6.73
3	10.64	0.367	41.06	21.20	1009.640	14.46
4	13.77	0.447	24.15	12.47	688.794	9.87
5	15.97	0.550	27.66	14.28	988.017	14.15
6	20.87	0.906	25.56	13.20	1593.215	22.82
7	23.99	0.740	39.97	20.64	1777.622	25.46
7	26.79	0.549	193.67	100.00	6981.455	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/15/2020 2:53:59 PM
Printed by: wet


Ident: CCB
Analysis from: 5/15/2020 10:07:30 AM
File: _2020-05-15_

Last save: 5/15/2020 10:34:30 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25607

Last save: 5/15/2020 10:03

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/15/2020 2:54:07 PM
Printed by: wet


Ident: LB109192BLW
Analysis from: 5/15/2020 10:37:07 AM
File: _2020-05-15_

Last save: 5/15/2020 2:51:04 PM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25608

Last save: 5/15/2020 10:03

SAMPLE:
:
Vial number: 4
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/15/2020 2:54:14 PM
Printed by: wet


Ident: LB109192BSW
Analysis from: 5/15/2020 11:06:46 AM
File: _2020-05-15_

Last save: 5/15/2020 2:51:04 PM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25609

Last save: 5/15/2020 10:03

SAMPLE:
: AM/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.542	10.85	5.65	454.377	6.55
2	8.75	0.302	23.67	12.33	468.011	6.75
3	10.59	0.399	39.28	20.46	1040.696	15.00
4	13.68	0.450	24.34	12.68	687.423	9.91
5	15.78	0.562	27.26	14.20	978.598	14.10
6	20.87	0.906	25.54	13.31	1540.528	22.20
7	24.02	0.650	41.02	21.37	1769.004	25.49
7	26.50	0.544	191.95	100.00	6938.637	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/15/2020 2:55:01 PM
Printed by: wet


Ident: L2676-01
Analysis from: 5/15/2020 12:05:58 PM
File: _2020-05-15_

Last save: 5/15/2020 2:54:38 PM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25611

Last save: 5/15/2020 11:33

SAMPLE:
: AM/AP
Vial number: 45
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.81	0.734	1.05	0.74	59.876	1.48
2	8.72	0.325	95.25	67.50	2080.010	51.27
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.93	0.532	44.82	31.76	1917.276	47.26
7	26.50	0.227	141.12	100.00	4057.162	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/15/2020 2:56:12 PM
Printed by: wet


Ident: L2676-02
Analysis from: 5/15/2020 12:35:06 PM
File: _2020-05-15_

Last save: 5/15/2020 2:56:12 PM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25612

Last save: 5/15/2020 11:33

SAMPLE:
: AM/AP
Vial number: 46
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.78	0.676	0.72	0.08	35.089	0.18
2	8.70	0.297	725.06	85.24	13858.135	73.06
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.83	0.484	124.72	14.66	5074.274	26.75
7	26.50	0.208	850.51	99.98	18967.498	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/15/2020 2:56:41 PM
Printed by: wet


Ident: L2676-03
Analysis from: 5/15/2020 1:04:15 PM
File: _2020-05-15_

Last save: 5/15/2020 2:56:25 PM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25613

Last save: 5/15/2020 11:33

SAMPLE:
: AM/AP
Vial number: 47
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.72	0.339	71.08	29.82	1567.810	18.77
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.95	0.483	166.83	69.98	6778.085	81.15
7	26.50	0.117	237.91	99.80	8345.895	99.92

This report has been created by IC Net
METROHM LTD

Report date: 5/15/2020 3:00:14 PM
Printed by: wet


Ident: L2676-03DUP
Analysis from: 5/15/2020 1:33:24 PM
File: _2020-05-15_

Last save: 5/15/2020 3:00:14 PM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25614

Last save: 5/15/2020 11:33

SAMPLE:
: AM/AP
Vial number: 48
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.70	0.342	71.01	29.72	1569.978	18.83
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.96	0.477	167.92	70.28	6769.704	81.17
7	26.50	0.117	238.93	99.99	8339.682	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 8:42:33 AM
Printed by: wet


Ident: CCV
Analysis from: 5/15/2020 4:03:35 PM
File: _2020-05-15_

Last save: 5/15/2020 4:35:16 PM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25619

Last save: 5/15/2020 2:58:

SAMPLE:
: AM/AP
Vial number: 34
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.593	11.09	5.19	486.565	6.89
2	8.57	0.237	24.91	11.65	468.898	6.64
3	10.28	0.300	40.40	18.89	999.596	14.15
4	13.20	0.300	28.04	13.11	693.312	9.82
5	15.15	0.364	32.59	15.24	986.037	13.96
6	20.73	0.592	32.27	15.09	1718.399	24.33
7	23.91	0.468	43.93	20.54	1697.546	24.04
7	27.00	0.408	213.24	99.72	7050.352	99.84

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 8:43:13 AM
Printed by: wet


Ident: CCB
Analysis from: 5/15/2020 4:43:50 PM
File: _2020-05-15_

Last save: 5/18/2020 8:43:08 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25621

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 1
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 8:45:45 AM
Printed by: wet


Ident: L2676-03MS
Analysis from: 5/15/2020 5:13:58 PM
File: _2020-05-15_

Last save: 5/18/2020 8:45:39 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25622

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 49
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.81	0.636	12.96	2.84	596.416	3.79
2	8.56	0.269	97.07	21.26	1915.328	12.16
3	10.27	0.595	24.96	5.47	868.739	5.51
4	13.20	0.273	29.13	6.38	665.901	4.23
5	15.12	0.344	33.14	7.26	954.568	6.06
6	20.81	0.731	45.51	9.97	2786.306	17.68
7	23.88	0.453	213.73	46.82	7968.777	50.58
7	27.00	0.472	456.50	100.00	15756.036	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 8:46:07 AM
Printed by: wet


Ident: L2676-03MSD
Analysis from: 5/15/2020 5:43:38 PM
File: _2020-05-15_

Last save: 5/18/2020 8:45:57 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25623

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 50
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.625	13.32	2.95	603.534	3.85
2	8.55	0.272	96.36	21.33	1917.599	12.24
3	10.25	0.640	24.02	5.32	882.825	5.63
4	13.17	0.273	29.32	6.49	669.190	4.27
5	15.08	0.343	33.21	7.35	955.472	6.10
6	20.83	0.768	43.09	9.54	2713.365	17.32
7	23.89	0.454	212.42	47.02	7927.212	50.59
7	27.00	0.482	451.76	100.00	15669.197	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 11:59:40 AM
Printed by: wet


Ident: CCV
Analysis from: 5/16/2020 2:18:55 PM
File: _2020-05-16_

Last save: 5/18/2020 11:56:36 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25664

Last save: 5/16/2020 8:08:

SAMPLE:
: AM/AP
Vial number: 120
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.75	0.659	9.90	4.63	471.202	7.09
2	8.57	0.219	26.09	12.19	473.700	7.13
3	10.29	0.264	41.98	19.62	960.808	14.47
4	13.24	0.285	28.83	13.47	684.367	10.30
5	15.22	0.345	33.86	15.82	979.811	14.75
6	20.61	0.587	30.78	14.38	1439.546	21.68
7	23.78	0.514	42.56	19.89	1632.031	24.57
7	28.00	0.410	214.00	100.00	6641.466	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 11:59:48 AM
Printed by: wet


Ident: CCB
Analysis from: 5/16/2020 2:49:42 PM
File: _2020-05-16_

Last save: 5/18/2020 8:43:42 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25665

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 35
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:20:07 PM
Printed by: wet


Ident: L2649-04DLX4
Analysis from: 5/16/2020 3:20:30 PM
File: _2020-05-16_

Last save: 5/18/2020 11:16:32 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25666

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 91
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.57	0.219	35.30	57.31	638.007	38.79
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.82	0.519	26.28	42.68	1006.688	61.21
7	28.00	0.105	61.58	99.99	1644.695	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:23:26 PM
Printed by: wet


Ident: L2676-01DLX2
Analysis from: 5/16/2020 3:51:08 PM
File: _2020-05-16_

Last save: 5/18/2020 11:16:32 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25667

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 92
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.74	0.728	0.34	0.53	16.903	1.07
2	8.55	0.229	45.09	70.43	839.631	52.92
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.82	0.516	18.58	29.03	730.083	46.02
7	28.00	0.210	64.01	99.99	1586.617	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:23:51 PM
Printed by: wet

Ident: L2676-02DLX2
Analysis from: 5/16/2020 4:21:49 PM
File: _2020-05-16_

Last save: 5/18/2020 11:16:32 AM


Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25668

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 93
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

Before

*Software did not
integrate
whole
peak*



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.53	0.201	363.39	86.69	6004.812	73.22
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.71	0.506	55.77	13.30	2196.403	26.78
7	28.00	0.101	419.16	100.00	8201.215	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:24:18 PM
Printed by: wet

Ident: L2676-02DLX2
Analysis from: 5/16/2020 4:21:49 PM
File: 2020-05-16_
Modified! Manual peaks!
Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25668


Last save: 5/18/2020 11:16:32 AM

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 93
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

After

5/18/20



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.53	0.201	363.39	86.65	6004.812	72.01
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.72	0.507	55.96	13.34	2334.466	27.99
7	28.00	0.101	419.35	100.00	8339.278	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:24:33 PM
Printed by: wet


Ident: L2676-02DL2X20
Analysis from: 5/16/2020 4:52:27 PM
File: _2020-05-16_

Last save: 5/18/2020 11:16:33 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25669

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 94
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.55	0.216	29.55	84.25	529.765	70.94
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.84	0.530	5.52	15.74	217.019	29.06
7	28.00	0.107	35.07	99.99	746.785	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:25:10 PM
Printed by: wet

Ident: L2676-03DLX2
Analysis from: 5/16/2020 5:23:05 PM
File: _2020-05-16_

Last save: 5/18/2020 12:25:10 PM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25670


Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 95
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

Before

*Software did not
integrate
whole
peak*

*Am
5/18/20*



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.53	0.231	38.85	34.31	729.521	20.39
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.80	0.499	74.38	65.69	2849.189	79.61
7	28.00	0.104	113.23	100.00	3578.710	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:25:24 PM
Printed by: wet


Ident: L2676-03DLX2
Analysis from: 5/16/2020 5:23:05 PM
File: 2020-05-16_
Modified! Manual peaks!
Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25670

Last save: 5/18/2020 12:25:10 PM

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 95
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

After
AM
5/18/20



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.53	0.231	38.85	34.22	729.521	19.19
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.80	0.501	74.69	65.78	3072.838	80.81
7	28.00	0.105	113.54	100.00	3802.359	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:35:10 PM
Printed by: wet

Ident: L2649-05DLX2
Analysis from: 5/16/2020 6:55:07 PM
File: _2020-05-16_

Last save: 5/18/2020 11:16:33 AM


Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25673

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 36
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

Before
Software did not
integrate whole peak

AM
5/18/20



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.51	0.224	18.25	19.33	336.064	10.36
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.78	0.500	76.17	80.67	2907.662	89.64
7	28.00	0.103	94.42	100.00	3243.725	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:35:23 PM
Printed by: wet


Ident: L2649-05DLX2
Analysis from: 5/16/2020 6:55:07 PM
File: 2020-05-16_
Modified! Manual peaks!
Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25673

Last save: 5/18/2020 11:16:33 AM

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 36
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

After 5/18/20
AM



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.51	0.224	18.25	19.26	336.064	9.65
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.78	0.502	76.50	80.74	3147.560	90.35
7	28.00	0.104	94.75	100.00	3483.624	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:38:08 PM
Printed by: wet

Ident: L2649-01-DLX10
Analysis from: 5/16/2020 7:25:45 PM
File: _2020-05-16_


Last save: 5/18/2020 11:16:33 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25674

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 37
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

*Before
software did not
integrate whole
peak
AM
5/18/20*



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.51	0.209	51.39	72.69	893.051	54.27
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.82	0.524	19.30	27.30	752.514	45.73
7	28.00	0.105	70.69	100.00	1645.565	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:38:20 PM
Printed by: wet


Ident: L2649-01-DLX10
Analysis from: 5/16/2020 7:25:45 PM
File: 2020-05-16_
Modified! Manual peaks!
Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25674

Last save: 5/18/2020 11:16:33 AM

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 37
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

After AM
5/18/20



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.51	0.209	51.39	72.58	893.051	51.77
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.82	0.526	19.41	27.42	832.065	48.23
7	28.00	0.105	70.80	100.00	1725.116	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:40:19 PM
Printed by: wet

Ident: L2649-08-DLX5
Analysis from: 5/16/2020 7:56:24 PM
File: _2020-05-16_

Last save: 5/18/2020 12:40:18 PM


Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25675

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 38
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

Before
Software did not
integrate whole
peak

AM
5/18/20



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.52	0.215	41.90	60.40	748.233	42.12
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.84	0.518	27.47	39.60	1028.080	57.88
7	28.00	0.105	69.37	100.00	1776.313	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:40:38 PM
Printed by: wet


Ident: L2649-08-DLX5
Analysis from: 5/16/2020 7:56:24 PM
File: 2020-05-16_
Modified! Manual peaks!
Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25675

Last save: 5/18/2020 12:40:18 PM

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 38
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

After
AM
5/18/20



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.52	0.215	41.90	60.27	748.233	39.47
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.84	0.521	27.63	39.73	1147.502	60.53
7	28.00	0.105	69.53	100.00	1895.736	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:42:57 PM
Printed by: wet


Ident: CCV
Analysis from: 5/16/2020 8:27:04 PM
File: _2020-05-16_

Last save: 5/18/2020 12:42:56 PM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25676

Last save: 5/16/2020 8:08:

SAMPLE:
: AM/AP
Vial number: 34
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.74	0.646	10.33	4.71	482.195	7.16
2	8.51	0.215	26.47	12.07	472.919	7.03
3	10.18	0.260	43.48	19.82	975.183	14.49
4	13.04	0.278	29.89	13.62	684.226	10.17
5	14.95	0.334	35.25	16.07	984.269	14.62
6	20.62	0.591	31.93	14.55	1508.862	22.42
7	23.82	0.519	42.04	19.16	1622.991	24.11
7	28.00	0.406	219.40	100.00	6730.646	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:43:10 PM
Printed by: wet


Ident: CCB
Analysis from: 5/16/2020 8:57:41 PM
File: _2020-05-16_

Last save: 5/18/2020 8:43:42 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25677

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 35
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:49:25 PM
Printed by: wet

Ident: L2649-02-DLX4
Analysis from: 5/16/2020 9:28:21 PM
File: _2020-05-16_

Last save: 5/18/2020 11:16:33 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25678


Last save: 5/15/2020 4:35:

SAMPLE:
:
Vial number: 39
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

AM/AP

Before

Software did not
integrate whole
peak
AM
5/18/2020



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.51	0.222	39.46	56.21	721.382	37.79
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.85	0.521	30.74	43.79	1187.703	62.21
7	28.00	0.106	70.20	100.00	1909.085	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:49:42 PM
Printed by: wet

Ident: L2649-02-DLX4
Analysis from: 5/16/2020 9:28:21 PM
File: 2020-05-16_
Modified! Manual peaks!
Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25678


Last save: 5/18/2020 11:16:33 AM

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 39
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

After

AM
5/18/20



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.51	0.222	39.46	56.07	721.382	35.33
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.85	0.523	30.92	43.93	1320.589	64.67
7	28.00	0.106	70.38	100.00	2041.971	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:50:08 PM
Printed by: wet

Ident: L2649-03DLX4
Analysis from: 5/16/2020 9:59:08 PM
File: _2020-05-16_


Last save: 5/18/2020 11:16:33 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25679

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 89
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

Before software did not
integrate whole
peak
AM
5/18/20



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.48	0.216	37.12	56.44	663.084	37.21
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.75	0.515	28.65	43.56	1119.066	62.79
7	28.00	0.105	65.76	100.00	1782.151	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:50:29 PM
Printed by: wet

Ident: L2649-03DLX4
Analysis from: 5/16/2020 9:59:08 PM
File: _2020-05-16_
Modified! Manual peaks!
Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25679


Last save: 5/18/2020 11:16:33 AM

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 89
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

After

AM
5/18/20



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.48	0.216	37.12	56.32	663.084	35.21
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.75	0.518	28.78	43.68	1220.333	64.79
7	28.00	0.105	65.90	100.00	1883.418	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:51:17 PM
Printed by: wet


Ident: L2649-06DLX20
Analysis from: 5/16/2020 10:29:46 PM
File: _2020-05-16_

Last save: 5/18/2020 12:51:14 PM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25680

Last save: 5/15/2020 4:35:

SAMPLE:
: AM/AP
Vial number: 90
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.50	0.209	44.23	80.36	763.934	68.60
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.83	0.539	3.05	5.55	119.199	10.70
7	28.00	0.107	47.28	85.91	883.132	79.31

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:52:24 PM
Printed by: wet


Ident: CCV
Analysis from: 5/16/2020 11:31:07 PM
File: _2020-05-16_

Last save: 5/18/2020 12:52:23 PM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25682

Last save: 5/16/2020 8:08:

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.73	0.641	9.97	4.68	460.764	7.09
2	8.50	0.214	25.95	12.17	462.499	7.11
3	10.18	0.258	42.85	20.09	948.186	14.58
4	13.04	0.278	29.24	13.71	667.155	10.26
5	14.95	0.334	34.39	16.13	960.420	14.77
6	20.60	0.585	30.13	14.13	1409.065	21.67
7	23.80	0.524	40.68	19.08	1593.234	24.51
7	28.00	0.405	213.22	100.00	6501.324	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/18/2020 12:52:32 PM
Printed by: wet


Ident: CCB
Analysis from: 5/17/2020 12:01:45 AM
File: _2020-05-17_

Last save: 5/18/2020 8:43:42 AM

Method: AnionsIC2-050420.mtw
Run operator: wet
Analysis number: 25683

Last save: 5/15/2020 4:35:

SAMPLE:
:
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD