

L13109346

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : *JS*

Instrument ID : Konelab

5/28/2020 12:48

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.006	0.0	0.101	
ICB1	-0.009	0.0	0.004	
CCV1	1.008	0.0	0.101	
CCB1	-0.011	0.0	0.004	
PB129102BL	-0.009	0.0	0.004	
PB129102BS	1.002	0.0	0.101	
L2747-01	-0.008	0.0	0.004	
L2747-02	-0.005	0.0	0.005	
L2747-02DUP	-0.007	0.0	0.005	
L2747-03	-0.007	0.0	0.005	
L2747-04	0.002	0.0	0.005	
L2747-05	-0.005	0.0	0.005	
L2747-06	0.005	0.0	0.006	
L2747-07	-0.007	0.0	0.004	
CCV2	1.002	0.0	0.101	
CCB2	-0.012	0.0	0.004	
L2747-08	-0.007	0.0	0.005	
L2747-09MS	1.009	0.0	0.101	
L2747-10MSD	1.012	0.0	0.102	
CCV3	1.011	0.0	0.102	
CCB3	-0.009	0.0	0.004	

N	21
Mean	0.331
SD	0.4896
CV%	147.76

Aquakem v. 7.2AQ1

Results from time period:

Thu May 28 10:37:42 2020

Thu May 28 12:45:42 2020

Dh

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0494	mg/l	5/28/2020 11:07:55	
0.1PPM	A	Phenolics-	P	0.1018	mg/l	5/28/2020 11:07:56	
0.25PPM	A	Phenolics-	P	0.2503	mg/l	5/28/2020 11:07:57	
0.50PPM	A	Phenolics-	P	0.5037	mg/l	5/28/2020 11:07:58	
1.0PPM	A	Phenolics-	P	0.9917	mg/l	5/28/2020 11:07:59	
2.0PPM	A	Phenolics-	P	2.0031	mg/l	5/28/2020 11:08:00	
ICV1	S	Phenolics-	P	1.0059	mg/l	5/28/2020 12:28:48	
ICB1	S	Phenolics-	P	-0.0095	mg/l	5/28/2020 12:28:49	
CCV1	S	Phenolics-	P	1.0081	mg/l	5/28/2020 12:28:52	
CCB1	S	Phenolics-	P	-0.0112	mg/l	5/28/2020 12:28:53	
PB129102BL	S	Phenolics-	P	-0.0087	mg/l	5/28/2020 12:28:55	
PB129102BS	S	Phenolics-	P	1.0019	mg/l	5/28/2020 12:28:56	
L2747-01	S	Phenolics-	P	-0.0075	mg/l	5/28/2020 12:28:58	
L2747-02	S	Phenolics-	P	-0.0054	mg/l	5/28/2020 12:38:49	
L2747-02DUP	S	Phenolics-	P	-0.0069	mg/l	5/28/2020 12:38:50	
L2747-03	S	Phenolics-	P	-0.0071	mg/l	5/28/2020 12:38:51	
L2747-04	S	Phenolics-	P	0.0017	mg/l	5/28/2020 12:38:52	
L2747-05	S	Phenolics-	P	-0.0045	mg/l	5/28/2020 12:38:53	
L2747-06	S	Phenolics-	P	0.0045	mg/l	5/28/2020 12:38:54	
L2747-07	S	Phenolics-	P	-0.0075	mg/l	5/28/2020 12:38:55	
CCV2	S	Phenolics-	P	1.0024	mg/l	5/28/2020 12:38:57	
CCB2	S	Phenolics-	P	-0.0121	mg/l	5/28/2020 12:38:58	
L2747-08	S	Phenolics-	P	-0.0072	mg/l	5/28/2020 12:39:00	
L2747-09MS	S	Phenolics-	P	1.0085	mg/l	5/28/2020 12:45:36	
L2747-10MSD	S	Phenolics-	P	1.0119	mg/l	5/28/2020 12:45:38	
CCV3	S	Phenolics-	P	1.0111	mg/l	5/28/2020 12:45:40	
CCB3	S	Phenolics-	P	-0.0094	mg/l	5/28/2020 12:45:42	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : DK

Instrument ID : Konelab

5/28/2020 11:12


Test Phenolics-

Accepted 5/28/2020 11:12

Factor 10.48
Bias 0.005

Coeff. of det. 0.999966

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0494	0.0500	
2	Phenol-2	0.015	0.1018	0.1000	
3	Phenol-2	0.029	0.2503	0.2500	
4	Phenol-2	0.053	0.5037	0.5000	
5	Phenol-2	0.100	0.9917	1.0000	
6	Phenol-2	0.196	2.0031	2.0000	