

TitriSoft Report

Ident	Amount	DateTime	Alkalinity	Final pH (sample)	Initial pH (sample)	slope	offset
LB109516BLW	100	11/06/2020 15:53	0.00	3.22	3.25	102.3	7.17
LB109516BSW	100	11/06/2020 16:05	44.12	4.30	8.42	102.3	7.17
L2920-01	50	11/06/2020 16:12	0.00	4.43	4.42	102.3	7.17
L2920-01DUP	50	11/06/2020 16:14	0.00	4.41	4.39	102.3	7.17
L2920-03	50	11/06/2020 16:17	149.28	4.35	5.92	102.3	7.17
L2920-04	50	11/06/2020 16:21	3.36	4.41	4.52	102.3	7.17
L2920-05	50	11/06/2020 16:24	180.96	4.20	6.37	102.3	7.17
L2920-06	50	11/06/2020 16:29	814.40	4.41	6.27	102.3	7.17
L2920-03DUP	50	11/06/2020 16:37	148.48	4.29	6.21	102.3	7.17
L2920-07	50	11/06/2020 16:41	145.76	4.39	6.51	102.3	7.17
L2920-08	50	11/06/2020 16:46	154.32	4.39	6.58	102.3	7.17
L2920-09	50	11/06/2020 16:50	1043.44	4.42	6.72	102.3	7.17
L2920-10	50	11/06/2020 16:59	65.12	4.41	5.60	102.3	7.17
L2920-11	50	11/06/2020 17:03	294.96	4.37	6.27	102.3	7.17
L2920-12	50	11/06/2020 17:08	415.84	4.40	6.87	102.3	7.17
L2949-01	50	11/06/2020 17:17	107.28	4.29	5.73	102.3	7.17
L2949-02	50	11/06/2020 17:21	386.48	4.38	6.44	102.3	7.17
L2949-03	50	11/06/2020 17:26	398.40	4.27	6.46	102.3	7.17
L2949-04	50	11/06/2020 17:32	470.80	4.34	6.70	102.3	7.17

06/11/2020
AM