

CB 109681

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : Sh

Instrument ID : Konelab

6/22/2020 10:58

Test: ^{Reactive}~~Total~~ CN

Mg/lc

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.488	0.0	0.077	
ICB1	-0.837	0.0	0.001	
CCV1	244.141	0.0	0.193	
CCB1	-0.181	0.0	0.002	
PB129646BL	-0.951	0.0	0.001	
L3051-02	-1.168	0.0	0.001	
L3051-02DUP	-1.149	0.0	0.001	
L3056-01	-0.840	0.0	0.001	
L3056-02	-0.617	0.0	0.001	
CCV2	252.608	0.0	0.199	
CCB2	-1.052	0.0	0.001	

N	11
Mean	53.313
SD	100.7309
CV%	188.94

Aquakem v. 7.2AQ1

Results from time period:

Mon Jun 22 10:15:46 2020

Mon Jun 22 10:54:37 2020

Sh

Sample Id	Sam/Ctr/c	Test short name	Test type	Result	Result unit	Result date and time
0.OPPBCN	A	REACTIVE CN	P	-1.3953	µg/l	6/22/2020 10:15:46
5.OPPBCN	A	REACTIVE CN	P	3.7752	µg/l	6/22/2020 10:15:47
10PPBCN	A	REACTIVE CN	P	8.9453	µg/l	6/22/2020 10:15:48
50PPBCN	A	REACTIVE CN	P	50.7411	µg/l	6/22/2020 10:15:49
100PPBCN	A	REACTIVE CN	P	101.6322	µg/l	6/22/2020 10:15:50
250PPBCN	A	REACTIVE CN	P	253.3374	µg/l	6/22/2020 10:15:51
500PPBCN	A	REACTIVE CN	P	497.9641	µg/l	6/22/2020 10:15:52
ICV1	S	REACTIVE CN	P	96.4877	µg/l	6/22/2020 10:48:49
ICB1	S	REACTIVE CN	P	-0.8368	µg/l	6/22/2020 10:48:50
CCV1	S	REACTIVE CN	P	244.1406	µg/l	6/22/2020 10:48:53
CCB1	S	REACTIVE CN	P	-0.1809	µg/l	6/22/2020 10:48:54
PB129646BL	S	REACTIVE CN	P	-0.9514	µg/l	6/22/2020 10:48:57
L3051-02	S	REACTIVE CN	P	-1.1675	µg/l	6/22/2020 10:48:58
L3051-02DUP	S	REACTIVE CN	P	-1.1492	µg/l	6/22/2020 10:54:31
L3056-01	S	REACTIVE CN	P	-0.8401	µg/l	6/22/2020 10:54:32
L3056-02	S	REACTIVE CN	P	-0.617	µg/l	6/22/2020 10:54:33
CCV2	S	REACTIVE CN	P	252.6082	µg/l	6/22/2020 10:54:34
CCB2	S	REACTIVE CN	P	-1.0524	µg/l	6/22/2020 10:54:37

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by :

Instrument ID : Konelab

6/22/2020 10:23


Test ~~Total~~ ^{River} CN

Accepted 6/22/2020 10:23

Factor 1279
Bias 0.002

Coeff. of det. 0.999888

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	-1.3953	0.0000	
2	5.0PPBCN	0.005	3.7752	5.0000	
3	10PPBCN	0.009	8.9453	10.0000	
4	50PPBCN	0.041	50.7411	50.0000	
5	100PPBCN	0.081	101.6322	100.0000	
6	250PPBCN	0.200	253.3374	250.0000	
7	500PPBCN	0.391	497.9641	500.0000	