

LB109720

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : Jh

Instrument ID : Konelab

6/24/2020 15:26

Test: ^{Result} Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.754	0.0	0.075	
ICB1	0.187	0.0	0.002	
CCV1	239.661	0.0	0.187	
CCB1	0.179	0.0	0.002	
PB129737BL	0.290	0.0	0.002	
L2818-10	0.191	0.0	0.002	
L2818-10DUP	-0.051	0.0	0.001	
L3089-02	0.055	0.0	0.002	
L3089-05	0.178	0.0	0.002	
L3089-08	0.050	0.0	0.002	
PB129738BL	0.030	0.0	0.002	
L3087-02	0.275	0.0	0.002	
L3087-02DUP	1.530	0.0	0.003	
CCV2	252.771	0.0	0.197	
CCB2	1.024	0.0	0.002	

N 15
Mean 39.408
SD 87.4306
CV% 221.86

Aquakem v. 7.2AQ1

Results from time period:

Wed Jun 24 15:22:03 2020

Wed Jun 24 15:26:25 2020

dh

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time
0.OPPBCN	A	REACTIVE CN	P	-0.5285	µg/l	6/24/2020 13:38:19
5.OPPBCN	A	REACTIVE CN	P	4.3143	µg/l	6/24/2020 13:38:20
10PPBCN	A	REACTIVE CN	P	9.557	µg/l	6/24/2020 13:38:21
50PPBCN	A	REACTIVE CN	P	49.76	µg/l	6/24/2020 13:38:22
100PPBCN	A	REACTIVE CN	P	100.5134	µg/l	6/24/2020 13:38:23
250PPBCN	A	REACTIVE CN	P	252.8934	µg/l	6/24/2020 13:38:24
500PPBCN	A	REACTIVE CN	P	498.4903	µg/l	6/24/2020 13:38:25
ICV1	S	REACTIVE CN	P	94.7538	µg/l	6/24/2020 15:22:03
ICB1	S	REACTIVE CN	P	0.1873	µg/l	6/24/2020 15:22:04
CCV1	S	REACTIVE CN	P	239.6611	µg/l	6/24/2020 15:22:05
CCB1	S	REACTIVE CN	P	0.1788	µg/l	6/24/2020 15:22:06
PB129737BL	S	REACTIVE CN	P	0.29	µg/l	6/24/2020 15:22:07
L2818-10	S	REACTIVE CN	P	0.1907	µg/l	6/24/2020 15:22:08
L2818-10DUP	S	REACTIVE CN	P	-0.051	µg/l	6/24/2020 15:22:09
L3089-02	S	REACTIVE CN	P	0.0551	µg/l	6/24/2020 15:22:10
L3089-05	S	REACTIVE CN	P	0.1782	µg/l	6/24/2020 15:22:11
L3089-08	S	REACTIVE CN	P	0.0504	µg/l	6/24/2020 15:22:12
PB129738BL	S	REACTIVE CN	P	0.0299	µg/l	6/24/2020 15:22:13
L3087-02	S	REACTIVE CN	P	0.2748	µg/l	6/24/2020 15:26:22
L3087-02DUP	S	REACTIVE CN	P	1.5302	µg/l	6/24/2020 15:26:23
CCV2	S	REACTIVE CN	P	252.7715	µg/l	6/24/2020 15:26:24
CCB2	S	REACTIVE CN	P	1.0237	µg/l	6/24/2020 15:26:25

LB109720

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : Jh

Instrument ID : Konelab

6/24/2020 14:12


Test ~~Total~~ ^{Reactive} CN

Accepted 6/24/2020 13:48

Factor 1290
Bias 0.001

Coeff. of det. 0.999942

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	-0.5285	0.0000	
2	5.0PPBCN	0.005	4.3143	5.0000	
3	10PPBCN	0.009	9.5570	10.0000	
4	50PPBCN	0.040	49.7600	50.0000	
5	100PPBCN	0.079	100.5134	100.0000	
6	250PPBCN	0.198	252.8934	250.0000	
7	500PPBCN	0.388	498.4903	500.0000	