

LB 109765

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : Dh

Instrument ID : Konelab

6/26/2020 14:18

Test: Total CN *Mg/L*

Sample Id	Result	Dil. 1 +	Response	Errors

ICV1	94.588	0.0	0.072	
ICB1	1.599	0.0	0.002	
CCV1	247.748	0.0	0.187	
CCB1	1.489	0.0	0.002	
PB129806BL	1.442	0.0	0.002	
L3125-01	1.373	0.0	0.002	
L3125-01DUP	1.462	0.0	0.002	
CCV2	247.846	0.0	0.187	
CCB2	1.745	0.0	0.002	

N	9
Mean	66.588
SD	107.1576
CV%	160.93

Aquakem v. 7.2AQ1

Results from time period:

Fri Jun 26 14:13:49 2020

Fri Jun 26 14:13:57 2020

i)h

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time
0.0PPBCN	A	Total CN	P	0.7006	µg/l	6/26/2020 11:06:19
5.0PPBCN	A	Total CN	P	5.7347	µg/l	6/26/2020 11:06:20
10PPBCN	A	Total CN	P	10.8302	µg/l	6/26/2020 11:06:21
50PPBCN	A	Total CN	P	49.7182	µg/l	6/26/2020 11:06:22
100PPBCN	A	Total CN	P	97.0969	µg/l	6/26/2020 11:06:23
250PPBCN	A	Total CN	P	250.6691	µg/l	6/26/2020 11:06:24
500PPBCN	A	Total CN	P	500.2503	µg/l	6/26/2020 11:06:25
ICV1	S	Total CN	P	94.5879	µg/l	6/26/2020 14:13:49
ICB1	S	Total CN	P	1.5994	µg/l	6/26/2020 14:13:50
CCV1	S	Total CN	P	247.7485	µg/l	6/26/2020 14:13:51
CCB1	S	Total CN	P	1.4894	µg/l	6/26/2020 14:13:52
PB129806BL	S	Total CN	P	1.4425	µg/l	6/26/2020 14:13:53
L3125-01	S	Total CN	P	1.3734	µg/l	6/26/2020 14:13:54
L3125-01DUP	S	Total CN	P	1.4622	µg/l	6/26/2020 14:13:55
CCV2	S	Total CN	P	247.8463	µg/l	6/26/2020 14:13:56
CCB2	S	Total CN	P	1.7447	µg/l	6/26/2020 14:13:57

LB109765

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : dh

Instrument ID : Konelab

6/26/2020 14:08


Test Total CN

Accepted 6/26/2020 11:06

Factor 1331
Bias 0.001

Coeff. of det. 0.999948

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	0.7006	0.0000	
2	5.0PPBCN	0.005	5.7347	5.0000	
3	10PPBCN	0.009	10.8302	10.0000	
4	50PPBCN	0.038	49.7182	50.0000	
5	100PPBCN	0.074	97.0969	100.0000	
6	250PPBCN	0.189	250.6691	250.0000	
7	500PPBCN	0.377	500.2503	500.0000	