

LB109884

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : DK

Instrument ID : Konelab

7/6/2020 16:06

Test: Phenolics-

mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.018	0.0	0.099	
ICB1	0.005	0.0	0.005	
CCV1	1.036	0.0	0.101	
CCB1	0.002	0.0	0.004	
PB129920BL	0.003	0.0	0.004	
PB129920BS	1.011	0.0	0.098	
L3198-01	0.010	0.0	0.005	
L3198-02	0.009	0.0	0.005	
L3198-03	0.004	0.0	0.005	
L3198-04	0.008	0.0	0.005	
L3198-05	0.008	0.0	0.005	
L3198-05DUP	0.005	0.0	0.005	
L3198-06MS	0.944	0.0	0.092	
L3198-07MSD	0.954	0.0	0.093	
CCV2	1.040	0.0	0.101	
CCB2	0.004	0.0	0.004	
L3198-08	0.013	0.0	0.005	
CCV3	1.027	0.0	0.100	
CCB3	0.008	0.0	0.005	

N	19
Mean	0.374
SD	0.4951
CV%	132.31

Aquakem v. 7.2AQ1

Results from time period:

Mon Jul 06 14:46:45 2020

Mon Jul 06 16:04:15 2020

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time
0.05PPM	A	Phenolics-	P	0.0476	mg/l	7/6/2020 14:46:45
0.1PPM	A	Phenolics-	P	0.1001	mg/l	7/6/2020 14:46:46
0.25PPM	A	Phenolics-	P	0.2527	mg/l	7/6/2020 14:46:47
0.50PPM	A	Phenolics-	P	0.4984	mg/l	7/6/2020 14:46:48
1.0PPM	A	Phenolics-	P	1.0021	mg/l	7/6/2020 14:46:49
2.0PPM	A	Phenolics-	P	1.9991	mg/l	7/6/2020 14:46:50
ICV1	S	Phenolics-	P	1.0185	mg/l	7/6/2020 15:56:53
ICB1	S	Phenolics-	P	0.0048	mg/l	7/6/2020 15:56:54
CCV1	S	Phenolics-	P	1.0359	mg/l	7/6/2020 15:56:55
CCB1	S	Phenolics-	P	0.0023	mg/l	7/6/2020 15:56:56
PB129920BL	S	Phenolics-	P	0.0029	mg/l	7/6/2020 15:56:57
PB129920BS	S	Phenolics-	P	1.0115	mg/l	7/6/2020 15:56:58
L3198-01	S	Phenolics-	P	0.01	mg/l	7/6/2020 15:56:59
L3198-02	S	Phenolics-	P	0.0088	mg/l	7/6/2020 15:57:00
L3198-03	S	Phenolics-	P	0.0038	mg/l	7/6/2020 15:57:01
L3198-04	S	Phenolics-	P	0.0083	mg/l	7/6/2020 15:57:02
L3198-05	S	Phenolics-	P	0.0083	mg/l	7/6/2020 15:57:03
L3198-05DUP	S	Phenolics-	P	0.0047	mg/l	7/6/2020 15:57:04
L3198-06MS	S	Phenolics-	P	0.9441	mg/l	7/6/2020 16:04:07
L3198-07MSD	S	Phenolics-	P	0.9539	mg/l	7/6/2020 16:04:08
CCV2	S	Phenolics-	P	1.0401	mg/l	7/6/2020 16:04:11
CCB2	S	Phenolics-	P	0.0036	mg/l	7/6/2020 16:04:12
L3198-08	S	Phenolics-	P	0.0127	mg/l	7/6/2020 16:04:13
CCV3	S	Phenolics-	P	1.0267	mg/l	7/6/2020 16:04:14
CCB3	S	Phenolics-	P	0.0079	mg/l	7/6/2020 16:04:15

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : Dh

Instrument ID : Konelab

7/6/2020 14:54
=====


Test Phenolics-

Accepted 7/6/2020 14:54

Factor 10.74
Bias 0.004

Coeff. of det. 0.999992

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0476	0.0500	
2	Phenol-2	0.013	0.1001	0.1000	
3	Phenol-2	0.028	0.2527	0.2500	
4	Phenol-2	0.051	0.4984	0.5000	
5	Phenol-2	0.098	1.0021	1.0000	
6	Phenol-2	0.190	1.9991	2.0000	