

LB110101

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : De

Instrument ID : Konelab

7/20/2020 12:08

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.073	0.0	0.077	
ICB1	-0.260	0.0	0.003	
CCV1	245.055	0.0	0.192	
CCB1	-0.471	0.0	0.003	
PB130320BL	-0.413	0.0	0.003	
PB130320BS	96.423	0.0	0.077	
MIDPB130320	246.438	0.0	0.193	
L3216-12	1.594	0.0	0.004	
CCV2	242.717	0.0	0.190	
CCB2	-0.316	0.0	0.003	

N	10
Mean	92.684
SD	111.7152
CV%	120.53

Aquakem v. 7.2AQ1

Results from time period:

Mon Jul 20 11:34:16 2020

Mon Jul 20 11:57:04 2020

124

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.7507	µg/l	7/20/2020 11:34:16	
5.0PPBCN	A	Total CN	P	4.1778	µg/l	7/20/2020 11:34:17	
10PPBCN	A	Total CN	P	9.4724	µg/l	7/20/2020 11:34:18	
50PPBCN	A	Total CN	P	50.6086	µg/l	7/20/2020 11:34:19	
100PPBCN	A	Total CN	P	99.6073	µg/l	7/20/2020 11:34:20	
250PPBCN	A	Total CN	P	253.6962	µg/l	7/20/2020 11:34:21	
500PPBCN	A	Total CN	P	498.1883	µg/l	7/20/2020 11:34:22	
ICV1	S	Total CN	P	96.073	µg/l	7/20/2020 11:56:55	
ICB1	S	Total CN	P	-0.2604	µg/l	7/20/2020 11:56:56	
CCV1	S	Total CN	P	245.0549	µg/l	7/20/2020 11:56:57	
CCB1	S	Total CN	P	-0.4706	µg/l	7/20/2020 11:56:58	
PB130320BL	S	Total CN	P	-0.4128	µg/l	7/20/2020 11:56:59	
PB130320BS	S	Total CN	P	96.4226	µg/l	7/20/2020 11:57:00	
MIDPB130320	S	Total CN	P	246.4381	µg/l	7/20/2020 11:57:01	
L3216-12	S	Total CN	P	1.5939	µg/l	7/20/2020 11:57:02	
CCV2	S	Total CN	P	242.7174	µg/l	7/20/2020 11:57:03	
CCB2	S	Total CN	P	-0.3159	µg/l	7/20/2020 11:57:04	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : *h*

Instrument ID : Konelab

7/20/2020 12:55


Test Total CN

Accepted 7/20/2020 11:34

Factor 1299
Bias 0.003

Coeff. of det. 0.999908

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.002	-0.7507	0.0000	
2	5.0PPBCN	0.006	4.1778	5.0000	
3	10PPBCN	0.010	9.4724	10.0000	
4	50PPBCN	0.042	50.6086	50.0000	
5	100PPBCN	0.080	99.6073	100.0000	
6	250PPBCN	0.198	253.6962	250.0000	
7	500PPBCN	0.387	498.1883	500.0000	