

LB110296

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AM

Instrument ID : Konelab

7/30/2020 13:36

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.495	0.0	0.080	
ICB1	1.194	0.0	0.003	
CCV1	243.522	0.0	0.196	
CCB1	1.001	0.0	0.003	
PB130631BL	0.892	0.0	0.003	
L3183-04 L3180-04	0.839	0.0	0.003	
L3184-04 DUP L3180-04	0.836	0.0	0.003	
L3183-02	0.850	0.0	0.003	
CCV2	244.971	0.0	0.197	
CCB2	1.162	0.0	0.003	

N	10
Mean	59.276
SD	102.0290
CV%	172.12

Aquakem v. 7.2AQ1

Results from time period:

Thu Jul 30 13:36:10 2020

Thu Jul 30 13:36:19 2020

Sample Id	Sam/Ctr/c	Test short	nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive	CN	P	-0.2826	µg/l	7/30/2020 9:41:53	
5.OPPBCN	A	Reactive	CN	P	5.2832	µg/l	7/30/2020 9:41:54	
10PPBCN	A	Reactive	CN	P	8.099	µg/l	7/30/2020 9:41:55	
50PPBCN	A	Reactive	CN	P	51.3032	µg/l	7/30/2020 9:41:56	
100PPBCN	A	Reactive	CN	P	100.4169	µg/l	7/30/2020 9:41:57	
250PPBCN	A	Reactive	CN	P	250.7177	µg/l	7/30/2020 9:41:58	
500PPBCN	A	Reactive	CN	P	499.4626	µg/l	7/30/2020 9:41:59	
ICV1	S	Reactive	CN	P	97.4951	µg/l	7/30/2020 13:36:10	
ICB1	S	Reactive	CN	P	1.1938	µg/l	7/30/2020 13:36:11	
CCV1	S	Reactive	CN	P	243.522	µg/l	7/30/2020 13:36:12	
CCB1	S	Reactive	CN	P	1.0011	µg/l	7/30/2020 13:36:13	
PB130631E	S	Reactive	CN	P	0.8922	µg/l	7/30/2020 13:36:14	
L3180-04	S	Reactive	CN	P	0.8387	µg/l	7/30/2020 13:36:15	
L3180-04D	S	Reactive	CN	P	0.8357	µg/l	7/30/2020 13:36:16	
L3183-02	S	Reactive	CN	P	0.8498	µg/l	7/30/2020 13:36:17	
CCV2	S	Reactive	CN	P	244.971	µg/l	7/30/2020 13:36:18	
CCB2	S	Reactive	CN	P	1.1616	µg/l	7/30/2020 13:36:19	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : _____

Instrument ID : Konelab

7/30/2020 15:54


Test Total CN

Accepted 7/30/2020 9:42

Factor 1258
Bias 0.002

Coeff. of det. 0.999969

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.002	-0.2826	0.0000	
2	5.0PPBCN	0.007	5.2832	5.0000	
3	10PPBCN	0.009	8.0990	10.0000	
4	50PPBCN	0.043	51.3032	50.0000	
5	100PPBCN	0.082	100.4169	100.0000	
6	250PPBCN	0.202	250.7177	250.0000	
7	500PPBCN	0.399	499.4626	500.0000	