

LB110403

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH  
284 Sheffield Street,  
Mountainside, NJ 07092  
Reviewed by : Am

Instrument ID : Konelab

8/6/2020 15:20

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	99.047	0.0	0.081	
ICB1	1.595	0.0	0.004	
CCV1	254.869	0.0	0.204	
CCB1	1.561	0.0	0.004	
PB130779BL	1.363	0.0	0.004	
PB130779BS	105.142	0.0	0.086	
L3502-01	0.482	0.0	0.003	
L3502-01DUP	0.626	0.0	0.004	
L3502-01MS	38.483	0.0	0.033	
L3502-01MSD	36.725	0.0	0.032	
L3568-02	0.025	0.0	0.003	
L3568-05	1.148	0.0	0.004	
L3573-01	5.398	0.0	0.007	
L3573-03	5.510	0.0	0.007	
CCV2	248.465	0.0	0.199	
CCB2	1.861	0.0	0.005	
L3574-01	4.945	0.0	0.007	
L3574-03	4.674	0.0	0.007	
L3575-01	14.023	0.0	0.014	
L3576-01	8.233	0.0	0.010	
CCV3	258.579	0.0	0.207	
CCB3	1.566	0.0	0.004	

N 22  
Mean 49.742  
SD 88.1563  
CV% 177.23



Aquakem v. 7.2AQ1

Results from time period:

Thu Aug 06 14:43:13 2020

Thu Aug 06 16:03:46 2020

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.6152	µg/l	8/6/2020 8:32:27	
5.OPPBCN	A	Total CN	P	4.6777	µg/l	8/6/2020 8:32:28	
10PPBCN	A	Total CN	P	9.7513	µg/l	8/6/2020 8:32:29	
50PPBCN	A	Total CN	P	50.0497	µg/l	8/6/2020 8:32:30	
100PPBCN	A	Total CN	P	100.4635	µg/l	8/6/2020 8:32:31	
250PPBCN	A	Total CN	P	251.5251	µg/l	8/6/2020 8:32:32	
500PPBCN	A	Total CN	P	499.148	µg/l	8/6/2020 8:32:33	
ICV1	S	Total CN	P	99.0474	µg/l	8/6/2020 14:43:13	
ICB1	S	Total CN	P	1.5953	µg/l	8/6/2020 14:43:14	
CCV1	S	Total CN	P	254.8688	µg/l	8/6/2020 15:04:21	
CCB1	S	Total CN	P	1.5614	µg/l	8/6/2020 15:04:22	
PB130779BL	S	Total CN	P	1.3626	µg/l	8/6/2020 15:04:23	
PB130779BS	S	Total CN	P	105.1421	µg/l	8/6/2020 15:08:39	
L3502-01	S	Total CN	P	0.4816	µg/l	8/6/2020 15:08:40	
L3502-01DUP	S	Total CN	P	0.6258	µg/l	8/6/2020 15:08:41	
L3502-01MS	S	Total CN	P	38.4834	µg/l	8/6/2020 15:08:42	
L3502-01MSD	S	Total CN	P	36.7247	µg/l	8/6/2020 15:16:14	
L3568-02	S	Total CN	P	0.0248	µg/l	8/6/2020 15:16:15	
L3568-05	S	Total CN	P	1.1475	µg/l	8/6/2020 15:16:16	
L3573-01	S	Total CN	P	5.3984	µg/l	8/6/2020 15:16:17	
L3573-03	S	Total CN	P	5.5097	µg/l	8/6/2020 15:16:18	
CCV2	S	Total CN	P	248.4649	µg/l	8/6/2020 15:16:19	
CCB2	S	Total CN	P	1.8608	µg/l	8/6/2020 15:16:20	
L3574-01	S	Total CN	P	4.9448	µg/l	8/6/2020 15:16:21	
L3574-03	S	Total CN	P	4.6739	µg/l	8/6/2020 15:16:22	
L3575-01	S	Total CN	P	14.0231	µg/l	8/6/2020 15:16:23	
L3576-01	S	Total CN	P	8.2334	µg/l	8/6/2020 15:16:24	
CCV3	S	Total CN	P	258.5786	µg/l	8/6/2020 15:19:37	
CCB3	S	Total CN	P	1.5661	µg/l	8/6/2020 15:19:38	
LOWPB130779	S	Total CN	P	7.5751	µg/l	8/6/2020 16:03:43	
HIGHPB130779	S	Total CN	P	495.2289	µg/l	8/6/2020 16:03:44	
CCV4	S	Total CN	P	253.8458	µg/l	8/6/2020 16:03:45	
CCB4	S	Total CN	P	1.1791	µg/l	8/6/2020 16:03:46	

