

LB110606

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : Am

Instrument ID : Konelab

8/19/2020 16:18

Test: Phenolics-

Sample Id	Result ^{mg/L}	Dil. 1 +	Response	Errors
ICV1	0.947	0.0	0.113	
ICB1	0.005	0.0	0.006	
CCV1	0.939	0.0	0.112	
CCB1	-0.001	0.0	0.005	
PB1301094BL	0.000	0.0	0.006	
PB131094BS	0.939	0.0	0.112	
L3723-01	-0.001	0.0	0.005	
L3723-01DUP	0.002	0.0	0.006	
L3723-01MS	0.825	0.0	0.099	
L3723-01MSD	0.823	0.0	0.099	
CCV2	0.984	0.0	0.117	
CCB2	-0.000	0.0	0.006	

N	12
Mean	0.455
SD	0.4767
CV%	104.73

Aquakem v. 7.2AQ1

Results from time period:

Wed Aug 19 15:20:46 2020

Wed Aug 19 16:17:06 2020

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0536	mg/l	8/19/2020 15:20:46	
0.1PPM	A	Phenolics-	P	0.1022	mg/l	8/19/2020 15:20:47	
0.25PPM	A	Phenolics-	P	0.2516	mg/l	8/19/2020 15:20:48	
0.50PPM	A	Phenolics-	P	0.4982	mg/l	8/19/2020 15:20:49	
1.0PPM	A	Phenolics-	P	0.9887	mg/l	8/19/2020 15:20:50	
2.0PPM	A	Phenolics-	P	2.0057	mg/l	8/19/2020 15:20:51	
ICV1	S	Phenolics-	P	0.9468	mg/l	8/19/2020 15:53:34	
ICB1	S	Phenolics-	P	0.0052	mg/l	8/19/2020 15:53:35	
CCV1	S	Phenolics-	P	0.939	mg/l	8/19/2020 15:53:36	
CCB1	S	Phenolics-	P	-0.0008	mg/l	8/19/2020 15:53:37	
PB1301094BL	S	Phenolics-	P	0.0003	mg/l	8/19/2020 15:53:38	
PB131094BS	S	Phenolics-	P	0.9387	mg/l	8/19/2020 15:53:39	
L3723-01	S	Phenolics-	P	-0.0011	mg/l	8/19/2020 16:17:01	
L3723-01DUP	S	Phenolics-	P	0.0021	mg/l	8/19/2020 16:17:02	
L3723-01MS	S	Phenolics-	P	0.8252	mg/l	8/19/2020 16:17:03	
L3723-01MSD	S	Phenolics-	P	0.8234	mg/l	8/19/2020 16:17:04	
CCV2	S	Phenolics-	P	0.9835	mg/l	8/19/2020 16:17:05	
CCB2	S	Phenolics-	P	-0.0004	mg/l	8/19/2020 16:17:06	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : Am

Instrument ID : Konelab

8/19/2020 15:23


Test Phenolics-

Accepted 8/19/2020 15:23

Factor 8.818
Bias 0.006

Coeff. of det. 0.999934

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.012	0.0536	0.0500	
2	Phenol-2	0.017	0.1022	0.1000	
3	Phenol-2	0.034	0.2516	0.2500	
4	Phenol-2	0.062	0.4982	0.5000	
5	Phenol-2	0.118	0.9887	1.0000	
6	Phenol-2	0.233	2.0057	2.0000	