

LB110651

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AM

Instrument ID : Konelab

8/21/2020 15:24

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	99.746	0.0	0.076	
ICB1	0.585	0.0	0.002	
CCV1	250.308	0.0	0.189	
CCB1	0.254	0.0	0.001	
PB131147BL	0.371	0.0	0.001	
PB131147BS	100.232	0.0	0.076	
MIDPB131147	254.656	0.0	0.192	
L3723-01	7.982	0.0	0.007	
L3723-01DUP	7.766	0.0	0.007	
CCV2	259.373	0.0	0.196	
CCB2	0.684	0.0	0.002	

N	11
Mean	89.269
SD	112.8012
CV%	126.36

Aquakem v. 7.2AQ1

Results from time period:

Fri Aug 21 15:15:33 2020

Fri Aug 21 15:18:57 2020

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.8976	µg/l	8/21/2020 11:23:43	
5.0PPBCN	A	Total CN	P	4.3145	µg/l	8/21/2020 11:23:44	
10PPBCN	A	Total CN	P	9.4213	µg/l	8/21/2020 11:23:45	
50PPBCN	A	Total CN	P	50.0637	µg/l	8/21/2020 11:23:46	
100PPBCN	A	Total CN	P	99.9283	µg/l	8/21/2020 11:23:47	
250PPBCN	A	Total CN	P	254.2867	µg/l	8/21/2020 11:23:48	
500PPBCN	A	Total CN	P	497.8831	µg/l	8/21/2020 11:23:49	
ICV1	S	Total CN	P	99.7458	µg/l	8/21/2020 15:15:33	
ICB1	S	Total CN	P	0.5854	µg/l	8/21/2020 15:15:34	
CCV1	S	Total CN	P	250.308	µg/l	8/21/2020 15:15:35	
CCB1	S	Total CN	P	0.2538	µg/l	8/21/2020 15:15:36	
PB131147BL	S	Total CN	P	0.3706	µg/l	8/21/2020 15:15:37	
PB131147BS	S	Total CN	P	100.2318	µg/l	8/21/2020 15:15:38	
MIDPB131147	S	Total CN	P	254.6564	µg/l	8/21/2020 15:15:39	
L3723-01	S	Total CN	P	7.9816	µg/l	8/21/2020 15:15:42	
L3723-01DUP	S	Total CN	P	7.7658	µg/l	8/21/2020 15:15:43	
CCV2	S	Total CN	P	259.373	µg/l	8/21/2020 15:18:56	
CCB2	S	Total CN	P	0.6838	µg/l	8/21/2020 15:18:57	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AM

Instrument ID : Konelab

8/21/2020 11:25


Test Total CN

Accepted 8/21/2020 11:25

Factor 1333
Bias 0.001

Coeff. of det. 0.999881

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.000	-0.8976	0.0000	
2	5.0PPBCN	0.004	4.3145	5.0000	
3	10PPBCN	0.008	9.4213	10.0000	
4	50PPBCN	0.039	50.0637	50.0000	
5	100PPBCN	0.076	99.9283	100.0000	
6	250PPBCN	0.192	254.2867	250.0000	
7	500PPBCN	0.375	497.8831	500.0000	