

LBH1022

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AM

Instrument ID : Konelab

9/17/2020 12:36

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.982	0.0	0.106	
ICB1	0.000	0.0	0.004	
CCV1	0.995	0.0	0.107	
CCB1	-0.000	0.0	0.004	
PB131565BL	-0.003	0.0	0.004	
PB131565BS	0.980	0.0	0.106	
L3974-26	1.225	0.0	0.131	
CCV2	0.994	0.0	0.107	
CCB2	-0.002	0.0	0.004	

N	9
Mean	0.575
SD	0.5514
CV%	95.96

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AM

Instrument ID : Konelab

9/17/2020 11:20


Test Phenolics-

Accepted 9/17/2020 11:20

Factor 9.633
Bias 0.004

Coeff. of det. 0.999836

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0502	0.0500	
2	Phenol-2	0.014	0.0972	0.1000	
3	Phenol-2	0.031	0.2561	0.2500	
4	Phenol-2	0.057	0.5079	0.5000	
5	Phenol-2	0.106	0.9823	1.0000	
6	Phenol-2	0.212	2.0062	2.0000	

Aquakem v. 7.2AQ1

Results from time period:

Thu Sep 17 12:34:26 2020

Thu Sep 17 12:34:35 2020

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0502	mg/l	9/17/2020 11:19:24	
0.1PPM	A	Phenolics-	P	0.0972	mg/l	9/17/2020 11:19:25	
0.25PPM	A	Phenolics-	P	0.2561	mg/l	9/17/2020 11:19:26	
0.50PPM	A	Phenolics-	P	0.5079	mg/l	9/17/2020 11:19:27	
1.0PPM	A	Phenolics-	P	0.9823	mg/l	9/17/2020 11:19:28	
2.0PPM	A	Phenolics-	P	2.0062	mg/l	9/17/2020 11:19:29	
ICV1	S	Phenolics-	P	0.9825	mg/l	9/17/2020 12:34:26	
ICB1	S	Phenolics-	P	0.0001	mg/l	9/17/2020 12:34:27	
CCV1	S	Phenolics-	P	0.995	mg/l	9/17/2020 12:34:28	
CCB1	S	Phenolics-	P	-0.0003	mg/l	9/17/2020 12:34:29	
PB131565BL	S	Phenolics-	P	-0.0029	mg/l	9/17/2020 12:34:30	
PB131565BS	S	Phenolics-	P	0.9803	mg/l	9/17/2020 12:34:31	
L3974-26	S	Phenolics-	P	1.2247	mg/l	9/17/2020 12:34:33	
CCV2	S	Phenolics-	P	0.9939	mg/l	9/17/2020 12:34:34	
CCB2	S	Phenolics-	P	-0.0016	mg/l	9/17/2020 12:34:35	