

LB111037

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AM

Instrument ID : Konelab

9/18/2020 10:19

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.789	0.0	0.081	
ICB1	-0.249	0.0	0.000	
CCV1	245.436	0.0	0.203	
CCB1	-0.451	0.0	-0.000	
PB131641BL	-0.300	0.0	0.000	
PB131641BS	97.913	0.0	0.081	
L3974-25	750.195	0.0	0.619	Test limit high
L3974-25DLX2	364.656	0.0	0.301	
CCV2	253.125	0.0	0.209	
CCB2	-0.551	0.0	-0.000	

N	10
Mean	180.756
SD	238.6303
CV%	132.02

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AM

Instrument ID : Konelab

9/18/2020 9:42

Test Total CN


Accepted 9/18/2020 9:42

Factor 1213

Bias 0

Coeff. of det. 0.999978

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	-0.000	-0.5892	0.0000	
2	5.0PPBCN	0.004	4.3432	5.0000	
3	10PPBCN	0.008	9.6015	10.0000	
4	50PPBCN	0.042	50.7002	50.0000	
5	100PPBCN	0.084	101.6237	100.0000	
6	250PPBCN	0.206	249.4017	250.0000	
7	500PPBCN	0.412	499.9189	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Fri Sep 18 09:41:39 2020

Fri Sep 18 10:17:37 2020

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.5892	µg/l	9/18/2020 9:41:39	
5.OPPBCN	A	Total CN	P	4.3432	µg/l	9/18/2020 9:41:40	
10PPBCN	A	Total CN	P	9.6015	µg/l	9/18/2020 9:41:41	
50PPBCN	A	Total CN	P	50.7002	µg/l	9/18/2020 9:41:42	
100PPBCN	A	Total CN	P	101.6237	µg/l	9/18/2020 9:41:43	
250PPBCN	A	Total CN	P	249.4017	µg/l	9/18/2020 9:41:44	
500PPBCN	A	Total CN	P	499.9189	µg/l	9/18/2020 9:41:45	
ICV1	S	Total CN	P	97.7887	µg/l	9/18/2020 10:14:13	
ICB1	S	Total CN	P	-0.2486	µg/l	9/18/2020 10:14:14	
CCV1	S	Total CN	P	245.4364	µg/l	9/18/2020 10:14:15	
CCB1	S	Total CN	P	-0.4509	µg/l	9/18/2020 10:14:16	
PB131641BL	S	Total CN	P	-0.3	µg/l	9/18/2020 10:14:17	
PB131641BS	S	Total CN	P	97.9133	µg/l	9/18/2020 10:14:18	
L3974-25	S	Total CN	P	750.1954	µg/l	9/18/2020 10:14:21	
L3974-25DLX2	S	Total CN	P	364.6559	µg/l	9/18/2020 10:14:22	
CCV2	S	Total CN	P	253.1245	µg/l	9/18/2020 10:17:36	
CCB2	S	Total CN	P	-0.5507	µg/l	9/18/2020 10:17:37	