

LB114362

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AM

Instrument ID : Konelab

10/7/2020 16:21

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.413	0.0	0.085	
ICB1	0.132	0.0	0.001	
CCV1	240.223	0.0	0.208	
CCB1	-0.149	0.0	0.000	
PB132183BL	-0.113	0.0	0.000	
PB132183BS	99.453	0.0	0.086	
MIDPB132183	243.211	0.0	0.210	
L4214-10	2.630	0.0	0.003	
L4214-11	4.343	0.0	0.004	
CCV2	244.374	0.0	0.211	
CCB2	0.330	0.0	0.001	

N	11
Mean	84.713
SD	108.1866
CV%	127.71

Aquakem v. 7.2AQ1

Results from time period:

Wed Oct 07 15:56:53 2020

Wed Oct 07 16:20:44 2020

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.6804	µg/l	10/7/2020 9:09:56	
5.OPPBCN	A	Total CN	P	5.5399	µg/l	10/7/2020 9:09:57	
10PPBCN	A	Total CN	P	9.6685	µg/l	10/7/2020 9:09:58	
50PPBCN	A	Total CN	P	49.1542	µg/l	10/7/2020 9:09:59	
100PPBCN	A	Total CN	P	101.5956	µg/l	10/7/2020 9:10:00	
250PPBCN	A	Total CN	P	249.9109	µg/l	10/7/2020 9:10:01	
500PPBCN	A	Total CN	P	499.8113	µg/l	10/7/2020 9:10:02	
ICV1	S	Total CN	P	97.4133	µg/l	10/7/2020 16:20:34	
ICB1	S	Total CN	P	0.1321	µg/l	10/7/2020 16:20:35	
CCV1	S	Total CN	P	240.2233	µg/l	10/7/2020 16:20:36	
CCB1	S	Total CN	P	-0.1489	µg/l	10/7/2020 16:20:37	
PB132183BL	S	Total CN	P	-0.1132	µg/l	10/7/2020 16:20:38	
PB132183BS	S	Total CN	P	99.4527	µg/l	10/7/2020 16:20:39	
MIDPB132183	S	Total CN	P	243.2106	µg/l	10/7/2020 16:20:40	
L4214-10	S	Total CN	P	2.6303	µg/l	10/7/2020 16:20:41	
L4214-11	S	Total CN	P	4.343	µg/l	10/7/2020 16:20:42	
CCV2	S	Total CN	P	244.3738	µg/l	10/7/2020 16:20:43	
CCB2	S	Total CN	P	0.3304	µg/l	10/7/2020 16:20:44	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : Am

Instrument ID : Konelab

10/7/2020 9:17


Test Total CN

Accepted 10/7/2020 9:17

Factor 1159
Bias 0.001

Coeff. of det. 0.999980

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	-0.000	-0.6804	0.0000	
2	5.0PPBCN	0.005	5.5399	5.0000	
3	10PPBCN	0.009	9.6685	10.0000	
4	50PPBCN	0.043	49.1542	50.0000	
5	100PPBCN	0.088	101.5956	100.0000	
6	250PPBCN	0.216	249.9109	250.0000	
7	500PPBCN	0.432	499.8113	500.0000	