

LB111493

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AM

Instrument ID : Konelab

10/14/2020 11:43

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.029	0.0	0.109	
ICB1	-0.000	0.0	0.004	
CCV1	1.034	0.0	0.110	
CCB1	-0.000	0.0	0.004	
PB132325BL	-0.000	0.0	0.004	
PB132325BS	1.034	0.0	0.110	
L4214-03	0.027	0.0	0.007	
PB132327BL	-0.003	0.0	0.004	
PB132327BS	1.032	0.0	0.110	
L4214-09	0.025	0.0	0.007	
CCV2	1.028	0.0	0.109	
CCB2	-0.002	0.0	0.004	

N	12
Mean	0.434
SD	0.5278
CV%	121.72

Aquakem v. 7.2AQ1

Results from time period:

Wed Oct 14 11:32:20 2020

Wed Oct 14 11:39:13 2020

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0546	mg/l	10/14/2020 10:04:35	
0.1PPM	A	Phenolics-	P	0.1013	mg/l	10/14/2020 10:04:36	
0.25PPM	A	Phenolics-	P	0.2532	mg/l	10/14/2020 10:04:37	
0.50PPM	A	Phenolics-	P	0.4954	mg/l	10/14/2020 10:04:38	
1.0PPM	A	Phenolics-	P	0.9899	mg/l	10/14/2020 10:04:39	
2.0PPM	A	Phenolics-	P	2.0057	mg/l	10/14/2020 10:04:40	
ICV1	S	Phenolics-	P	1.0291	mg/l	10/14/2020 11:32:21	
ICB1	S	Phenolics-	P	0	mg/l	10/14/2020 11:32:22	
CCV1	S	Phenolics-	P	1.0338	mg/l	10/14/2020 11:32:25	
CCB1	S	Phenolics-	P	-0.0002	mg/l	10/14/2020 11:32:26	
PB132325BL	S	Phenolics-	P	-0.0003	mg/l	10/14/2020 11:32:28	
PB132325BS	S	Phenolics-	P	1.0339	mg/l	10/14/2020 11:32:29	
L4214-03	S	Phenolics-	P	0.0271	mg/l	10/14/2020 11:32:30	
PB132327BL	S	Phenolics-	P	-0.0029	mg/l	10/14/2020 11:39:06	
PB132327BS	S	Phenolics-	P	1.0323	mg/l	10/14/2020 11:39:07	
L4214-09	S	Phenolics-	P	0.0249	mg/l	10/14/2020 11:39:09	
CCV2	S	Phenolics-	P	1.0284	mg/l	10/14/2020 11:39:11	
CCB2	S	Phenolics-	P	-0.0022	mg/l	10/14/2020 11:39:12	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AM

Instrument ID : Konelab

10/14/2020 10:05


Test Phenolics-

Accepted 10/14/2020 10:05

Factor 9.805
 Bias 0.004

Coeff. of det. 0.999932

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0546	0.0500	
2	Phenol-2	0.015	0.1013	0.1000	
3	Phenol-2	0.030	0.2532	0.2500	
4	Phenol-2	0.055	0.4954	0.5000	
5	Phenol-2	0.105	0.9899	1.0000	
6	Phenol-2	0.209	2.0057	2.0000	