

LB111615

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AM

Instrument ID : Konelab

10/20/2020 15:54

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.986	0.0	0.113	
ICB1	-0.001	0.0	0.005	
CCV1	0.989	0.0	0.113	
CCB1	0.000	0.0	0.005	
PB132489BL	-0.000	0.0	0.005	
PB132489BS	0.981	0.0	0.112	
L4457-02	0.013	0.0	0.007	
L4457-02DUP	0.011	0.0	0.006	
L4457-02MS	0.979	0.0	0.112	
L4457-02MSD	0.981	0.0	0.112	
L4457-03	0.003	0.0	0.006	
L4457-04	0.010	0.0	0.006	
CCV2	0.986	0.0	0.113	
CCB2	0.008	0.0	0.006	

N	14
Mean	0.425
SD	0.5024
CV%	118.28

Aquakem v. 7.2AQ1

Results from time period:

Tue Oct 20 15:00:49 2020

Tue Oct 20 15:48:57 2020

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0481	mg/l	10/20/2020 15:00:49	
0.1PPM	A	Phenolics-	P	0.0996	mg/l	10/20/2020 15:00:50	
0.25PPM	A	Phenolics-	P	0.2537	mg/l	10/20/2020 15:00:51	
0.50PPM	A	Phenolics-	P	0.4981	mg/l	10/20/2020 15:00:52	
1.0PPM	A	Phenolics-	P	1.001	mg/l	10/20/2020 15:00:53	
2.0PPM	A	Phenolics-	P	1.9996	mg/l	10/20/2020 15:00:54	
ICV1	S	Phenolics-	P	0.9861	mg/l	10/20/2020 15:38:53	
ICB1	S	Phenolics-	P	-0.0012	mg/l	10/20/2020 15:38:54	
CCV1	S	Phenolics-	P	0.9891	mg/l	10/20/2020 15:38:57	
CCB1	S	Phenolics-	P	0.0001	mg/l	10/20/2020 15:38:58	
PB132489BL	S	Phenolics-	P	-0.0003	mg/l	10/20/2020 15:39:00	
PB132489BS	S	Phenolics-	P	0.9806	mg/l	10/20/2020 15:39:01	
L4457-02	S	Phenolics-	P	0.0134	mg/l	10/20/2020 15:39:02	
L4457-02DUP	S	Phenolics-	P	0.0114	mg/l	10/20/2020 15:39:03	
L4457-02MS	S	Phenolics-	P	0.9792	mg/l	10/20/2020 15:48:47	
L4457-02MSD	S	Phenolics-	P	0.9809	mg/l	10/20/2020 15:48:48	
L4457-03	S	Phenolics-	P	0.0027	mg/l	10/20/2020 15:48:51	
L4457-04	S	Phenolics-	P	0.0097	mg/l	10/20/2020 15:48:52	
CCV2	S	Phenolics-	P	0.9863	mg/l	10/20/2020 15:48:54	
CCB2	S	Phenolics-	P	0.0084	mg/l	10/20/2020 15:48:55	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : Am

Instrument ID : Konelab

10/20/2020 15:01


Test Phenolics-

Accepted 10/20/2020 15:01

Factor 9.16
Bias 0.005

Coeff. of det. 0.999992

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0481	0.0500	
2	Phenol-2	0.016	0.0996	0.1000	
3	Phenol-2	0.033	0.2537	0.2500	
4	Phenol-2	0.060	0.4981	0.5000	
5	Phenol-2	0.115	1.0010	1.0000	
6	Phenol-2	0.224	1.9996	2.0000	