

IC-1
90567
11/05/2020
AM


Clear table

ident	concentra tion F-	concentrati on CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-10-21_	10/21/20 9:52		AM/AP
STD2	0.1510	0.3230	0.3550	1.2840	0.2870	0.2070	1.7940	2020-10-21_	10/21/20 10:24		AM/AP
STD3	0.4240	0.6210	0.6280	2.1740	0.5530	1.0370	3.2180	2020-10-21_	10/21/20 10:55		AM/AP
STD4	0.7410	1.2040	1.2010	4.0050	1.0200	2.2220	6.0340	2020-10-21_	10/21/20 11:27		AM/AP
STD5	2.1010	2.9840	2.9670	9.7680	2.4740	5.4800	15.0250	2020-10-21_	10/21/20 11:59		AM/AP
STD6	3.9820	6.1310	6.1160	19.8290	5.0570	10.3340	30.7650	2020-10-21_	10/21/20 12:31		AM/AP
STD7	4.9840	7.6160	7.6410	25.3730	6.3360	12.4270	37.9110	2020-10-21_	10/21/20 13:03		AM/AP
ICV	2.0270	2.9300	2.8810	9.6560	2.3950	5.1140	14.2850	2020-10-21_	10/21/20 13:35		AM/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-10-21_	10/21/20 14:21		AM/AP
CCV	2.1110	3.0920	3.0880	10.5410	2.5970	5.3300	15.5750	2020-11-04_	11/4/20 9:46		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-11-04_	11/4/20 10:19		AM/AP
LB111830BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-11-04_	11/4/20 10:51		AM/AP
LB111830BSW	2.0790	2.9710	2.9770	10.1270	2.4900	5.1810	14.7640	2020-11-04_	11/4/20 11:23		AM/AP
L4214-07	0.1130	0.2320	0.2810	0.8440	0.1840	0.1470	1.0850	2020-11-04_	11/4/20 13:02		AM/AP
L4214-08	0.1980	0.3760	0.4260	1.3430	0.3260	0.4220	1.9510	2020-11-04_	11/4/20 13:42		AM/AP
L4214-08RE	0.4900	0.6840	0.7380	2.3980	0.5870	0.9260	3.4530	2020-11-04_	11/4/20 14:14		AM/AP
CCV	2.1880	3.1090	3.1000	10.4920	2.6230	5.1600	14.9490	2020-11-04_	11/4/20 16:34		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-11-04_	11/4/20 17:06		AM/AP

CALIBRATION OF COMPONENT F-

IC-1
300.0
10/21/2020
AM

Method: ANIONS 10-21-20.mtw
Equation: $Q = 0.918264 \cdot A + 0.765631$
RSD: 3.031 %
Correlation coefficient: 0.999612



K3 = 0 K2 = 0 K1 = 0.918264 K0 = 0.765631

Base: Area

Ref.channel: Cond

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0	0	0	0	0			No
2	0.1601	3.153	0.2	20	5.718			Yes
3	0.3213	8.596	0.4	20	5.718			Yes
4	0.6734	14.98	0.8	20	5.718			Yes
5	1.956	44.71	2	20	5.718			Yes
6	4.097	85.91	4	20	5.718			Yes
7	5.106	107.7	5	20	5.718			Yes

CALIBRATION OF COMPONENT CL-

Method: ANIONS 10-21-20.mtw
 Equation: $Q = 1.39326 \cdot A + 1.50457$
 RSD: 3.440 %
 Correlation coefficient: 0.999500



K3 = 0 K2 = 0 K1 = 1.39326 K0 = 1.50457

Base: Area

Ref.channel: Cond

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0	0	0	0	0			No
2	0.2646	4.081	0.3	20	8.798			Yes
3	0.5382	8.181	0.6	20	8.798			Yes
4	1.059	16.23	1.2	20	8.798			Yes
5	2.687	40.11	3	20	8.798			Yes
6	5.722	83.67	6	20	8.798			Yes
7	7.384	108.2	7.5	20	8.798			Yes

CALIBRATION OF COMPONENT
NO2

Method: ANIONS 10-21-20.mtw
Equation: $Q = 0.599456 \cdot A + 2.8051$
RSD: 4.284 %
Correlation coefficient: 0.999224




K3 = 0 K2 = 0 K1 = 0.599456 K0 = 2.8051
Base: Area
Ref.channel: Cond
ISTD:
Formula: Linear
Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0	0	0	0	0		0	No
2	0.4521	8.493	0.3	20	10.57			Yes
3	0.9262	17.06	0.6	20	10.57			Yes
4	1.892	34.98	1.2	20	10.57			Yes
5	4.821	90.05	3	20	10.57			Yes
6	9.957	191.7	6	20	10.57			Yes
7	12.59	250.2	7.5	20	10.57			Yes

CALIBRATION OF COMPONENT BR-

Method: ANIONS 10-21-20.mtw
 Equation: $Q = 3.21197 \cdot A + 7.27565$
 RSD: 3.349 %
 Correlation coefficient: 0.999526



K3 = 0 K2 = 0 K1 = 3.21197 K0 = 7.27565

Base: Area

Ref.channel: Cond

ISTD:

Formula: Linear


Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0	0	0	0	0			No
2	0.2296	5.45	1	20	13.6	Yes		Yes
3	0.4785	11.11	2	20	13.6	Yes		Yes
4	0.9866	22.33	4	20	13.6	Yes		Yes
5	2.588	57.47	10	20	13.6	Yes		Yes
6	5.519	120.4	20	20	13.6	Yes		Yes
7	7.168	155.7	25	20	13.6	Yes		Yes

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0	0	No
2	0.2729	8.015	0.25	20	15.54	Yes	
3	0.6046	17.62	0.5	20	15.54	Yes	
4	1.184	33.79	1	20	15.54	Yes	
5	3.003	82.5	2.5	20	15.54	Yes	
6	6.303	171.5	5	20	15.54	Yes	
7	8.206	220.8	6.25	20	15.54	Yes	

CALIBRATION OF COMPONENT HPO4

Method: ANIONS 10-21-20.mtw
 Equation: $Q = 1.33739 \cdot A - 5.3391$
 RSD: 3.688 %
 Correlation coefficient: 0.999425




K3 = 0 K2 = 0 K1 = 1.33739 K0 = -5.3391
 Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0	0	0	0	0			No
2	0.1611	7.495	0.5	20	21.75			Yes
3	0.41	19.31	1	20	21.75			Yes
4	0.8275	36.22	2	20	21.75			Yes
5	1.949	81.85	5	20	21.75			Yes
6	3.747	152.8	10	20	21.75			Yes
7	4.71	189.8	12.5	20	21.75			Yes

CALIBRATION OF COMPONENT SO4

Method: ANIONS 10-21-20.mtw
 Equation: $Q = 1.8684 \cdot A + 7.87171$
 RSD: 2.868 %
 Correlation coefficient: 0.999652



K3 = 0 K2 = 0 K1 = 1.8684 K0 = 7.87171

Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0	0	0	0	0		0	No
2	0.3746	15.35	1.5	20	25.43		25.43	Yes
3	0.7421	30.36	3	20	25.43		25.43	Yes
4	1.489	58.99	6	20	25.43		25.43	Yes
5	3.819	149.7	15	20	25.43		25.43	Yes
6	8.032	314.3	30	20	25.43		25.43	Yes
7	10.39	401.6	37.5	20	25.43		25.43	Yes

Report date: 10/21/2020 1:44:14 PM
Printed by: wet


Ident: STD1
Analysis from: 10/21/2020 9:52:10 AM
File: _2020-10-21_

Last save: 10/21/2020 10:20:58 AM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 71947

Last save: 10/21/2020 9:42:32 AM

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/21/2020 1:44:20 PM
Printed by: wet


Ident: STD2
Analysis from: 10/21/2020 10:24:04 AM
File: _2020-10-21_

Last save: 10/21/2020 10:52:52 AM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 71948

Last save: 10/21/2020 10:20:58 AM

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.71	0.271	0.16	8.36	3.153	6.06	0.
2	8.88	0.227	0.26	13.81	4.081	7.84	0.
3	10.80	0.275	0.45	23.62	8.493	16.32	0.
4	14.03	0.351	0.23	11.95	5.450	10.47	0.
5	16.28	0.433	0.27	14.22	8.015	15.40	0.
6	21.71	0.552	0.16	8.40	7.495	14.40	0.
7	25.11	0.517	0.37	19.54	15.351	29.50	0.
7	29.00	0.375	1.91	99.89	52.038	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 10/21/2020 1:44:25 PM
 Printed by: wet


Ident: STD3
 Analysis from: 10/21/2020 10:55:58 AM
 File: _2020-10-21_

Last save: 10/21/2020 11:24:46 AM

Method: ANIONS 10-21-20.mtw
 Run operator: wet
 Analysis number: 71949

Last save: 10/21/2020 10:52:52 AM

SAMPLE:
 : AM/AP
 Vial number: 4
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.72	0.280	0.32	7.98	8.596	7.66	0.
2	8.86	0.224	0.54	13.38	8.181	7.29	0.
3	10.77	0.272	0.93	23.02	17.058	15.20	0.
4	13.90	0.340	0.48	11.87	11.112	9.90	0.
5	16.09	0.425	0.60	15.03	17.620	15.70	0.
6	21.88	0.549	0.41	10.20	19.308	17.20	0.
7	25.37	0.522	0.74	18.45	30.363	27.05	0.
7	29.00	0.373	4.02	99.94	112.238	100.00	0.

This report has been created by IC Net
 METROHM LTD

Report date: 10/21/2020 1:44:31 PM
Printed by: wet


Ident: STD4
Analysis from: 10/21/2020 11:27:53 AM
File: _2020-10-21_

Last save: 10/21/2020 11:56:42 AM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 71950

Last save: 10/21/2020 11:24:46 AM

SAMPLE:
: AM/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.79	0.281	0.67	8.29	14.980	6.89	0.
2	8.90	0.221	1.06	13.05	16.234	7.46	0.
3	10.77	0.273	1.89	23.32	34.976	16.08	0.
4	13.86	0.331	0.99	12.16	22.331	10.27	0.
5	16.00	0.418	1.18	14.60	33.792	15.54	0.
6	21.78	0.524	0.83	10.20	36.221	16.65	0.
7	25.20	0.505	1.49	18.35	58.987	27.12	0.
7	29.00	0.365	8.11	99.97	217.521	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 10/21/2020 1:44:36 PM
Printed by: wet


Ident: STD5
Analysis from: 10/21/2020 11:59:47 AM
File: _2020-10-21_

Last save: 10/21/2020 12:28:36 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 71951

Last save: 10/21/2020 11:56:42 AM

SAMPLE:
: AM/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.70	0.270	1.96	9.39	44.712	8.18	0.
2	8.82	0.219	2.69	12.90	40.105	7.34	0.
3	10.67	0.278	4.82	23.15	90.046	16.48	0.
4	13.72	0.322	2.59	12.43	57.468	10.52	0.
5	15.79	0.403	3.00	14.42	82.499	15.10	0.
6	21.65	0.508	1.95	9.36	81.846	14.98	0.
7	25.08	0.504	3.82	18.34	149.663	27.39	0.
7	29.00	0.358	20.82	99.99	546.339	100.00	0.


This report has been created by IC Net
METROHM LTD

Report date: 10/21/2020 1:44:42 PM
Printed by: wet

Ident: STD6
Analysis from: 10/21/2020 12:31:41 PM
File: _2020-10-21_ Last save: 10/21/2020 1:00:30 PM

Method: ANIONS 10-21-20.mtw Last save: 10/21/2020 12:28:36 PM
Run operator: wet
Analysis number: 71952

SAMPLE:
: AM/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.76	0.267	4.10	9.44	85.906	7.67	0.
2	8.86	0.212	5.72	13.19	83.670	7.47	0.
3	10.70	0.288	9.96	22.96	191.675	17.11	0.
4	13.75	0.314	5.52	12.72	120.376	10.75	0.
5	15.78	0.399	6.30	14.53	171.462	15.31	0.
6	21.76	0.500	3.75	8.64	152.838	13.64	0.
7	25.22	0.504	8.03	18.52	314.262	28.05	0.
7	29.00	0.355	43.37	99.99	1120.189	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 10/21/2020 1:44:48 PM
Printed by: wet


Ident: STD7
Analysis from: 10/21/2020 1:03:35 PM
File: _2020-10-21_

Last save: 10/21/2020 1:32:24 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 71953

Last save: 10/21/2020 1:00:30 PM

SAMPLE:
: AM/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.74	0.277	5.11	9.19	107.725	7.51	0.
2	8.85	0.213	7.38	13.29	108.249	7.55	0.
3	10.67	0.302	12.59	22.66	250.238	17.45	0.
4	13.74	0.312	7.17	12.90	155.727	10.86	0.
5	15.73	0.395	8.20	14.77	220.791	15.40	0.
6	21.80	0.495	4.71	8.48	189.836	13.24	0.
7	25.28	0.497	10.39	18.70	401.603	28.00	0.
7	29.00	0.356	55.55	100.00	1434.168	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 10/21/2020 3:45:13 PM
Printed by: wet


Ident: ICV
Analysis from: 10/21/2020 1:35:29 PM
File: _2020-10-21_

Last save: 10/21/2020 2:04:18 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 71954

Last save: 10/21/2020 1:32:24 PM

SAMPLE:
: AM/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.74	0.274	1.91	9.16	43.317	7.95	0.
2	8.86	0.221	2.70	12.93	40.982	7.52	0.
3	10.76	0.278	4.87	23.33	91.454	16.77	0.
4	13.78	0.315	2.64	12.64	57.860	10.61	0.
5	15.83	0.392	3.03	14.51	82.430	15.12	0.
6	21.94	0.508	1.90	9.12	80.470	14.76	0.
7	25.39	0.500	3.82	18.29	148.697	27.27	0.
7	29.00	0.355	20.87	99.99	545.210	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 10/21/2020 3:45:20 PM
Printed by: wet


Ident: ICB
Analysis from: 10/21/2020 2:21:03 PM
File: _2020-10-21_

Last save: 10/21/2020 2:49:52 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 71955

Last save: 10/21/2020 1:32:24 PM

SAMPLE:
: AM/AP
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 11/5/2020 9:07:43 AM
Printed by: wet


Ident: CCV
Analysis from: 11/4/2020 9:46:00 AM
File: _2020-11-04_

Last save: 11/4/2020 10:14:50 AM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72172

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.68	0.430	1.41	8.61	45.145	7.70	0.
2	8.97	0.252	2.26	13.80	43.303	7.39	0.
3	11.02	0.345	3.83	23.36	98.358	16.79	0.
4	14.57	0.422	1.94	11.84	63.370	10.82	0.
5	16.98	0.546	2.14	13.04	89.524	15.28	0.
6	20.97	0.626	1.60	9.76	83.707	14.29	0.
7	24.59	0.618	3.21	19.57	162.508	27.74	0.
7	29.00	0.463	16.38	99.98	585.914	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/5/2020 9:07:53 AM
Printed by: wet


Ident: CCB
Analysis from: 11/4/2020 10:19:24 AM
File: _2020-11-04_

Last save: 11/4/2020 10:48:12 AM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72173

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 11/5/2020 9:08:00 AM
Printed by: wet


Ident: LB111830BLW
Analysis from: 11/4/2020 10:51:18 AM
File: _2020-11-04_

Last save: 11/4/2020 2:23:56 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72174

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 4
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 11/5/2020 9:08:06 AM
Printed by: wet


Ident: LB111830BSW
Analysis from: 11/4/2020 11:23:12 AM
File: _2020-11-04_

Last save: 11/4/2020 2:23:56 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72175

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.68	0.418	1.45	8.94	44.458	7.90	0.
2	8.90	0.245	2.21	13.60	41.567	7.39	0.
3	10.87	0.331	3.80	23.46	94.658	16.83	0.
4	14.25	0.395	1.96	12.11	60.791	10.81	0.
5	16.54	0.504	2.18	13.46	85.762	15.25	0.
6	21.01	0.626	1.56	9.61	81.468	14.48	0.
7	24.60	0.616	3.05	18.80	153.822	27.34	0.
7	29.00	0.448	16.21	99.99	562.526	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/5/2020 9:09:14 AM
Printed by: wet


Ident: L4214-07
Analysis from: 11/4/2020 1:02:49 PM
File: _2020-11-04_

Last save: 11/4/2020 1:39:16 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72180

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.66	0.404	0.06	6.81	1.628	5.44	0.
2	8.87	0.257	0.11	13.08	2.244	7.50	0.
3	10.85	0.306	0.20	23.52	4.702	15.71	0.
4	14.18	0.398	0.10	11.27	2.988	9.98	0.
5	16.51	0.487	0.12	13.63	4.772	15.95	0.
6	21.13	0.719	0.11	12.39	6.190	20.68	0.
7	24.73	0.653	0.16	18.61	7.404	24.74	0.
7	29.00	0.461	0.86	99.32	29.928	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/5/2020 9:09:20 AM
Printed by: wet


Ident: L4214-08
Analysis from: 11/4/2020 1:42:59 PM
File: _2020-11-04_

Last save: 11/4/2020 2:11:48 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72181

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.77	0.439	0.12	7.09	3.475	5.78	0.
2	9.00	0.251	0.22	13.18	4.311	7.17	0.
3	10.98	0.309	0.40	23.76	9.545	15.87	0.
4	14.29	0.387	0.20	11.78	6.095	10.13	0.
5	16.59	0.472	0.25	15.01	9.759	16.22	0.
6	21.40	0.689	0.18	10.34	10.305	17.13	0.
7	25.03	0.646	0.32	18.66	16.667	27.71	0.
7	29.00	0.456	1.69	99.83	60.157	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/5/2020 9:09:26 AM
Printed by: wet


Ident: L4214-08RE
Analysis from: 11/4/2020 2:14:54 PM
File: _2020-11-04_

Last save: 11/4/2020 2:43:42 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72182

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.70	0.461	0.26	7.84	9.846	8.16	0.
2	8.91	0.260	0.44	13.43	8.732	7.23	0.
3	10.91	0.389	0.71	21.77	19.950	16.53	0.
4	14.15	0.379	0.42	12.75	12.666	10.49	0.
5	16.41	0.469	0.49	14.92	18.922	15.68	0.
6	21.29	0.646	0.33	10.03	17.840	14.78	0.
7	24.92	0.633	0.63	19.14	32.749	27.13	0.
7	29.00	0.462	3.27	99.89	120.706	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/5/2020 9:09:32 AM
Printed by: wet


Ident: CCV
Analysis from: 11/4/2020 4:34:36 PM
File: _2020-11-04_

Last save: 11/5/2020 9:04:28 AM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72186

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.66	0.414	1.56	8.89	46.816	8.08	0.
2	8.75	0.237	2.37	13.52	43.543	7.51	0.
3	10.61	0.314	4.12	23.47	98.733	17.04	0.
4	13.75	0.352	2.24	12.75	63.065	10.88	0.
5	15.87	0.445	2.55	14.51	90.435	15.60	0.
6	21.15	0.634	1.58	9.03	81.150	14.00	0.
7	24.78	0.620	3.13	17.82	155.811	26.88	0.
7	29.00	0.431	17.54	99.99	579.553	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/5/2020 9:09:39 AM
Printed by: wet


Ident: CCB
Analysis from: 11/4/2020 5:06:30 PM
File: _2020-11-04_

Last save: 11/4/2020 5:35:18 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72187

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD