

Clear table

Je-1
300.0
11/06/2020
AM

ident	concentra tion F-	concentra on CL-	concentra on NO2	concentra on BR-	concentra on NO3	concentra on HPO4	concentra on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-10-21_	10/21/20 9:52		AM/AP
STD2	0.1510	0.3230	0.3550	1.2840	0.2870	0.2070	1.7940	2020-10-21_	10/21/20 10:24		AM/AP
STD3	0.4240	0.6210	0.6280	2.1740	0.5530	1.0370	3.2180	2020-10-21_	10/21/20 10:55		AM/AP
STD4	0.7410	1.2040	1.2010	4.0050	1.0200	2.2220	6.0340	2020-10-21_	10/21/20 11:27		AM/AP
STD5	2.1010	2.9840	2.9670	9.7680	2.4740	5.4800	15.0250	2020-10-21_	10/21/20 11:59		AM/AP
STD6	3.9820	6.1310	6.1160	19.8290	5.0570	10.3340	30.7650	2020-10-21_	10/21/20 12:31		AM/AP
STD7	4.9840	7.6160	7.6410	25.3730	6.3360	12.4270	37.9110	2020-10-21_	10/21/20 13:03		AM/AP
ICV	2.0270	2.9300	2.8810	9.6560	2.3950	5.1140	14.2850	2020-10-21_	10/21/20 13:35		AM/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-10-21_	10/21/20 14:21		AM/AP
CCV	2.0850	3.0950	3.0520	10.4100	2.5840	5.1890	15.1790	2020-11-05_	11/5/20 9:23		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-11-05_	11/5/20 9:56		AM/AP
LB111845BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-11-05_	11/5/20 10:28		AM/AP
LB111845BSW	1.9860	2.9960	3.0130	10.2220	2.5480	4.8980	14.5930	2020-11-05_	11/5/20 11:00		AM/AP
L4634-01	0.0000	72.4480	0.0000	0.0000	1.9330	0.0000	19.8330	2020-11-05_	11/5/20 11:46		AM/AP
L4634-02	0.0000	65.3010	0.0000	0.0000	2.1420	0.0000	19.0640	2020-11-05_	11/5/20 12:18		AM/AP
L4650-01	0.0000	0.0000	0.0000	0.0000	0.2230	0.0000	0.0000	2020-11-05_	11/5/20 12:54		AM/AP
L4650-02	0.0000	0.0000	0.0000	0.0000	0.1140	0.0000	0.0000	2020-11-05_	11/5/20 13:26		AM/AP
L4650-03	0.0000	0.0000	0.0000	0.0000	0.1800	0.0000	0.0000	2020-11-05_	11/5/20 13:58		AM/AP
L4652-01	0.0000	0.0000	0.0000	0.0000	0.1740	0.0000	0.0000	2020-11-05_	11/5/20 14:30		AM/AP
L4634-01MS	2.6220	69.1600	3.0220	10.4970	4.3990	5.0100	33.4860	2020-11-05_	11/5/20 15:02		AM/AP
L4634-01MSD	2.4970	69.2260	3.0250	10.6740	4.4250	5.5440	33.4130	2020-11-05_	11/5/20 15:34		AM/AP
CCV	2.1380	3.0230	3.0310	10.3900	2.5830	5.1060	14.5340	2020-11-05_	11/5/20 16:05		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-11-05_	11/5/20 16:37		AM/AP
L4634-01DUP	0.0000	73.2900	0.0000	0.0000	1.9360	0.0000	19.4620	2020-11-05_	11/5/20 17:09		AM/AP
CCV	2.1490	3.0280	3.0950	10.3320	2.5550	5.0140	14.1780	2020-11-05_	11/5/20 17:41		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-11-05_	11/5/20 18:52		AM/AP

Report date: 11/6/2020 10:46:06 AM
Printed by: wet


Ident: CCV
Analysis from: 11/5/2020 9:23:36 AM
File: _2020-11-05_

Last save: 11/5/2020 9:52:24 AM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72198

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.64	0.440	1.38	8.35	44.587	7.73	0.
2	8.89	0.243	2.28	13.80	43.349	7.52	0.
3	10.90	0.331	3.85	23.31	97.136	16.85	0.
4	14.32	0.390	2.01	12.16	62.557	10.85	0.
5	16.63	0.500	2.24	13.56	89.061	15.45	0.
6	20.93	0.613	1.60	9.69	81.585	14.15	0.
7	24.54	0.613	3.16	19.12	158.272	27.45	0.
7	29.00	0.447	16.51	99.99	576.546	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:46:15 AM
Printed by: wet


Ident: CCB
Analysis from: 11/5/2020 9:56:59 AM
File: _2020-11-05_

Last save: 11/5/2020 10:25:48 AM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72199

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:46:24 AM
Printed by: wet


Ident: LB111845BLW
Analysis from: 11/5/2020 10:28:53 AM
File: _2020-11-05_

Last save: 11/5/2020 1:09:12 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72200

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 4
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:47:29 AM
Printed by: wet


Ident: LB111845BSW
Analysis from: 11/5/2020 11:00:48 AM
File: _2020-11-05_

Last save: 11/5/2020 1:09:12 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72201

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.67	0.424	1.40	8.48	42.417	7.59	0.
2	8.81	0.238	2.23	13.55	41.922	7.50	0.
3	10.74	0.318	3.90	23.68	95.854	17.16	0.
4	14.02	0.369	2.07	12.53	61.385	10.99	0.
5	16.22	0.472	2.32	14.09	87.782	15.71	0.
6	21.10	0.626	1.52	9.20	77.243	13.83	0.
7	24.71	0.614	3.04	18.43	152.000	27.21	0.
7	29.00	0.437	16.48	99.98	558.603	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:47:41 AM
Printed by: wet


Ident: L4634-01
Analysis from: 11/5/2020 11:46:45 AM
File: _2020-11-05_

Last save: 11/5/2020 12:15:34 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72202

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	8.86	0.217	61.67	91.37	1038.896	79.11	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	16.18	0.461	1.78	2.64	66.184	5.04	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	24.77	0.617	4.04	5.99	208.089	15.85	0.
7	29.00	0.185	67.49	99.99	1313.169	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:47:49 AM
Printed by: wet


Ident: L4634-02
Analysis from: 11/5/2020 12:18:39 PM
File: _2020-11-05_

Last save: 11/5/2020 12:47:28 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72203

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 13
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	8.78	0.207	57.53	90.74	936.304	77.40	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	16.03	0.455	2.00	3.15	73.519	6.08	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	24.67	0.622	3.87	6.11	199.854	16.52	0.
7	29.00	0.183	63.40	100.00	1209.677	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:47:58 AM
Printed by: wet


Ident: L4650-01
Analysis from: 11/5/2020 12:54:32 PM
File: _2020-11-05_

Last save: 11/5/2020 1:23:20 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72204

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	0.00	0.000	0.00	0.00	0.000	0.00	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	16.09	0.446	0.15	99.43	6.156	100.00	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	29.00	0.064	0.15	99.43	6.156	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:48:04 AM
Printed by: wet


Ident: L4650-02
Analysis from: 11/5/2020 1:26:26 PM
File: _2020-11-05_

Last save: 11/5/2020 1:55:14 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72205

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 15
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	0.00	0.000	0.00	0.00	0.000	0.00	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	16.27	0.446	0.06	98.65	2.315	100.00	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	29.00	0.064	0.06	98.65	2.315	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:48:11 AM
Printed by: wet


Ident: L4650-03
Analysis from: 11/5/2020 1:58:20 PM
File: _2020-11-05_

Last save: 11/5/2020 2:27:08 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72206

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 16
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	0.00	0.000	0.00	0.00	0.000	0.00	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	16.81	0.471	0.11	99.78	4.625	100.00	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	29.00	0.067	0.11	99.78	4.625	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:48:21 AM
Printed by: wet


Ident: L4652-01
Analysis from: 11/5/2020 2:30:14 PM
File: _2020-11-05_

Last save: 11/5/2020 2:59:02 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72207

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 17
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	0.00	0.000	0.00	0.00	0.000	0.00	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	15.94	0.439	0.12	99.79	4.413	100.00	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	29.00	0.063	0.12	99.79	4.413	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:48:30 AM
Printed by: wet


Ident: L4634-01MS
Analysis from: 11/5/2020 3:02:08 PM
File: _2020-11-05_

Last save: 11/5/2020 3:30:58 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72208

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.73	0.463	1.68	2.10	56.273	3.14	0.
2	8.79	0.214	59.62	74.45	991.691	55.31	0.
3	10.62	0.381	3.49	4.36	96.138	5.36	0.
4	13.72	0.332	2.33	2.90	63.094	3.52	0.
5	15.72	0.426	4.42	5.52	152.782	8.52	0.
6	21.20	0.663	1.47	1.84	78.910	4.40	0.
7	24.79	0.609	7.06	8.82	354.236	19.76	0.
7	29.00	0.441	80.08	99.99	1793.126	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:48:38 AM
Printed by: wet


Ident: L4634-01MSD
Analysis from: 11/5/2020 3:34:03 PM
File: _2020-11-05_

Last save: 11/5/2020 4:02:52 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72209

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 12
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.68	0.459	1.61	2.00	53.561	2.97	0.
2	8.72	0.213	60.08	74.40	992.648	55.13	0.
3	10.55	0.377	3.53	4.37	96.245	5.34	0.
4	13.64	0.335	2.35	2.91	64.201	3.57	0.
5	15.65	0.423	4.47	5.53	153.689	8.54	0.
6	21.19	0.660	1.63	2.01	86.897	4.83	0.
7	24.78	0.609	7.09	8.78	353.448	19.63	0.
7	29.00	0.439	80.76	100.01	1800.689	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:48:44 AM
Printed by: wet


Ident: CCV
Analysis from: 11/5/2020 4:05:57 PM
File: _2020-11-05_

Last save: 11/5/2020 4:34:46 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72210

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.68	0.431	1.46	8.46	45.723	8.05	0.
2	8.72	0.234	2.30	13.33	42.308	7.45	0.
3	10.54	0.308	4.06	23.52	96.437	16.99	0.
4	13.58	0.343	2.25	13.03	62.433	11.00	0.
5	15.65	0.432	2.56	14.81	89.016	15.68	0.
6	21.19	0.631	1.62	9.36	80.351	14.16	0.
7	24.82	0.631	3.01	17.46	151.367	26.67	0.
7	29.00	0.430	17.26	99.98	567.635	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:48:53 AM
Printed by: wet


Ident: CCB
Analysis from: 11/5/2020 4:37:51 PM
File: _2020-11-05_

Last save: 11/5/2020 5:06:40 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72211

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:49:02 AM
Printed by: wet


Ident: L4634-01DUP
Analysis from: 11/5/2020 5:09:45 PM
File: _2020-11-05_

Last save: 11/5/2020 5:38:34 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72212

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 10
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	8.86	0.214	63.25	91.42	1050.989	79.54	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	15.83	0.417	1.94	2.80	66.311	5.02	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	25.06	0.624	4.00	5.78	204.112	15.45	0.
7	29.00	0.179	69.18	99.99	1321.412	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:49:12 AM
Printed by: wet


Ident: CCV
Analysis from: 11/5/2020 5:41:39 PM
File: _2020-11-05_

Last save: 11/5/2020 6:10:28 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72213

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.67	0.427	1.47	8.51	45.970	8.16	0.
2	8.69	0.233	2.31	13.35	42.387	7.52	0.
3	10.50	0.312	4.09	23.62	98.574	17.49	0.
4	13.57	0.337	2.27	13.11	62.068	11.01	0.
5	15.61	0.419	2.58	14.92	88.045	15.62	0.
6	21.29	0.649	1.57	9.08	78.976	14.01	0.
7	24.98	0.624	3.01	17.41	147.556	26.18	0.
7	29.00	0.429	17.31	99.99	563.575	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 11/6/2020 10:49:24 AM
Printed by: wet


Ident: CCB
Analysis from: 11/5/2020 6:52:44 PM
File: _2020-11-05_

Last save: 11/5/2020 7:21:32 PM

Method: ANIONS 10-21-20.mtw
Run operator: wet
Analysis number: 72217

Last save: 11/2/2020 10:43:58 AM

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD