

LB111973


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37867



A= 0.0000e+000

B= 2.6441e-004 *slope*C= 1.4485e-002 *y-intercept*

Rho= 0.9999758

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.024	0.024	37	0.000	37					1
0.05	0.050					244					9
0.20	0.200	0.217	0.017	765	0.0 %	765					-2
2.50	2.500	2.456	-0.044	9232	0.0 %	9232					0
5.00	5.000	4.996	-0.004	18839	0.0 %	18839					0
7.50	7.500	7.481	-0.019	28238	0.0 %	28238					0
10.0	10.000	10.027	0.027	37867	0.0 %	37867					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	37	-		0 ISM02.4	PPB	11/15/2020 14:08	Std
	0.2 S0.2	765	-		0.2 ISM02.4	PPB	11/15/2020 14:10	Std
	2.5 S2.5	9232	-		2.5 ISM02.4	PPB	11/15/2020 14:12	Std
	5 S5.0	18839	-		5 ISM02.4	PPB	11/15/2020 14:14	Std
	7.5 S7.5	28238	-		7.5 ISM02.4	PPB	11/15/2020 14:17	Std
	10 S10.0	37867	-		10 ISM02.4	PPB	11/15/2020 14:19	Std
ICV	ICV	14993	3.9788 -		ISM02.4	PPB	11/15/2020 14:24	SMPL
ICB	ICB	-64	-0.0024 -		ISM02.4	PPB	11/15/2020 14:27	SMPL
CCV034	CCV034	18699	4.9587 -		ISM02.4	PPB	11/15/2020 14:29	SMPL
CCB034	CCB034	-82	-0.0072 -		ISM02.4	PPB	11/15/2020 14:31	SMPL
PB133013BL	PBS013	81	0.0359 -		ISM02.4	PPB	11/15/2020 14:33	SMPL
L4760-01	MCOAA0	1332	0.3667 -		ISM02.4	PPB	11/15/2020 14:40	SMPL
L4760-02	MCOAA0D	1169	0.3236 -		ISM02.4	PPB	11/15/2020 14:43	SMPL
L4760-03	MCOAA0S	9570	2.5449 -		ISM02.4	PPB	11/15/2020 14:48	SMPL
L4760-04	MCOAA1	2165	0.5869 -		ISM02.4	PPB	11/15/2020 14:50	SMPL
L4760-05	MCOAA2	1720	0.4693 -		ISM02.4	PPB	11/15/2020 14:52	SMPL
L4760-06	MCOAA3	1411	0.3876 -		ISM02.4	PPB	11/15/2020 14:55	SMPL
L4760-07	MCOAA4	1187	0.3283 -		ISM02.4	PPB	11/15/2020 14:57	SMPL
L4760-08	MCOAA5	3486	0.9362 -		ISM02.4	PPB	11/15/2020 14:59	SMPL
L4760-09	MCOAA6	1399	0.3844 -		ISM02.4	PPB	11/15/2020 15:01	SMPL
L4760-10	MCOAA7	3098	0.8336 -		ISM02.4	PPB	11/15/2020 15:04	SMPL
L4760-11	MCOAA8	1311	0.3611 -		ISM02.4	PPB	11/15/2020 15:06	SMPL
L4760-12	MCOAA9	1680	0.4587 -		ISM02.4	PPB	11/15/2020 15:08	SMPL
L4760-13	MCOAB0	1309	0.3606 -		ISM02.4	PPB	11/15/2020 15:10	SMPL
L4760-14	MCOAB1	1590	0.4349 -		ISM02.4	PPB	11/15/2020 15:13	SMPL
L4760-15	MCOAB2	1386	0.381 -		ISM02.4	PPB	11/15/2020 15:15	SMPL
L4760-16	MCOAB3	1596	0.4365 -		ISM02.4	PPB	11/15/2020 15:17	SMPL
L4760-17	MCOAB4	1142	0.3164 -		ISM02.4	PPB	11/15/2020 15:19	SMPL
L4760-18	MCOAB5	1226	0.3387 -		ISM02.4	PPB	11/15/2020 15:22	SMPL
L4760-19	MCOAB6	1434	0.3936 -		ISM02.4	PPB	11/15/2020 15:24	SMPL
CCV035	CCV035	19087	5.0613 -		ISM02.4	PPB	11/15/2020 15:26	SMPL
CCB035	CCB035	-54	0.0002 -		ISM02.4	PPB	11/15/2020 15:29	SMPL
L4760-20	MCOAB7	1632	0.446 -		ISM02.4	PPB	11/15/2020 15:31	SMPL
L4760-21	MCOAB8	2613	0.7054 -		ISM02.4	PPB	11/15/2020 15:33	SMPL
L4760-22	MCOAB9	1611	0.4404 -		ISM02.4	PPB	11/15/2020 15:35	SMPL
PB133014BL	PBS014	73	0.0338 -		ISM02.4	PPB	11/15/2020 15:38	SMPL
L4761-01	MCOAC2	1445	0.3966 -		ISM02.4	PPB	11/15/2020 15:45	SMPL
L4761-02	MCOAC2D	1429	0.3923 -		ISM02.4	PPB	11/15/2020 15:47	SMPL
L4761-03	MCOAC2S	10181	2.7064 -		ISM02.4	PPB	11/15/2020 15:55	SMPL
L4761-04	MCOAC3	1413	0.3881 -		ISM02.4	PPB	11/15/2020 15:57	SMPL
L4761-05	MCOAC4	1104	0.3064 -		ISM02.4	PPB	11/15/2020 15:59	SMPL
L4761-06	MCOAC6	2411	0.652 -		ISM02.4	PPB	11/15/2020 16:02	SMPL
L4761-07	MCOAC7	2172	0.5888 -		ISM02.4	PPB	11/15/2020 16:04	SMPL
L4761-08	MCOAC8	1538	0.4211 -		ISM02.4	PPB	11/15/2020 16:06	SMPL
L4761-09	MCOAC9	1486	0.4074 -		ISM02.4	PPB	11/15/2020 16:08	SMPL
CCV036	CCV036	18534	4.9151 -		ISM02.4	PPB	11/15/2020 16:11	SMPL

LB111973
INSTRUMENT ID : CV1

CCB036 CCB036 -38 0.0044 - ISM02.4 PPB 11/15/2020 16:13 SMPL