

LB112191

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AM

Instrument ID : Konelab

12/2/2020 13:05

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.934	0.0	0.088	
ICB1	-0.237	0.0	0.001	
CCV1	240.647	0.0	0.222	
CCB1	-0.313	0.0	0.001	
PB133353BL	-0.352	0.0	0.000	
PB133353BS	96.254	0.0	0.089	
LOWPB133353	9.363	0.0	0.009	
HIGHPB133353	476.930	0.0	0.440	
L4909-01	-0.206	0.0	0.001	
L4909-01DUP	-0.281	0.0	0.001	
L4909-01MS	34.083	0.0	0.032	
L4909-01MSD	33.889	0.0	0.032	
L4909-02	0.745	0.0	0.001	
L4910-03	39.525	0.0	0.037	
CCV2	240.199	0.0	0.222	
CCB2	-0.449	0.0	0.000	
L4910-05	0.006	0.0	0.001	
L4911-01	0.790	0.0	0.002	
L4911-02	-0.318	0.0	0.001	
L4911-03	0.427	0.0	0.001	
L4911-04	-0.282	0.0	0.001	
L4911-05	0.822	0.0	0.002	
CCV3	242.445	0.0	0.224	
CCB3	-0.009	0.0	0.001	

N	24
Mean	62.859
SD	119.6479
CV%	190.34

Aquakem v. 7.2AQ1

Results from time period:

Wed Dec 02 11:08:31 2020

Wed Dec 02 12:58:25 2020

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.7463	µg/l	12/2/2020 11:08:31	
5.OPPBCN	A	Total CN	P	4.3411	µg/l	12/2/2020 11:08:32	
10PPBCN	A	Total CN	P	9.8596	µg/l	12/2/2020 11:08:33	
50PPBCN	A	Total CN	P	50.0222	µg/l	12/2/2020 11:08:34	
100PPBCN	A	Total CN	P	100.5349	µg/l	12/2/2020 11:08:35	
250PPBCN	A	Total CN	P	252.1765	µg/l	12/2/2020 11:08:36	
500PPBCN	A	Total CN	P	498.812	µg/l	12/2/2020 11:08:37	
ICV1	S	Total CN	P	94.9343	µg/l	12/2/2020 12:19:04	
ICB1	S	Total CN	P	-0.2366	µg/l	12/2/2020 12:19:05	
CCV1	S	Total CN	P	240.6475	µg/l	12/2/2020 12:19:08	
CCB1	S	Total CN	P	-0.3132	µg/l	12/2/2020 12:19:09	
PB133353BL	S	Total CN	P	-0.3522	µg/l	12/2/2020 12:19:11	
PB133353BS	S	Total CN	P	96.2543	µg/l	12/2/2020 12:19:12	
LOWPB133353	S	Total CN	P	9.3632	µg/l	12/2/2020 12:19:13	
HIGHPB133353	S	Total CN	P	476.9303	µg/l	12/2/2020 12:26:36	
L4909-01	S	Total CN	P	-0.2058	µg/l	12/2/2020 12:26:37	
L4909-01DUP	S	Total CN	P	-0.281	µg/l	12/2/2020 12:26:38	
L4909-01MS	S	Total CN	P	34.0833	µg/l	12/2/2020 12:26:39	
L4909-01MSD	S	Total CN	P	33.8886	µg/l	12/2/2020 12:26:40	
L4909-02	S	Total CN	P	0.7448	µg/l	12/2/2020 12:26:43	
L4910-03	S	Total CN	P	39.5249	µg/l	12/2/2020 12:26:44	
CCV2	S	Total CN	P	240.1988	µg/l	12/2/2020 12:26:46	
CCB2	S	Total CN	P	-0.4488	µg/l	12/2/2020 12:34:11	
L4910-05	S	Total CN	P	0.0064	µg/l	12/2/2020 12:34:13	
L4911-01	S	Total CN	P	0.7904	µg/l	12/2/2020 12:34:14	
L4911-02	S	Total CN	P	-0.3179	µg/l	12/2/2020 12:34:15	
L4911-03	S	Total CN	P	0.4273	µg/l	12/2/2020 12:34:16	
L4911-04	S	Total CN	P	-0.2822	µg/l	12/2/2020 12:34:17	
L4911-05	S	Total CN	P	0.8221	µg/l	12/2/2020 12:34:18	
CCV3	S	Total CN	P	242.4447	µg/l	12/2/2020 12:34:19	
CCB3	S	Total CN	P	-0.0094	µg/l	12/2/2020 12:34:20	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AM

Instrument ID : Konelab

12/2/2020 11:11


Test Total CN

Accepted 12/2/2020 11:11

Factor 1086
Bias 0.001

Coeff. of det. 0.999964

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.000	-0.7463	0.0000	
2	5.0PPBCN	0.005	4.3411	5.0000	
3	10PPBCN	0.010	9.8596	10.0000	
4	50PPBCN	0.047	50.0222	50.0000	
5	100PPBCN	0.093	100.5349	100.0000	
6	250PPBCN	0.233	252.1765	250.0000	
7	500PPBCN	0.460	498.8120	500.0000	