

IC-1
300.0
12/15/2020
AM


Clear table

ident	concentra tion F-	concentrati on CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-12-02_	12/2/20 10:09		AM/AP
STD2	0.1840	0.3570	0.3530	1.1630	0.3340	0.3530	1.8540	2020-12-02_	12/2/20 10:42		AM/AP
STD3	0.3610	0.6160	0.6170	2.0920	0.5430	1.1460	3.1500	2020-12-02_	12/2/20 11:14		AM/AP
STD4	0.7540	1.1790	1.1710	3.9500	0.9920	2.4780	5.8570	2020-12-02_	12/2/20 11:46		AM/AP
STD5	2.0080	2.8890	2.8540	9.6530	2.3670	5.7690	14.3300	2020-12-02_	12/2/20 12:19		AM/AP
STD6	3.9410	5.9670	6.0000	19.9680	4.8850	10.5350	29.9820	2020-12-02_	12/2/20 12:51		AM/AP
STD7	4.9910	7.5750	7.5540	25.2520	6.3060	12.3390	37.9140	2020-12-02_	12/2/20 13:24		AM/AP
ICV	1.9730	2.8850	2.9050	9.5930	2.3960	5.2260	14.2020	2020-12-02_	12/2/20 13:56		AM/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-12-02_	12/2/20 14:28		AM/AP
CCV	2.0730	2.9730	2.9490	9.6920	2.4090	6.3400	14.3090	2020-12-14_	12/14/20 14:52		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-12-14_	12/14/20 15:25		AM/AP
LB112368BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-12-14_	12/14/20 15:58		AM/AP
LB112368BSW	2.1090	3.0310	3.0520	9.8010	2.4700	6.4940	14.8000	2020-12-14_	12/14/20 16:31		AM/AP
L5086-01	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-12-14_	12/14/20 17:04		AM/AP
L5086-01DUP	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-12-14_	12/14/20 17:37		AM/AP
L5086-01MS	2.1120	3.0850	2.9510	9.8210	2.4720	6.4300	14.7730	2020-12-14_	12/14/20 18:09		AM/AP
L5086-01MSD	2.0630	2.9700	2.9500	9.7800	2.4960	6.1440	14.7470	2020-12-14_	12/14/20 18:42		AM/AP
CCV	2.0180	2.9120	2.9590	9.7510	2.4250	6.4970	14.3840	2020-12-14_	12/14/20 19:15		AM/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2020-12-14_	12/14/20 19:48		AM/AP

IC-1
300.0
12/07/2020
AM

CALIBRATION OF COMPONENT F-

Method: ANIONS 12-02-20.mtw
Equation: $Q = 0.864883 \cdot A + 0.961723$
RSD: 1.505 %
Correlation coefficient: 0.999904




K3 = 0 K2 = 0 K1 = 0.864883 K0 = 0.961723
Base: Area
Ref.channel: Cond
ISTD:
Formula: Linear
Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		No
2	0.1661	4.301	0.2	20	5.822	Yes	
3	0.3632	7.742	0.4	20	5.822	Yes	
4	0.7806	16.47	0.8	20	5.822	Yes	
5	1.984	45.76	2	20	5.822	Yes	
6	4.055	91.49	4	20	5.822	Yes	
7	5.22	114.3	5	20	5.822	Yes	

CALIBRATION OF COMPONENT CL-

Method: ANIONS 12-02-20.mtw
 Equation: $Q = 1.28717 \cdot A + 1.96684$
 RSD: 2.705 %
 Correlation coefficient: 0.999691




K3 = 0 K2 = 0 K1 = 1.28717 K0 = 1.96684
 Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		No
2	0.3639	4.253	0.3	20	9.31	Yes	
3	0.7336	8.285	0.6	20	9.31	Yes	
4	1.486	16.85	1.2	20	9.31	Yes	
5	3.846	43.23	3	20	9.31	Yes	
6	8.409	91.04	6	20	9.31	Yes	
7	10.8	116.2	7.5	20	9.31	Yes	

CALIBRATION OF COMPONENT NO2

Method: ANIONS 12-02-20.mtw
Equation: $Q = 0.57842 \cdot A + 1.85748$
RSD: 2.383 %
Correlation coefficient: 0.999760




K3 = 0 K2 = 0 K1 = 0.57842 K0 = 1.85748
Base: Area
Ref.channel: Cond
ISTD:
Formula: Linear
Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0	0	0	0	0			No
2	0.5783	9.2	0.3	20	11.45			Yes
3	1.161	18.72	0.6	20	11.45			Yes
4	2.306	37.82	1.2	20	11.45			Yes
5	5.746	96.45	3	20	11.45			Yes
6	11.64	203.7	6	20	11.45			Yes
7	14.49	258	7.5	20	11.45			Yes

CALIBRATION OF COMPONENT BR-

Method: ANIONS 12-02-20.mtw
 Equation: $Q = 2.90254 \cdot A + 6.15045$
 RSD: 2.698 %
 Correlation coefficient: 0.999692




K3 = 0 K2 = 0 K1 = 2.90254 K0 = 6.15045

Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0	0	0	0	0			No
2	0.3437	6.105	1	20	15.03			Yes
3	0.7157	12.61	2	20	15.03			Yes
4	1.473	25.24	4	20	15.03			Yes
5	3.863	63.94	10	20	15.03			Yes
6	8.545	134.7	20	20	15.03			Yes
7	11.03	171.9	25	20	15.03			Yes

CALIBRATION OF COMPONENT NO3

Method: ANIONS 12-02-20.mtw
 Equation: $Q = 0.509768 \cdot A + 1.59737$
 RSD: 2.893 %
 Correlation coefficient: 0.999646



K3 = 0 K2 = 0 K1 = 0.509768 K0 = 1.59737

Base: Area

Ref.channel: Cond

ISTD:

Formula: Linear


Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		No
2	0.4111	9.135	0.25	20	17.65	Yes	
3	0.8474	17.72	0.5	20	17.65	Yes	
4	1.727	35.56	1	20	17.65	Yes	
5	4.457	90.37	2.5	20	17.65	Yes	
6	9.681	192.2	5	20	17.65	Yes	
7	12.42	244.3	6.25	20	17.65	Yes	

±□□□

CALIBRATION OF COMPONENT HPO4

Method: ANIONS 12-02-20.mtw
 Equation: $Q = 1.0249 \cdot A - 7.74053$
 RSD: 5.063 %
 Correlation coefficient: 0.998916




K3 = 0 K2 = 0 K1 = 1.0249 K0 = -7.74053

Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0	0	0	0	0			No
2	0.2964	10.99	0.5	20	23.03			Yes
3	0.7489	26.23	1	20	23.03			Yes
4	1.525	51.13	2	20	23.03			Yes
5	3.509	110.9	5	20	23.03			Yes
6	6.863	202.7	10	20	23.03			Yes
7	8.544	248.3	12.5	20	23.03			Yes

CALIBRATION OF COMPONENT SO4

Method: ANIONS 12-02-20.mtw
 Equation: $Q = 1.64644 \cdot A + 11.1988$
 RSD: 3.355 %
 Correlation coefficient: 0.999524



K3 = 0 K2 = 0 K1 = 1.64644 K0 = 11.1988

Base: Area

Ref.channel: Cond

ISTD:

Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File	
1	0	0	0	0	0			No
2	0.5957	16.71	1.5	20	26.99			Yes
3	1.179	32.58	3	20	26.99			Yes
4	2.37	64.55	6	20	26.99			Yes
5	6.138	165.9	15	20	26.99			Yes
6	13.37	355.4	30	20	26.99			Yes
7	17.1	453.8	37.5	20	26.99			Yes

Report date: 12/7/2020 9:00:15 AM
Printed by: wet


Ident: STD1
Analysis from: 12/2/2020 10:09:42 AM
File: _2020-12-02_

Last save: 12/2/2020 10:39:00 AM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72516

Last save: 12/2/2020 9:54:32 AM

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/7/2020 9:00:22 AM
Printed by: wet


Ident: STD2
Analysis from: 12/2/2020 10:42:06 AM
File: _2020-12-02_

Last save: 12/2/2020 11:11:24 AM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72517

Last save: 12/2/2020 10:39:00 AM

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.81	0.308	0.17	6.01	4.301	7.09	0.
2	9.18	0.172	0.36	13.20	4.253	7.01	0.
3	11.24	0.223	0.58	20.96	9.200	15.16	0.
4	14.68	0.262	0.34	12.45	6.105	10.06	0.
5	17.15	0.309	0.41	14.90	9.135	15.05	0.
6	22.80	0.449	0.30	10.73	10.989	18.10	0.
7	26.80	0.421	0.60	21.61	16.713	27.54	0.
7	29.50	0.306	2.75	99.86	60.694	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 12/7/2020 9:00:27 AM
Printed by: wet


Ident: STD3
Analysis from: 12/2/2020 11:14:30 AM
File: _2020-12-02_

Last save: 12/2/2020 11:43:48 AM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72518

Last save: 12/2/2020 11:35:38 AM

SAMPLE:
: AM/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.80	0.277	0.36	6.31	7.742	6.25	0.
2	9.17	0.168	0.73	12.76	8.285	6.69	0.
3	11.20	0.224	1.16	20.19	18.715	15.11	0.
4	14.60	0.256	0.72	12.44	12.611	10.18	0.
5	17.02	0.302	0.85	14.73	17.718	14.30	0.
6	22.80	0.437	0.75	13.01	26.225	21.17	0.
7	26.82	0.418	1.18	20.50	32.579	26.30	0.
7	29.50	0.297	5.75	99.94	123.876	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 12/7/2020 9:00:33 AM
Printed by: wet


Ident: STD4
Analysis from: 12/2/2020 11:46:54 AM
File: _2020-12-02_

Last save: 12/2/2020 12:16:12 PM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72519

Last save: 12/2/2020 11:43:48 AM

SAMPLE:
: AM/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.81	0.265	0.78	6.69	16.469	6.65	0.
2	9.15	0.168	1.49	12.74	16.851	6.81	0.
3	11.16	0.224	2.31	19.77	37.817	15.27	0.
4	14.52	0.253	1.47	12.62	25.244	10.19	0.
5	16.90	0.298	1.73	14.80	35.564	14.36	0.
6	22.82	0.424	1.52	13.07	51.128	20.65	0.
7	26.84	0.413	2.37	20.31	64.550	26.07	0.
7	29.50	0.292	11.67	100.00	247.624	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 12/7/2020 9:00:38 AM
Printed by: wet


Ident: STD5
Analysis from: 12/2/2020 12:19:18 PM
File: _2020-12-02_

Last save: 12/2/2020 12:48:36 PM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72520

Last save: 12/2/2020 12:16:12 PM

SAMPLE:
: AM/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.81	0.286	1.98	6.72	45.764	7.42	0.
2	9.12	0.164	3.85	13.02	43.232	7.01	0.
3	11.09	0.226	5.75	19.45	96.447	15.64	0.
4	14.42	0.242	3.86	13.08	63.942	10.37	0.
5	16.75	0.292	4.46	15.09	90.370	14.66	0.
6	22.77	0.409	3.51	11.88	110.869	17.98	0.
7	26.81	0.408	6.14	20.78	165.904	26.91	0.
7	29.50	0.290	29.54	100.00	616.527	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 12/7/2020 9:00:44 AM
Printed by: wet


Ident: STD6
Analysis from: 12/2/2020 12:51:42 PM
File: _2020-12-02_

Last save: 12/2/2020 1:21:00 PM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72521

Last save: 12/2/2020 12:48:36 PM

SAMPLE:
: AM/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.86	0.293	4.05	6.48	91.493	7.20	0.
2	9.13	0.160	8.41	13.44	91.044	7.16	0.
3	11.05	0.241	11.64	18.61	203.709	16.02	0.
4	14.39	0.233	8.55	13.66	134.718	10.60	0.
5	16.64	0.286	9.68	15.47	192.240	15.12	0.
6	22.78	0.394	6.86	10.97	202.703	15.94	0.
7	26.83	0.403	13.37	21.37	355.396	27.96	0.
7	29.50	0.287	62.57	100.00	1271.304	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 12/7/2020 9:00:50 AM
 Printed by: wet


Ident: STD7
 Analysis from: 12/2/2020 1:24:06 PM
 File: _2020-12-02_

Last save: 12/2/2020 1:53:24 PM

Method: ANIONS 12-02-20.mtw
 Run operator: wet
 Analysis number: 72522

Last save: 12/2/2020 1:21:00 PM

SAMPLE:
 : AM/AP
 Vial number: 8
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.85	0.278	5.22	6.56	114.303	7.11	0.
2	9.12	0.159	10.80	13.57	116.173	7.23	0.
3	11.02	0.246	14.49	18.20	257.976	16.06	0.
4	14.35	0.231	11.03	13.85	171.877	10.70	0.
5	16.57	0.283	12.42	15.60	244.292	15.20	0.
6	22.76	0.390	8.54	10.73	248.337	15.46	0.
7	26.80	0.402	17.10	21.49	453.755	28.24	0.
7	29.50	0.284	79.60	100.00	1606.713	100.00	0.

This report has been created by IC Net
 METROHM LTD

Report date: 12/7/2020 9:00:55 AM
Printed by: wet


Ident: ICV
Analysis from: 12/2/2020 1:56:30 PM
File: _2020-12-02_

Last save: 12/2/2020 2:25:48 PM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72523

Last save: 12/2/2020 1:53:24 PM

SAMPLE:
: AM/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.82	0.278	2.07	6.89	44.516	7.24	0.
2	9.11	0.164	3.91	13.04	43.303	7.04	0.
3	11.06	0.228	5.83	19.43	97.231	15.81	0.
4	14.37	0.239	3.95	13.18	63.985	10.40	0.
5	16.67	0.289	4.57	15.23	90.858	14.77	0.
6	22.80	0.417	3.44	11.47	109.538	17.81	0.
7	26.84	0.405	6.22	20.74	165.716	26.94	0.
7	29.50	0.289	29.97	99.99	615.147	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 12/7/2020 9:01:01 AM
Printed by: wet


Ident: ICB
Analysis from: 12/2/2020 2:28:54 PM
File: _2020-12-02_

Last save: 12/2/2020 2:58:12 PM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72524

Last save: 12/2/2020 1:53:24 PM

SAMPLE:
: AM/AP
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/15/2020 4:26:25 PM
Printed by: wet


Ident: CCV
Analysis from: 12/14/2020 2:52:35 PM
File: _2020-12-14_

Last save: 12/15/2020 4:26:16 PM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72669

Last save: 12/11/2020 9:10:14 PM

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.93	0.327	1.72	5.58	46.819	7.26	0.
2	9.25	0.171	3.90	12.66	44.660	6.93	0.
3	11.20	0.221	6.34	20.58	98.744	15.32	0.
4	14.42	0.240	3.99	12.95	64.667	10.03	0.
5	16.65	0.286	4.63	15.05	91.388	14.18	0.
6	22.72	0.427	3.92	12.73	131.265	20.37	0.
7	26.35	0.403	6.29	20.45	167.013	25.91	0.
7	30.00	0.296	30.79	100.00	644.556	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 12/15/2020 4:26:32 PM
Printed by: wet


Ident: CCB
Analysis from: 12/14/2020 3:25:29 PM
File: _2020-12-14_

Last save: 12/15/2020 4:26:16 PM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72670

Last save: 12/11/2020 9:10:14 PM

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/15/2020 4:26:40 PM
Printed by: wet


Ident: LB112368BLW
Analysis from: 12/14/2020 3:58:23 PM
File: _2020-12-14_

Last save: 12/14/2020 4:28:10 PM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72671

Last save: 12/11/2020 9:10:14 PM

SAMPLE:
: AM/AP
Vial number: 4
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/15/2020 4:30:56 PM
 Printed by: wet


Ident: LB112368BSW
 Analysis from: 12/14/2020 4:31:16 PM
 File: _2020-12-14_

Last save: 12/15/2020 4:30:52 PM

Method: ANIONS 12-02-20.mtw
 Run operator: wet
 Analysis number: 72672

Last save: 12/11/2020 9:10:14 PM

SAMPLE:
 : AM/AP
 Vial number: 5
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.87	0.326	1.80	5.72	47.647	7.20	0.
2	9.16	0.171	3.97	12.64	45.564	6.88	0.
3	11.13	0.224	6.37	20.29	102.324	15.46	0.
4	14.37	0.238	4.08	12.99	65.414	9.88	0.
5	16.60	0.289	4.74	15.08	93.781	14.17	0.
6	22.67	0.424	4.04	12.86	134.269	20.28	0.
7	26.30	0.411	6.41	20.42	172.975	26.13	0.
7	30.00	0.298	31.40	100.00	661.974	100.00	0.

This report has been created by IC Net
 METROHM LTD

Report date: 12/15/2020 4:26:55 PM
Printed by: wet


Ident: L5086-01
Analysis from: 12/14/2020 5:04:10 PM
File: _2020-12-14_

Last save: 12/14/2020 5:33:58 PM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72673

Last save: 12/14/2020 4:23:50 PM

SAMPLE:
: AM/AP
Vial number: 33
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/15/2020 4:27:03 PM
Printed by: wet


Ident: L5086-01DUP
Analysis from: 12/14/2020 5:37:03 PM
File: _2020-12-14_

Last save: 12/14/2020 6:06:52 PM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72674

Last save: 12/14/2020 4:23:50 PM

SAMPLE:
: AM/AP
Vial number: 34
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/15/2020 4:27:09 PM
Printed by: wet


Ident: L5086-01MS
Analysis from: 12/14/2020 6:09:56 PM
File: _2020-12-14_

Last save: 12/14/2020 6:39:44 PM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72675

Last save: 12/14/2020 4:23:50 PM

SAMPLE:
: AM/AP
Vial number: 35
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.89	0.328	1.79	5.74	47.724	7.25	0.
2	9.18	0.171	4.04	12.95	46.400	7.05	0.
3	11.16	0.220	6.22	19.91	98.832	15.02	0.
4	14.39	0.239	4.07	13.03	65.553	9.96	0.
5	16.63	0.287	4.76	15.25	93.865	14.26	0.
6	22.81	0.436	3.89	12.46	133.021	20.21	0.
7	26.52	0.407	6.45	20.66	172.646	26.24	0.
7	30.00	0.298	31.23	99.99	658.040	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 12/15/2020 4:27:15 PM
Printed by: wet


Ident: L5086-01MSD
Analysis from: 12/14/2020 6:42:50 PM
File: _2020-12-14_

Last save: 12/14/2020 7:12:38 PM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72676

Last save: 12/14/2020 4:23:50 PM

SAMPLE:
: AM/AP
Vial number: 36
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.79	0.324	1.88	6.06	46.601	7.17	0.
2	9.00	0.168	3.95	12.70	44.617	6.87	0.
3	10.93	0.220	6.17	19.85	98.785	15.20	0.
4	14.18	0.239	4.05	13.04	65.269	10.04	0.
5	16.43	0.289	4.75	15.29	94.777	14.58	0.
6	22.43	0.425	3.83	12.33	127.454	19.61	0.
7	26.05	0.406	6.45	20.73	172.341	26.52	0.
7	30.00	0.296	31.10	99.99	649.844	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 12/15/2020 4:27:23 PM
 Printed by: wet


Ident: CCV
 Analysis from: 12/14/2020 7:15:44 PM
 File: _2020-12-14_

Last save: 12/14/2020 7:45:32 PM

Method: ANIONS 12-02-20.mtw
 Run operator: wet
 Analysis number: 72677

Last save: 12/14/2020 4:23:50 PM

SAMPLE:
 : AM/AP
 Vial number: 2
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.87	0.324	1.81	5.86	45.544	7.03	0.
2	9.08	0.167	3.88	12.59	43.715	6.75	0.
3	11.03	0.219	6.22	20.17	99.095	15.30	0.
4	14.27	0.241	4.01	13.00	65.072	10.05	0.
5	16.54	0.287	4.64	15.06	92.014	14.21	0.
6	22.66	0.440	3.93	12.73	134.343	20.74	0.
7	26.35	0.402	6.35	20.59	167.921	25.93	0.
7	30.00	0.297	30.84	100.00	647.705	100.00	0.

This report has been created by IC Net
 METROHM LTD

Report date: 12/15/2020 4:27:31 PM
Printed by: wet


Ident: CCB
Analysis from: 12/14/2020 7:48:38 PM
File: _2020-12-14_

Last save: 12/14/2020 8:18:26 PM

Method: ANIONS 12-02-20.mtw
Run operator: wet
Analysis number: 72678

Last save: 12/14/2020 4:23:50 PM

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD