

LB112404

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AM

Instrument ID : Konelab

12/16/2020 13:57

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.195	0.0	0.087	
ICB1	0.618	0.0	0.001	
CCV1	242.442	0.0	0.220	
CCB1	0.821	0.0	0.001	
PB133621BL	0.588	0.0	0.001	
PB133621BS	96.348	0.0	0.088	
MIDPB133621	241.041	0.0	0.219	
L5028-07	5.328	0.0	0.005	
L5028-07DUP	5.341	0.0	0.005	
L5051-07	5.816	0.0	0.005	
CCV2	252.195	0.0	0.229	
CCB2	0.618	0.0	0.001	

N	12
Mean	78.863
SD	106.2959
CV%	134.79

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AM

Instrument ID : Konelab

12/16/2020 11:35


Test Total CN

Accepted 12/16/2020 9:56

Factor 1101
 Bias 0

Coeff. of det. 0.999998

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.000	-0.0682	0.0000	
2	5.0PPBCN	0.005	5.0205	5.0000	
3	10PPBCN	0.009	10.0304	10.0000	
4	50PPBCN	0.045	49.5256	50.0000	
5	100PPBCN	0.091	100.3394	100.0000	
6	250PPBCN	0.228	250.3472	250.0000	
7	500PPBCN	0.454	499.8052	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Wed Dec 16 13:47:54 2020

Wed Dec 16 13:53:43 2020

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date	Stat
0.OPPBCN	A	Amendable CN	P	-0.0682	µg/l	44181.41	
5.OPPBCN	A	Amendable CN	P	5.0205	µg/l	44181.41	
10PPBCN	A	Amendable CN	P	10.0304	µg/l	44181.41	
50PPBCN	A	Amendable CN	P	49.5256	µg/l	44181.41	
100PPBCN	A	Amendable CN	P	100.3394	µg/l	44181.41	
250PPBCN	A	Amendable CN	P	250.3472	µg/l	44181.41	
500PPBCN	A	Amendable CN	P	499.8052	µg/l	44181.41	
ICV1	S	Amendable CN	P	95.1951	µg/l	44181.57	
ICB1	S	Amendable CN	P	0.618	µg/l	44181.57	
CCV1	S	Amendable CN	P	242.4424	µg/l	44181.57	
CCB1	S	Amendable CN	P	0.8212	µg/l	44181.58	
PB133621BL	S	Amendable CN	P	0.5885	µg/l	44181.58	
PB133621BS	S	Amendable CN	P	96.3476	µg/l	44181.58	
MIDPB133621	S	Amendable CN	P	241.0411	µg/l	44181.58	
L5028-07	S	Amendable CN	P	5.328	µg/l	44181.58	
L5028-07DUP	S	Amendable CN	P	5.341	µg/l	44181.58	
L5051-07	S	Amendable CN	P	5.8165	µg/l	44181.58	
CCV2	S	Amendable CN	P	252.1948	µg/l	44181.58	
CCB2	S	Amendable CN	P	0.6176	µg/l	44181.58	