

LB112882


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55304



A= 0.0000e+000

B= 1.8049e-004

C= -1.3423e-002

Rho= 0.9999563

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.005	-0.005	47	0.000	47					-
0.2	0.200	0.198	-0.002	1169	0.0 %	1169					-
2.5	2.500	2.463	-0.037	13718	0.0 %	13718					-
5.0	5.000	5.068	0.068	28154	0.0 %	28154					-
7.5	7.500	7.508	0.008	41674	0.0 %	41674					-
10.0	10.000	9.968	-0.032	55304	0.0 %	55304					0

LB112882 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Jan 2021 10:28:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.0000	7471B	PPB	22 Jan 2021 11:01:46	S	Std
0.2 - 1	50.2	1169	-	0.2000	7471B	PPB	22 Jan 2021 11:04:03	S	Std
2.5 - 1	52.5	13718	-	2.5000	7471B	PPB	22 Jan 2021 11:06:21	S	Std
5.0 - 1	55	28154	-	5.0000	7471B	PPB	22 Jan 2021 11:08:38	S	Std
7.5 - 1	57.5	41674	-	7.5000	7471B	PPB	22 Jan 2021 11:10:54	S	Std
10.0 - 1	510	55304	-	10.0000	7471B	PPB	22 Jan 2021 11:15:42	S	Std
ICV79 - 1	ICV79	22203	3.9940	-	7471B	PPB	22 Jan 2021 11:18:27	U	SMPL
ICB79 - 1	ICB79	-54	-0.0232	-	7471B	PPB	22 Jan 2021 11:20:42	U	SMPL
CCV61 - 1	CCV61	28426	5.1172	-	7471B	PPB	22 Jan 2021 11:22:57	U	SMPL
CCB61 - 1	CCB61	-41	-0.0208	-	7471B	PPB	22 Jan 2021 11:25:12	U	SMPL
CRA - 1	CRA	1157	0.1954	-	7471B	PPB	22 Jan 2021 11:27:27	U	SMPL
HighStd - 1	HighStd	56491	10.1826	-	7471B	PPB	22 Jan 2021 11:33:39	U	SMPL
ChkStd - 1	ChkStd	40133	7.2302	-	7471B	PPB	22 Jan 2021 11:35:54	U	SMPL
PB134260BL - 1	PBS	-67	-0.0255	-	7471B	PPB	22 Jan 2021 11:38:11	U	SMPL
PB134260BS - 1	LCSS	21372	3.8440	-	7471B	PPB	22 Jan 2021 11:40:29	U	SMPL
M1179-02 - 1	HS-TP1	687	0.1106	-	7471B	PPB	22 Jan 2021 11:42:45	U	SMPL
M1179-02DUP - 1	HS-TP1DUP	442	0.0664	-	7471B	PPB	22 Jan 2021 11:45:01	U	SMPL
M1179-02MS - 1	HS-TP1MS	20282	3.6473	-	7471B	PPB	22 Jan 2021 11:47:18	U	SMPL
M1179-02MSD - 1	HS-TP1MSD	20094	3.6133	-	7471B	PPB	22 Jan 2021 11:49:35	U	SMPL
M1179-04 - 1	HS-TP2	900	0.1490	-	7471B	PPB	22 Jan 2021 11:51:50	U	SMPL
CCV62 - 1	CCV62	28630	5.1540	-	7471B	PPB	22 Jan 2021 11:54:06	U	SMPL
CCB62 - 1	CCB62	-33	-0.0194	-	7471B	PPB	22 Jan 2021 11:56:21	U	SMPL
M1179-06 - 1	HS-TP3	805	0.1319	-	7471B	PPB	22 Jan 2021 11:58:36	U	SMPL
M1179-08 - 1	HS-TP4	375	0.0543	-	7471B	PPB	22 Jan 2021 12:00:52	U	SMPL
M1188-02 - 1	TP-230	104261	18.8047	-	7471B	PPB	22 Jan 2021 12:03:07	U	SMPL
M1188-05 - 1	TP-SWITCH-GEAR	15594	2.8011	-	7471B	PPB	22 Jan 2021 12:05:23	U	SMPL
M1191-01 - 1	0040-042-COMP-B	4299	0.7625	-	7471B	PPB	22 Jan 2021 12:07:41	U	SMPL
M1191-02 - 1	0041-043-COMP-C	28514	5.1331	-	7471B	PPB	22 Jan 2021 12:09:58	U	SMPL
M1191-03 - 1	0042-044-COMP-D	29030	5.2262	-	7471B	PPB	22 Jan 2021 12:12:14	U	SMPL
M1191-04 - 1	0043-045-COMP-E	23023	4.1420	-	7471B	PPB	22 Jan 2021 12:14:32	U	SMPL
M1191-05 - 1	0047-052-COMP-I	3709	0.6560	-	7471B	PPB	22 Jan 2021 12:16:49	U	SMPL
M1193-01 - 1	OR-02-012121	1936	0.3360	-	7471B	PPB	22 Jan 2021 12:19:05	U	SMPL
CCV63 - 1	CCV63	28830	5.1901	-	7471B	PPB	22 Jan 2021 12:21:21	U	SMPL
CCB63 - 1	CCB63	-37	-0.0201	-	7471B	PPB	22 Jan 2021 12:23:36	U	SMPL
M1179-02LX5 - 1		142	0.0122	-	7471B	PPB	22 Jan 2021 12:25:52	U	SMPL
M1179-02A - 1		22158	3.9859	-	7471B	PPB	22 Jan 2021 12:28:08	U	SMPL
M1188-02DLX5 - 1		20930	3.7642	-	7471B	PPB	22 Jan 2021 12:30:23	U	SMPL
CCV64 - 1	CCV64	28715	5.1694	-	7471B	PPB	22 Jan 2021 12:32:39	U	SMPL
CCB64 - 1	CCB64	-47	-0.0219	-	7471B	PPB	22 Jan 2021 12:34:55	U	SMPL