

LB118552

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AM Instrument ID : Konelab

3/17/2021 16:30

Test: Phenolics-

Sample Id	Result ^{mg/L}	Dil. 1 +	Response	Errors
ICV1	0.935	0.0	0.102	
ICB1	-0.007	0.0	0.005	
CCV1	0.939	0.0	0.102	
CCB1	-0.005	0.0	0.005	
PB135100BL	-0.004	0.0	0.006	
PB135100BS	0.941	0.0	0.102	
M1458-03	0.019	0.0	0.008	
PB135102BL	-0.004	0.0	0.006	
PB135102BS	0.953	0.0	0.104	
M1458-09	0.018	0.0	0.008	
CCV2	0.962	0.0	0.104	
CCB2	-0.005	0.0	0.006	

N	12
Mean	0.395
SD	0.4865
CV%	123.13

Aquakem v. 7.2AQ1

Results from time period:

Wed Mar 17 16:09:08 2021

Wed Mar 17 16:19:21 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Phenolics-	P	-0.003	mg/l	3/17/2021 14:13:02	
0.05PPM	A	Phenolics-	P	0.0511	mg/l	3/17/2021 14:13:03	
0.1PPM	A	Phenolics-	P	0.1046	mg/l	3/17/2021 14:13:04	
0.25PPM	A	Phenolics-	P	0.2453	mg/l	3/17/2021 14:13:05	
0.50PPM	A	Phenolics-	P	0.5034	mg/l	3/17/2021 14:13:06	
1.0PPM	A	Phenolics-	P	0.9979	mg/l	3/17/2021 14:13:07	
2.0PPM	A	Phenolics-	P	2.0005	mg/l	3/17/2021 14:13:08	
ICV1	S	Phenolics-	P	0.9349	mg/l	3/17/2021 16:09:08	
ICB1	S	Phenolics-	P	-0.0074	mg/l	3/17/2021 16:09:10	
CCV1	S	Phenolics-	P	0.9389	mg/l	3/17/2021 16:09:13	
CCB1	S	Phenolics-	P	-0.0054	mg/l	3/17/2021 16:09:14	
PB135100BL	S	Phenolics-	P	-0.0042	mg/l	3/17/2021 16:09:16	
PB135100BS	S	Phenolics-	P	0.9411	mg/l	3/17/2021 16:09:18	
M1458-03	S	Phenolics-	P	0.0187	mg/l	3/17/2021 16:19:12	
PB135102BL	S	Phenolics-	P	-0.0037	mg/l	3/17/2021 16:19:13	
PB135102BS	S	Phenolics-	P	0.9531	mg/l	3/17/2021 16:19:14	
M1458-09	S	Phenolics-	P	0.0182	mg/l	3/17/2021 16:19:15	
CCV2	S	Phenolics-	P	0.9624	mg/l	3/17/2021 16:19:18	
CCB2	S	Phenolics-	P	-0.0054	mg/l	3/17/2021 16:19:20	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

3/17/2021 14:25

Reviewed by : AM Instrument ID : Konelab


Test Phenolics-

Accepted 3/17/2021 14:25

Factor 9.78
Bias 0.006

Coeff. of det. 0.999978

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.006	-0.0030	0.0000	
2	Phenol-2	0.011	0.0511	0.0500	
3	Phenol-2	0.017	0.1046	0.1000	
4	Phenol-2	0.031	0.2453	0.2500	
5	Phenol-2	0.058	0.5034	0.5000	
6	Phenol-2	0.108	0.9979	1.0000	
7	Phenol-2	0.211	2.0005	2.0000	

1% DI

-252100

03/17/2021
AM