

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
CCV	10.6830	0.2024	1.89	TOC	
CCB	0.2542	0.0423	16.66	TOC	
LB113881BLW	0.1978	0.0322	16.25	TOC	
LB113881BSW.....	10.5221...	0.0797..	0.76...	TOC	..
M1967-13	1.8192	0.1593	8.76	TOC	
CCV	10.7777	0.0894	0.83	TOC	
CCB.....	0.2368...	0.0322..	13.59...	TOC	..

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	Sample	2	2021/04/08 13:56	
TOC 0 - 20 ppmC	Sample	3	2021/04/08 14:18	
TOC 0 - 20 ppmC	Sample	4	2021/04/08 14:40	
TOC 0 - 20 ppmC	...Sample	..	5..2021/04/08 15:04	..
TOC 0 - 20 ppmC	Sample	7	2021/04/08 17:34	
TOC 0 - 20 ppmC	Sample	2	2021/04/08 17:58	
TOC 0 - 20 ppmC	...Sample	..	3..2021/04/08 18:20	..

```
=====
Sample ID:   CCV                               Mode:      TOC
Method:      TOC 0 - 20 ppmC                   Filename:   04081308
Cal. Curve:  TOC WATER 9060A                   Timestamp: 2021/04/08 13:56
Operator ID: AM AP                             Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	10.8849	5.4425	469180	-0.642	-0.442	135
2	10.7789	5.3895	464612	-0.596	-0.397	130
3	10.6534	5.3267	459203	-0.552	-0.353	128
4	10.4146	5.2073	448907	-0.584	-0.386	120

```
-----
<<<Statistics>>>   Mean:   10.6830   Std Dev:   0.2024   RSD: 1.89
=====
```

```
=====
Sample ID:   CCB                               Mode:      TOC
Method:      TOC 0 - 20 ppmC                   Filename:   04081308
Cal. Curve:  TOC WATER 9060A                   Timestamp: 2021/04/08 14:18
Operator ID: AM AP                             Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.2180	0.1090	9397	-0.796	-0.598	91
2	0.2706	0.1353	11666	-0.857	-0.660	92
3	0.2216	0.1108	9552	-0.868	-0.670	91
4	0.3065	0.1533	13213	-0.882	-0.683	92

```
-----
<<<Statistics>>>   Mean:    0.2542   Std Dev:    0.0423   RSD: 16.66
=====
```

```
=====
Sample ID:   LB113881BLW                     Mode:      TOC
Method:      TOC 0 - 20 ppmC                   Filename:   04081308
Cal. Curve:  TOC WATER 9060A                   Timestamp: 2021/04/08 14:40
Operator ID: AM AP                             Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.1579	0.0789	6804	-0.916	-0.717	89
2	0.1867	0.0933	8046	-0.942	-0.742	90
3	0.2170	0.1085	9355	-0.947	-0.752	90
4	0.2297	0.1148	9900	-0.961	-0.764	91

```
-----
<<<Statistics>>>   Mean:    0.1978   Std Dev:    0.0322   RSD: 16.25
=====
```

```
=====
Sample ID:   LB113881BSW                     Mode:      TOC
Method:      TOC 0 - 20 ppmC                   Filename:   04081308
Cal. Curve:  TOC WATER 9060A                   Timestamp: 2021/04/08 15:04
Operator ID: AM AP                             Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	10.5123	5.2562	453121	-0.870	-0.671	135
2	10.4657	5.2329	451111	-0.828	-0.631	131
3	10.4729	5.2365	451422	-0.806	-0.608	142
4	10.6376	5.3188	458519	-0.801	-0.604	129

```
-----
<<<Statistics>>>   Mean:   10.5221   Std Dev:    0.0797   RSD: 0.76
=====
```

```
=====
Sample ID:   M1967-13                         Mode:      TOC
Method:      TOC 0 - 20 ppmC                   Filename:   04081308
Cal. Curve:  TOC WATER 9060A                   Timestamp: 2021/04/08 17:34
Operator ID: AM AP                             Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning	Ending	Integration
-------	-------	------	----------	-----------	--------	-------------

				Baseline	Baseline	Time
1	1.9260	0.9630	83020	-0.541	-0.343	186
2	1.9084	0.9542	82260	-0.467	-0.268	171
3	1.8584	0.9292	80104	-0.407	-0.208	165
4	1.5841	0.7920	68279	-0.359	-0.161	154

<<<Statistics>>> Mean: 1.8192 Std Dev: 0.1593 RSD: 8.76

Sample ID:	CCV	Mode:	TOC
Method:	TOC 0 - 20 ppmC	Filename:	04081308
Cal. Curve:	TOC WATER 9060A	Timestamp:	2021/04/08 17:58
Operator ID:	AM AP	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	10.8670	5.4335	468410	-0.597	-0.400	141
2	10.8013	5.4006	465575	-0.541	-0.341	135
3	10.6539	5.3270	459223	-0.509	-0.312	124
4	10.7885	5.3943	465027	-0.513	-0.317	132

<<<Statistics>>> Mean: 10.7777 Std Dev: 0.0894 RSD: 0.83

Sample ID:	CCB	Mode:	TOC
Method:	TOC 0 - 20 ppmC	Filename:	04081308
Cal. Curve:	TOC WATER 9060A	Timestamp:	2021/04/08 18:20
Operator ID:	AM AP	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.1924	0.0962	8292	-0.671	-0.473	91
2	0.2468	0.1234	10638	-0.695	-0.500	93
3	0.2393	0.1196	10313	-0.683	-0.484	91
4	0.2689	0.1344	11590	-0.705	-0.505	92

<<<Statistics>>> Mean: 0.2368 Std Dev: 0.0322 RSD: 13.59

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
0.0PPM	10935	2079	19.01	TOC	
0.5PPM	30350	1328	4.38	TOC	
1.0PPM	50065	2517	5.03	TOC	
2.0PPM.....	88711...	2814..	3.17...	TOC	..
5.0PPM	202799	1586	0.78	TOC	
10.0PPM	442148	5317	1.20	TOC	
20.0PPM.....	868135...	4860..	0.56...	TOC	..
ICV	10.4223	0.2300	2.21	TOC	
ICB	0.2492	0.0428	17.18	TOC	
IC-20PPM.....	0.2729...	0.0352..	12.91...	TOC	..

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	TOC Standard	1	2021/03/01 14:46	
TOC 0 - 20 ppmC	TOC Standard	2	2021/03/01 15:09	
TOC 0 - 20 ppmC	TOC Standard	3	2021/03/01 15:32	
TOC 0 - 20 ppmC	...TOC Standard	.. 4..	2021/03/01 15:55	..
TOC 0 - 20 ppmC	TOC Standard	5	2021/03/01 16:19	
TOC 0 - 20 ppmC	TOC Standard	6	2021/03/01 16:44	
TOC 0 - 20 ppmC	...TOC Standard	.. 7..	2021/03/01 17:11	..
TOC 0 - 20 ppmC	Sample	8	2021/03/01 17:36	
TOC 0 - 20 ppmC	Sample	9	2021/03/01 17:58	
TOC 0 - 20 ppmC	...Sample	.. 10..	2021/03/01 18:20	..

```
=====
Sample ID: 0.0PPM Mode: TOC
Method: TOC 0 - 20 ppmC Filename: 03011425
Cal. Curve: TOC WATER 9060A Timestamp: 2021/03/01 14:46
Operator ID: AM AP Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			10294	-0.859	-0.664	99
2			9525	-0.848	-0.651	99
3			9903	-0.852	-0.654	98
4			14018	-0.871	-0.672	101

```
-----
<<<Statistics>>> Mean: 10935 Std Dev: 2079 RSD: 19.01
=====
```

```
=====
Sample ID: 0.5PPM Mode: TOC
Method: TOC 0 - 20 ppmC Filename: 03011425
Cal. Curve: TOC WATER 9060A Timestamp: 2021/03/01 15:09
Operator ID: AM AP Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			30892	-0.832	-0.637	106
2			28439	-0.813	-0.615	106
3			31486	-0.832	-0.632	106
4			30584	-0.819	-0.621	107

```
-----
<<<Statistics>>> Mean: 30350 Std Dev: 1328 RSD: 4.38
=====
```

```
=====
Sample ID: 1.0PPM Mode: TOC
Method: TOC 0 - 20 ppmC Filename: 03011425
Cal. Curve: TOC WATER 9060A Timestamp: 2021/03/01 15:32
Operator ID: AM AP Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			48722	-0.789	-0.593	111
2			49536	-0.766	-0.568	111
3			53756	-0.803	-0.605	112
4			48247	-0.776	-0.576	109

```
-----
<<<Statistics>>> Mean: 50065 Std Dev: 2517 RSD: 5.03
=====
```

```
=====
Sample ID: 2.0PPM Mode: TOC
Method: TOC 0 - 20 ppmC Filename: 03011425
Cal. Curve: TOC WATER 9060A Timestamp: 2021/03/01 15:55
Operator ID: AM AP Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			87722	-0.813	-0.613	114
2			85972	-0.798	-0.599	114
3			88535	-0.801	-0.602	115
4			92615	-0.814	-0.617	115

```
-----
<<<Statistics>>> Mean: 88711 Std Dev: 2814 RSD: 3.17
=====
```

```
=====
Sample ID: 5.0PPM Mode: TOC
Method: TOC 0 - 20 ppmC Filename: 03011425
Cal. Curve: TOC WATER 9060A Timestamp: 2021/03/01 16:19
Operator ID: AM AP Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning	Ending	Integration
-------	-------	------	----------	-----------	--------	-------------

			Baseline	Baseline	Time
1		204787	-0.783	-0.585	125
2		203365	-0.732	-0.534	124
3		201606	-0.708	-0.508	124
4		201438	-0.712	-0.514	121
<<<Statistics>>> Mean: 202799 Std Dev: 1586 RSD: 0.78					
=====					
Sample ID: 10.0PPM			Mode: TOC		
Method: TOC 0 - 20 ppmC			Filename: 03011425		
Cal. Curve: TOC WATER 9060A			Timestamp: 2021/03/01 16:44		
Operator ID: AM AP			Sample Type: TOC Standard		
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline Integration Time
1			444583	-0.834	-0.635 140
2			444269	-0.741	-0.544 156
3			434212	-0.749	-0.549 131
4			445526	-0.764	-0.567 133
<<<Statistics>>> Mean: 442148 Std Dev: 5317 RSD: 1.20					
=====					
Sample ID: 20.0PPM			Mode: TOC		
Method: TOC 0 - 20 ppmC			Filename: 03011425		
Cal. Curve: TOC WATER 9060A			Timestamp: 2021/03/01 17:11		
Operator ID: AM AP			Sample Type: TOC Standard		
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline Integration Time
1			871993	-0.802	-0.604 171
2			867580	-0.709	-0.511 171
3			861473	-0.666	-0.466 160
4			871494	-0.636	-0.440 160
<<<Statistics>>> Mean: 868135 Std Dev: 4860 RSD: 0.56					
=====					
Sample ID: ICV			Mode: TOC		
Method: TOC 0 - 20 ppmC			Filename: 03011425		
Cal. Curve: TOC WATER 9060A			Timestamp: 2021/03/01 17:36		
Operator ID: AM AP			Sample Type: Sample		
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline Integration Time
1	10.6416	5.3208	458692	-0.807	-0.608 148
2	10.1032	5.0516	435485	-0.730	-0.531 140
3	10.5158	5.2579	453270	-0.672	-0.472 133
4	10.4286	5.2143	449513	-0.723	-0.523 145
<<<Statistics>>> Mean: 10.4223 Std Dev: 0.2300 RSD: 2.21					
=====					
Sample ID: ICB			Mode: TOC		
Method: TOC 0 - 20 ppmC			Filename: 03011425		
Cal. Curve: TOC WATER 9060A			Timestamp: 2021/03/01 17:58		
Operator ID: AM AP			Sample Type: Sample		
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline Integration Time
1	0.2127	0.1064	9168	-0.833	-0.637 99
2	0.2987	0.1493	12875	-0.846	-0.647 100
3	0.2141	0.1071	9229	-0.812	-0.614 98
4	0.2711	0.1355	11684	-0.856	-0.657 100
<<<Statistics>>> Mean: 0.2492 Std Dev: 0.0428 RSD: 17.18					
=====					

Sample ID: IC-20PPM
Method: TOC 0 - 20 ppmC
Cal. Curve: TOC WATER 9060A
Operator ID: AM AP

Mode: TOC
Filename: 03011425
Timestamp: 2021/03/01 18:20
Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.2865	0.1433	12349	-0.902	-0.703	100
2	0.2592	0.1296	11171	-0.923	-0.727	100
3	0.2321	0.1161	10006	-0.935	-0.736	99
4	0.3140	0.1570	13534	-0.931	-0.731	101

<<<Statistics>>> Mean: 0.2729 Std Dev: 0.0352 RSD: 12.91

Cal. Curve ID: TOC WATER 9060A
 Created: 2021/03/01 17:11
 Calibration Factor (m): 8.621e+04
 Y Intercept (b): 4807
 r-squared: 0.99930

Standard ID	Y Raw Data	X Expected ug C	Measured ug C	Message	Date & Time
0.0PPM	10935	0.000	0.071	18	2021/03/01 14:46
0.5PPM	30350	0.250	0.296	5	2021/03/01 15:09
1.0PPM	50065	0.500	0.525	3	2021/03/01 15:32
2.0PPM	88711	1.000	0.973	8	2021/03/01 15:55
5.0PPM	202799	2.500	2.297	i	2021/03/01 16:19
10.0PPM	442148	5.000	5.073	0	2021/03/01 16:44
20.0PPM	868135	10.000	10.015		2021/03/01 17:11

03/02/2021
 AM