

LB 114246


7470A

INSTRUMENT ID: CVL

Linear

 μ Abs.:

35987



A= 0.0000e+000

B= 2.7605e-004

C= 1.9186e-002

Rho= 0.9998454

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.0	0.000	0.026	0.026	25	0.000	25					5
0.2	0.200	0.210	0.010	691	0.0 %	691					-1
2.5	2.500	2.468	-0.032	8870	0.0 %	8870					-2
5.0	5.000	4.921	-0.079	17756	0.0 %	17756					2
7.5	7.500	7.622	0.122	27543	0.0 %	27543					2
10.0	10.000	9.953	-0.047	35987	0.0 %	35987					0

LB114246 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 May 2021 13:01:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.00007470A	PPB		03 May 2021 13:31:21	S	Std
0.2 - 1	50.2	691	-	0.20007470A	PPB		03 May 2021 13:33:38	S	Std
2.5 - 1	52.5	8870	-	2.50007470A	PPB		03 May 2021 13:35:55	S	Std
5.0 - 1	55	17756	-	5.00007470A	PPB		03 May 2021 13:38:12	S	Std
7.5 - 1	57.5	27543	-	7.50007470A	PPB		03 May 2021 13:40:28	S	Std
10.0 - 1	59.5	35987	-	10.00007470A	PPB		03 May 2021 13:42:44	S	Std
ICV13 - 1	ICV13	14671	4.0691	-7470A	PPB		03 May 2021 13:45:30	U	SMPL
ICB13 - 1	ICB13	-208	-0.0382	-7470A	PPB		03 May 2021 13:47:44	U	SMPL
CCV30 - 1	CCV30	18416	5.1029	-7470A	PPB		03 May 2021 13:49:59	U	SMPL
CCB30 - 1	CCB30	-204	-0.0371	-7470A	PPB		03 May 2021 13:52:14	U	SMPL
CRA - 1	CRA	653	0.1994	-7470A	PPB		03 May 2021 13:54:30	U	SMPL
HighStd - 1	HighStd	36498	10.0943	-7470A	PPB		03 May 2021 13:56:45	U	SMPL
ChkStd - 1	ChkStd	25926	7.1760	-7470A	PPB		03 May 2021 13:59:00	U	SMPL
PB136101BL - 1	PBW	-308	-0.0658	-7470A	PPB		03 May 2021 14:01:17	U	SMPL
PB136101BS - 1	LCSW	13124	3.6420	-7470A	PPB		03 May 2021 14:03:35	U	SMPL
M2229-02 - 1	TP-6	-172	-0.0283	-7470A	PPB		03 May 2021 14:05:51	U	SMPL
M2230-02 - 1	JTY3-WC-S-1	-12	0.0159	-7470A	PPB		03 May 2021 14:08:09	U	SMPL
M2230-02DUP - 1	JTY3-WC-S-1DUP	-10	0.0164	-7470A	PPB		03 May 2021 14:10:25	U	SMPL
M2230-02MS - 1	JTY3-WC-S-1MS	13920	3.8618	-7470A	PPB		03 May 2021 14:12:43	U	SMPL
M2230-02MSD - 1	JTY3-WC-S-1MSD	12344	3.4267	-7470A	PPB		03 May 2021 14:14:57	U	SMPL
CCV31 - 1	CCV31	17972	4.9803	-7470A	PPB		03 May 2021 14:17:13	U	SMPL
CCB31 - 1	CCB31	-207	-0.0380	-7470A	PPB		03 May 2021 14:19:29	U	SMPL
PB136102BL - 1	PBW	-41	0.0079	-7470A	PPB		03 May 2021 14:21:46	U	SMPL
PB136102BS - 1	LCSW	12149	3.3729	-7470A	PPB		03 May 2021 14:24:01	U	SMPL
M2204-05 - 1	ANTIFREEZE-1906	181	0.0692	-7470A	PPB		03 May 2021 14:26:17	U	SMPL
M2204-05DUP - 1	ANTIFREEZE-1906DUP	240	0.0854	-7470A	PPB		03 May 2021 14:31:16	U	SMPL
M2204-05MS - 1	ANTIFREEZE-1906MS	12295	3.4132	-7470A	PPB		03 May 2021 14:36:07	U	SMPL
M2204-05MSD - 1	ANTIFREEZE-1906MSD	12511	3.4728	-7470A	PPB		03 May 2021 14:38:23	U	SMPL
PB136103BL - 1	PBW	-268	-0.0548	-7470A	PPB		03 May 2021 14:40:41	U	SMPL
PB136103BS - 1	LCSW	13227	3.6705	-7470A	PPB		03 May 2021 14:42:57	U	SMPL
M2202-08 - 1	TOTE-COMP	-149	-0.0219	-7470A	PPB		03 May 2021 14:45:15	U	SMPL
M2202-08DUP - 1	TOTE-COMP DUP	-38	0.0087	-7470A	PPB		03 May 2021 14:47:30	U	SMPL
CCV32 - 1	CCV32	18039	4.9988	-7470A	PPB		03 May 2021 14:49:45	U	SMPL
CCB32 - 1	CCB32	-182	-0.0311	-7470A	PPB		03 May 2021 14:52:01	U	SMPL
M2202-08MS - 1	TOTE-COMPMS	13430	3.7265	-7470A	PPB		03 May 2021 14:57:30	U	SMPL
M2202-08MSD - 1	TOTE-COMPMSD	13203	3.6638	-7470A	PPB		03 May 2021 14:59:45	U	SMPL
PB136101TB - 1	PB136079TB	-165	-0.0264	-7470A	PPB		03 May 2021 15:02:01	U	SMPL
PB136102TB - 1	PB136081TB	-38	0.0087	-7470A	PPB		03 May 2021 15:04:16	U	SMPL
M2230-02LX5 - 1		3	0.0200	-7470A	PPB		03 May 2021 15:06:32	U	SMPL
M2230-02A - 1		14680	4.0715	-7470A	PPB		03 May 2021 15:08:48	U	SMPL
M2204-05LX5 - 1		-60	0.0026	-7470A	PPB		03 May 2021 15:11:05	U	SMPL
M2204-05A - 1		14785	4.1005	-7470A	PPB		03 May 2021 15:13:22	U	SMPL
M2202-08LX5 - 1		-128	-0.0161	-7470A	PPB		03 May 2021 15:15:39	U	SMPL
M2202-08A - 1		14773	4.0972	-7470A	PPB		03 May 2021 15:17:54	U	SMPL
CCV33 - 1	CCV33	18254	5.0581	-7470A	PPB		03 May 2021 15:20:09	U	SMPL
CCB33 - 1	CCB33	-166	-0.0266	-7470A	PPB		03 May 2021 15:22:25	U	SMPL