

LB114518

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AM Instrument ID : Konelab

5/18/2021 13:47

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.982	0.0	0.102	
ICB1	-0.002	0.0	0.005	
CCV1	0.990	0.0	0.103	
CCB1	0.000	0.0	0.005	
PB136507BL	-0.000	0.0	0.005	
PB136507BS	0.988	0.0	0.103	
M2262-03	0.026	0.0	0.007	
PB136509BL	-0.001	0.0	0.005	
PB136509BS	0.986	0.0	0.103	
M2262-09	0.023	0.0	0.007	
CCV2	0.991	0.0	0.103	
CCB2	-0.001	0.0	0.005	

N	12
Mean	0.415
SD	0.5053
CV%	121.72

Aquakem v. 7.2AQ1

Results from time period:

Tue May 18 13:34:53 2021

Tue May 18 13:45:06 2021

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Phenolics-	P	0.0027	mg/l	5/18/2021 12:12:09	
0.05PPM	A	Phenolics-	P	0.0521	mg/l	5/18/2021 12:12:10	
0.1PPM	A	Phenolics-	P	0.1009	mg/l	5/18/2021 12:12:11	
0.25PPM	A	Phenolics-	P	0.2511	mg/l	5/18/2021 12:12:12	
0.50PPM	A	Phenolics-	P	0.5045	mg/l	5/18/2021 12:12:13	
1.0PPM	A	Phenolics-	P	0.98	mg/l	5/18/2021 12:12:14	
2.0PPM	A	Phenolics-	P	2.0086	mg/l	5/18/2021 12:12:15	
ICV1	S	Phenolics-	P	0.982	mg/l	5/18/2021 13:34:54	
ICB1	S	Phenolics-	P	-0.0024	mg/l	5/18/2021 13:34:56	
CCV1	S	Phenolics-	P	0.9896	mg/l	5/18/2021 13:34:58	
CCB1	S	Phenolics-	P	0.0004	mg/l	5/18/2021 13:35:00	
PB136507BL	S	Phenolics-	P	-0.0001	mg/l	5/18/2021 13:35:02	
PB136507BS	S	Phenolics-	P	0.9884	mg/l	5/18/2021 13:35:03	
M2262-03	S	Phenolics-	P	0.0255	mg/l	5/18/2021 13:44:56	
PB136509BL	S	Phenolics-	P	-0.0012	mg/l	5/18/2021 13:44:58	
PB136509BS	S	Phenolics-	P	0.9863	mg/l	5/18/2021 13:44:59	
M2262-09	S	Phenolics-	P	0.023	mg/l	5/18/2021 13:45:01	
CCV2	S	Phenolics-	P	0.9912	mg/l	5/18/2021 13:45:03	
CCB2	S	Phenolics-	P	-0.0009	mg/l	5/18/2021 13:45:05	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AM Instrument ID : Konelab

5/18/2021 12:12


Test Phenolics-

Accepted 5/18/2021 12:12

Factor 10.06
Bias 0.005

Coeff. of det. 0.999839

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors %RE
1	0.00PPM	0.005	0.0027	0.0000	4.2
2	Phenol-2	0.010	0.0521	0.0500	0.9
3	Phenol-2	0.015	0.1009	0.1000	0.4
4	Phenol-2	0.030	0.2511	0.2500	0.9
5	Phenol-2	0.055	0.5045	0.5000	0.4
6	Phenol-2	0.102	0.9800	1.0000	
7	Phenol-2	0.204	2.0086	2.0000	

05/18/2021
AM