


LB115428

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.9773e-004

C= -1.3321e-002

Rho= 0.9986164

Accept=Accepted

ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
	0.000	0.004	0.004	57	0.000	57					0.00
	0.200	0.205	0.005	732	0.0 %	732					0.7
	2.500	2.346	-0.154	7923	0.0 %	7923					1.9
	5.000	5.380	0.380	18116	0.0 %	18116					2.1
	7.500	7.260	-0.240	24431	0.0 %	24431					0.9
	10.000	10.005	0.005	33651	0.0 %	33651					0.0

LB115428 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Jul 2021 11:23:22

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.00007470A	PPB		20 Jul 2021 11:34:42	S	Std
0.2 - 1	50.2	732	-	0.20007470A	PPB		20 Jul 2021 11:36:59	S	Std
2.5 - 1	50.5	7923	-	2.50007470A	PPB		20 Jul 2021 11:39:15	S	Std
5.0 - 1	55	18116	-	5.00007470A	PPB		20 Jul 2021 11:41:32	S	Std
7.5 - 1	57.5	24431	-	7.50007470A	PPB		20 Jul 2021 11:43:49	S	Std
10.0 - 1	57.0	33651	-	10.00007470A	PPB		20 Jul 2021 11:50:30	S	Std
ICV12 - 1	ICV12	14293	4.2421	-7470A	PPB		20 Jul 2021 11:53:26	U	SMPL
ICB12 - 1	ICB12	-111	-0.0464	-7470A	PPB		20 Jul 2021 11:55:41	U	SMPL
CCV90 - 1	CCV90	17451	5.1823	-7470A	PPB		20 Jul 2021 12:00:28	U	SMPL
CCB90 - 1	CCB90	-124	-0.0502	-7470A	PPB		20 Jul 2021 12:02:43	U	SMPL
CRA - 1	CRA	702	0.1957	-7470A	PPB		20 Jul 2021 12:07:42	U	SMPL
HighStd - 1	HighStd	33328	9.9093	-7470A	PPB		20 Jul 2021 12:09:56	U	SMPL
ChkStd - 1	ChkStd	24076	7.1547	-7470A	PPB		20 Jul 2021 12:12:11	U	SMPL
PB137791BL - 1	PBW	-1	-0.0136	-7470A	PPB		20 Jul 2021 12:16:59	U	SMPL
PB137791BS - 1	LCSW	12790	3.7946	-7470A	PPB		20 Jul 2021 12:19:15	U	SMPL
M3061-01 - 1	PB-2021-WS-000-08	343	0.0888	-7470A	PPB		20 Jul 2021 12:29:08	U	SMPL
M3088-02 - 1	B1-(A-D)-(0-4-2)	268	0.0665	-7470A	PPB		20 Jul 2021 12:31:24	U	SMPL
M3089-02 - 1	PAR-N2-(A-D)-(0-12)	273	0.0680	-7470A	PPB		20 Jul 2021 12:33:41	U	SMPL
M3089-13 - 1	R99997-(A-D)-(0-3)	139	0.0281	-7470A	PPB		20 Jul 2021 12:42:51	U	SMPL
M3089-13DUP - 1	R99997-(A-D)-(0-3)DUP	127	0.0245	-7470A	PPB		20 Jul 2021 12:45:06	U	SMPL
CCV91 - 1	CCV91	18312	5.4386	-7470A	PPB		20 Jul 2021 12:47:21	U	SMPL
CCB91 - 1	CCB91	-99	-0.0428	-7470A	PPB		20 Jul 2021 12:49:37	U	SMPL
M3089-13MS - 1	R99997-(A-D)-(0-3)MS	15166	4.5020	-7470A	PPB		20 Jul 2021 12:51:52	U	SMPL
M3089-13MSD - 1	R99997-(A-D)-(0-3)MSD	14412	4.2775	-7470A	PPB		20 Jul 2021 12:54:08	U	SMPL
PB137792BL - 1	PBW	-15	-0.0178	-7470A	PPB		20 Jul 2021 12:56:23	U	SMPL
PB137792BS - 1	LCSW	15683	4.6559	-7470A	PPB		20 Jul 2021 12:58:39	U	SMPL
M3076-04 - 1	20210587-AMENDED-COM-1-SPLP	1108	0.3166	-7470A	PPB		20 Jul 2021 13:00:54	U	SMPL
M3076-04DUP - 1	20210587-AMENDED-COM-1-SPLPDUP	1137	0.3252	-7470A	PPB		20 Jul 2021 13:03:10	U	SMPL
M3076-04MS - 1	20210587-AMENDED-COM-1-SPLPMS	16651	4.9441	-7470A	PPB		20 Jul 2021 13:05:26	U	SMPL
M3076-04MSD - 1	20210587-AMENDED-COM-1-SPLPMSD	16190	4.8069	-7470A	PPB		20 Jul 2021 13:07:43	U	SMPL
PB137793BL - 1	PBW	-66	-0.0330	-7470A	PPB		20 Jul 2021 13:09:59	U	SMPL
PB137793BS - 1	LCSW	13837	4.1063	-7470A	PPB		20 Jul 2021 13:12:14	U	SMPL
CCV92 - 1	CCV92	18113	5.3794	-7470A	PPB		20 Jul 2021 13:14:30	U	SMPL
CCB92 - 1	CCB92	-93	-0.0410	-7470A	PPB		20 Jul 2021 13:16:45	U	SMPL
M3051-09 - 1	41213	912	0.2582	-7470A	PPB		20 Jul 2021 13:23:53	U	SMPL
M3051-09DUP - 1	41213DUP	865	0.2442	-7470A	PPB		20 Jul 2021 13:26:08	U	SMPL
M3051-09MS - 1	41213MS	16311	4.8429	-7470A	PPB		20 Jul 2021 13:28:24	U	SMPL
M3051-09MSD - 1	41213MSD	15721	4.6672	-7470A	PPB		20 Jul 2021 13:30:39	U	SMPL
M3051-10 - 1	COMP-41257-59	300	0.0760	-7470A	PPB		20 Jul 2021 13:32:55	U	SMPL
M3051-11 - 1	30817	42	-0.0008	-7470A	PPB		20 Jul 2021 13:35:11	U	SMPL
M3061-02 - 1	PB-2021-WW-000-08	368	0.0962	-7470A	PPB		20 Jul 2021 13:37:28	U	SMPL
M3076-05 - 1	20210515-FIELD-BLANK-1	39	-0.0017	-7470A	PPB		20 Jul 2021 13:39:45	U	SMPL
CCV93 - 1	CCV93	17095	5.0763	-7470A	PPB		20 Jul 2021 13:44:35	U	SMPL
CCB93 - 1	CCB93	-71	-0.0345	-7470A	PPB		20 Jul 2021 13:46:50	U	SMPL
PB137791TB - 1	PB137752TB	249	0.0608	-7470A	PPB		20 Jul 2021 13:51:24	U	SMPL
PB137792TB - 1	PB137760TB	387	0.1019	-7470A	PPB		20 Jul 2021 13:53:40	U	SMPL
M3089-13LX5 - 1		13	-0.0095	-7470A	PPB		20 Jul 2021 13:55:55	U	SMPL
M3089-13A - 1		14424	4.2811	-7470A	PPB		20 Jul 2021 13:58:10	U	SMPL
M3076-04LX5 - 1		185	0.0418	-7470A	PPB		20 Jul 2021 14:00:26	U	SMPL
M3076-04A - 1		15548	4.6157	-7470A	PPB		20 Jul 2021 14:02:41	U	SMPL
M3051-09LX5 - 1		369	0.0965	-7470A	PPB		20 Jul 2021 14:04:57	U	SMPL
M3051-09A - 1		16022	4.7568	-7470A	PPB		20 Jul 2021 14:07:13	U	SMPL
CCV94 - 1	CCV94	17676	5.2493	-7470A	PPB		20 Jul 2021 14:09:30	U	SMPL
CCB94 - 1	CCB94	-44	-0.0264	-7470A	PPB		20 Jul 2021 14:11:46	U	SMPL