


LB115816

7470A

INSTRUMENT ID: CV2



A= 0.0000e+000

B= 2.4472e-004

C= -3.6100e-002

Rho= 0.9999346

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.021	-0.021	61	0.000	61					-10
0.2	0.200	0.180	-0.020	885	0.0 %	885					2
2.5	2.500	2.539	0.039	10522	0.0 %	10522					1
5.0	5.000	5.059	0.059	20820	0.0 %	20820					-1
7.5	7.500	7.431	-0.069	30514	0.0 %	30514					0
10.0	10.000	10.012	0.012	41059	0.0 %	41059					

LB115816 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 12 Aug 2021 15:42:14					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	61	-	0.0000	7470A	PPB	12 Aug 2021 15:48:53	S	Std
0.2 - 1	50-2	885	-	0.2000	7470A	PPB	12 Aug 2021 15:51:11	S	Std
2.5 - 1	52-5	10522	-	2.5000	7470A	PPB	12 Aug 2021 15:53:28	S	Std
5.0 - 1	55	20820	-	5.0000	7470A	PPB	12 Aug 2021 15:55:44	S	Std
7.5 - 1	57-5	30514	-	7.5000	7470A	PPB	12 Aug 2021 15:58:00	S	Std
10.0 - 1	510	41059	-	10.0000	7470A	PPB	12 Aug 2021 16:00:16	S	Std
ICV52 - 1	ICV52	15459	3.7470	-	7470A	PPB	12 Aug 2021 16:03:03	U	SMPL
ICB52 - 1	ICB52	-49	-0.0481	-	7470A	PPB	12 Aug 2021 16:05:17	U	SMPL
CCV46 - 1	CCV46	20778	5.0486	-	7470A	PPB	12 Aug 2021 16:07:32	U	SMPL
CCB46 - 1	CCB46	-70	-0.0532	-	7470A	PPB	12 Aug 2021 16:09:47	U	SMPL
CRA - 1	CRA	823	0.1653	-	7470A	PPB	12 Aug 2021 16:12:02	U	SMPL
HighStd - 1	HighStd	40711	9.9266	-	7470A	PPB	12 Aug 2021 16:14:17	U	SMPL
ChkStd - 1	ChkStd	28365	6.9053	-	7470A	PPB	12 Aug 2021 16:16:32	U	SMPL
PB138365BL - 1	PBW	-95	-0.0593	-	7470A	PPB	12 Aug 2021 16:18:48	U	SMPL
PB138365BS - 1	LCSW	16304	3.9538	-	7470A	PPB	12 Aug 2021 16:21:04	U	SMPL
M3343-10 - 1	SPLP-AMENDED-COM-5	136	-0.0028	-	7470A	PPB	12 Aug 2021 16:23:20	U	SMPL
M3343-10DUP - 1	SPLP-AMENDED-COM-5DUP	196	0.0119	-	7470A	PPB	12 Aug 2021 16:25:36	U	SMPL
M3343-10MS - 1	SPLP-AMENDED-COM-5MS	15504	3.7580	-	7470A	PPB	12 Aug 2021 16:27:53	U	SMPL
M3343-10MSD - 1	SPLP-AMENDED-COM-5MSD	15555	3.7705	-	7470A	PPB	12 Aug 2021 16:30:10	U	SMPL
CCV47 - 1	CCV47	20345	4.9427	-	7470A	PPB	12 Aug 2021 16:32:25	U	SMPL
CCB47 - 1	CCB47	-71	-0.0535	-	7470A	PPB	12 Aug 2021 16:34:41	U	SMPL
PB138365TB - 1	PB138348TB	-44	-0.0469	-	7470A	PPB	12 Aug 2021 16:36:56	U	SMPL
M3343-10LX5 - 1		31	-0.0285	-	7470A	PPB	12 Aug 2021 16:42:09	U	SMPL
M3343-10A - 1		15796	3.8295	-	7470A	PPB	12 Aug 2021 16:44:23	U	SMPL
CCV48 - 1	CCV48	20469	4.9730	-	7470A	PPB	12 Aug 2021 16:46:39	U	SMPL
CCB48 - 1	CCB48	-99	-0.0603	-	7470A	PPB	12 Aug 2021 16:48:55	U	SMPL