

LB116808


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

48719



A= 0.0000e+000

B= 2.0553e-004

C= -1.2292e-002

Rho= 0.9999967

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.009	0.009	102	0.000	102					-
0.2	0.200	0.206	0.006	1060	0.0 %	1060					3
2.5	2.500	2.480	-0.020	12128	0.0 %	12128					-1
5.0	5.000	4.998	-0.002	24378	0.0 %	24378					0
7.5	7.500	7.507	0.007	36583	0.0 %	36583					8
10.0	10.000	10.001	0.001	48719	0.0 %	48719					

LB116808 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Oct 2021 15:49:51

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	102	-	0.00007471B	PPB		12 Oct 2021 15:52:45	S	Std
0.2 - 1	50.2	1060	-	0.20007471B	PPB		12 Oct 2021 15:55:02	S	Std
2.5 - 1	52.5	12128	-	2.50007471B	PPB		12 Oct 2021 15:57:19	S	Std
5.0 - 1	55	24378	-	5.00007471B	PPB		12 Oct 2021 15:59:36	S	Std
7.5 - 1	57.5	36583	-	7.50007471B	PPB		12 Oct 2021 16:01:53	S	Std
10.0 - 1	510	48719	-	10.00007471B	PPB		12 Oct 2021 16:06:49	S	Std
ICV40 - 1	ICV40	19370	3.9688	-7471B	PPB		12 Oct 2021 16:09:30	U	SMPL
ICB40 - 1	ICB40	-66	-0.0259	-7471B	PPB		12 Oct 2021 16:11:44	U	SMPL
CCV41 - 1	CCV41	24087	4.9383	-7471B	PPB		12 Oct 2021 16:14:00	U	SMPL
CCB41 - 1	CCB41	-70	-0.0267	-7471B	PPB		12 Oct 2021 16:16:15	U	SMPL
CRA - 1	CRA	1002	0.1936	-7471B	PPB		12 Oct 2021 16:18:31	U	SMPL
HighStd - 1	HighStd	48904	10.0388	-7471B	PPB		12 Oct 2021 16:20:46	U	SMPL
ChkStd - 1	ChkStd	32593	6.6865	-7471B	PPB		12 Oct 2021 16:23:02	U	SMPL
PB139845BL - 1	PBS	-170	-0.0472	-7471B	PPB		12 Oct 2021 16:25:20	U	SMPL
PB139845BS - 1	LCSS	22386	4.5887	-7471B	PPB		12 Oct 2021 16:27:38	U	SMPL
M4147-01 - 1	20-20211007-WC	29863	6.1254	-7471B	PPB		12 Oct 2021 16:29:54	U	SMPL
M4150-01 - 1	FDR-20211007-WC	3538	0.7149	-7471B	PPB		12 Oct 2021 16:32:12	U	SMPL
M4158-10 - 1	T3-COMP-(1-4)	3702	0.7486	-7471B	PPB		12 Oct 2021 16:34:30	U	SMPL
M4158-20 - 1	T8-COMP-(1-4)	5071	1.0299	-7471B	PPB		12 Oct 2021 16:36:46	U	SMPL
M4159-10 - 1	T9-COMP-(1-4)	4385	0.8889	-7471B	PPB		12 Oct 2021 16:39:02	U	SMPL
CCV42 - 1	CCV42	23547	4.8273	-7471B	PPB		12 Oct 2021 16:41:17	U	SMPL
CCB42 - 1	CCB42	-81	-0.0289	-7471B	PPB		12 Oct 2021 16:43:33	U	SMPL
M4159-20 - 1	T10-COMP-(1-4)	2006	0.4000	-7471B	PPB		12 Oct 2021 16:45:50	U	SMPL
M4161-01 - 1	T-1	2812	0.5657	-7471B	PPB		12 Oct 2021 16:48:05	U	SMPL
M4163-01 - 1	TR-04-100821	6378	1.2986	-7471B	PPB		12 Oct 2021 16:55:10	U	SMPL
M4163-01DUP - 1	TR-04-100821DUP	6181	1.2581	-7471B	PPB		12 Oct 2021 16:57:26	U	SMPL
M4163-01MS - 1	TR-04-100821MS	28492	5.8436	-7471B	PPB		12 Oct 2021 16:59:42	U	SMPL
M4163-01MSD - 1	TR-04-100821MSD	26195	5.3715	-7471B	PPB		12 Oct 2021 17:04:33	U	SMPL
M4164-01 - 1	BU-03-100821	1505	0.2970	-7471B	PPB		12 Oct 2021 17:06:49	U	SMPL
M4172-01 - 1	OK-3-101121	1318	0.2586	-7471B	PPB		12 Oct 2021 17:09:07	U	SMPL
M4163-01LX5 - 1		915	0.1758	-7471B	PPB		12 Oct 2021 17:13:12	U	SMPL
M4163-01A - 1		27664	5.6734	-7471B	PPB		12 Oct 2021 17:15:27	U	SMPL
CCV43 - 1	CCV43	24434	5.0096	-7471B	PPB		12 Oct 2021 17:17:43	U	SMPL
CCB43 - 1	CCB43	-71	-0.0269	-7471B	PPB		12 Oct 2021 17:19:58	U	SMPL