

LB119069


7471B

INSTRUMENT ID: CVL

Linear

 μ Abs.:

54435



A= 0.0000e+000

B= 1.8495e-004

C= 2.7227e-002

Rho= 0.9998381

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/d
0.0	0.000	0.030	0.030	16	0.000	16					-3
0.2	0.200	0.194	-0.006	901	0.0 %	901					2
2.5	2.500	2.543	0.043	13602	0.0 %	13602					-1
5.0	5.000	4.947	-0.053	26601	0.0 %	26601					-1
7.5	7.500	7.391	-0.109	39816	0.0 %	39816					-1
10.0	10.000	10.095	0.095	54435	0.0 %	54435					1

LB119069 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 09 Mar 2022 15:53:54					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	16	-	0.0000	7471B	PPB	09 Mar 2022 15:55:02	S	Std
0.2 - 1	50.2	901	-	0.2000	7471B	PPB	09 Mar 2022 15:57:19	S	Std
2.5 - 1	52.5	13602	-	2.5000	7471B	PPB	09 Mar 2022 16:04:29	S	Std
5.0 - 1	55	26601	-	5.0000	7471B	PPB	09 Mar 2022 16:06:46	S	Std
7.5 - 1	57.5	39816	-	7.5000	7471B	PPB	09 Mar 2022 16:09:04	S	Std
10.0 - 1	51.0	54435	-	10.0000	7471B	PPB	09 Mar 2022 16:21:33	S	Std
ICV26 - 1	ICV26	20404	3.8009	-	7471B	PPB	09 Mar 2022 16:26:47	U	SMPL
ICB26 - 1	ICB26	-45	0.0189	-	7471B	PPB	09 Mar 2022 16:29:02	U	SMPL
CCV29 - 1	CCV29	27209	5.0595	-	7471B	PPB	09 Mar 2022 16:31:18	U	SMPL
CCB29 - 1	CCB29	-51	0.0178	-	7471B	PPB	09 Mar 2022 16:33:33	U	SMPL
CRA - 1	CRA	881	0.1902	-	7471B	PPB	09 Mar 2022 16:35:49	U	SMPL
HighStd - 1	HighStd	54501	10.1071	-	7471B	PPB	09 Mar 2022 16:38:05	U	SMPL
ChkStd - 1	ChkStd	37482	6.9595	-	7471B	PPB	09 Mar 2022 16:40:20	U	SMPL
PB143212BL - 1	PBS	-75	0.0134	-	7471B	PPB	09 Mar 2022 16:42:38	U	SMPL
PB143212BS - 1	LCSS	20688	3.8534	-	7471B	PPB	09 Mar 2022 16:44:55	U	SMPL
N1845-01 - 1	SAMPLE-1	8146	1.5338	-	7471B	PPB	09 Mar 2022 16:47:11	U	SMPL
N1845-02 - 1	SAMPLE-2	6877	1.2991	-	7471B	PPB	09 Mar 2022 16:49:28	U	SMPL
N1845-03 - 1	SAMPLE-3	7600	1.4328	-	7471B	PPB	09 Mar 2022 16:51:45	U	SMPL
N1845-04 - 1	SAMPLE-4	7478	1.4103	-	7471B	PPB	09 Mar 2022 16:54:02	U	SMPL
N1845-05 - 1	SAMPLE-5	7003	1.3224	-	7471B	PPB	09 Mar 2022 16:56:18	U	SMPL
CCV30 - 1	CCV30	26696	4.9646	-	7471B	PPB	09 Mar 2022 16:58:34	U	SMPL
CCB30 - 1	CCB30	-48	0.0183	-	7471B	PPB	09 Mar 2022 17:00:50	U	SMPL
N1845-06 - 1	SAMPLE-6	10582	1.9843	-	7471B	PPB	09 Mar 2022 17:03:06	U	SMPL
N1845-07 - 1	SAMPLE-7	8503	1.5998	-	7471B	PPB	09 Mar 2022 17:05:21	U	SMPL
N1845-08 - 1	SAMPLE-8	6965	1.3154	-	7471B	PPB	09 Mar 2022 17:07:37	U	SMPL
N1845-09 - 1	SAMPLE-9	7118	1.3437	-	7471B	PPB	09 Mar 2022 17:09:53	U	SMPL
N1845-10 - 1	SAMPLE-10	9189	1.7267	-	7471B	PPB	09 Mar 2022 17:12:09	U	SMPL
N1845-11 - 1	SAMPLE-11	9427	1.7707	-	7471B	PPB	09 Mar 2022 17:14:26	U	SMPL
N1845-12 - 1	SAMPLE-12	8641	1.6254	-	7471B	PPB	09 Mar 2022 17:16:42	U	SMPL
N1845-13 - 1	SAMPLE-13	15079	2.8161	-	7471B	PPB	09 Mar 2022 17:18:59	U	SMPL
N1845-14 - 1	SAMPLE-14	10107	1.8965	-	7471B	PPB	09 Mar 2022 17:21:17	U	SMPL
N1845-15 - 1	SAMPLE-15	9045	1.7001	-	7471B	PPB	09 Mar 2022 17:23:32	U	SMPL
CCV31 - 1	CCV31	26813	4.9862	-	7471B	PPB	09 Mar 2022 17:25:48	U	SMPL
CCB31 - 1	CCB31	-43	0.0193	-	7471B	PPB	09 Mar 2022 17:28:04	U	SMPL
N1845-15DUP - 1	SAMPLE-15DUP	9062	1.7032	-	7471B	PPB	09 Mar 2022 17:30:20	U	SMPL
N1845-15MS - 1	SAMPLE-15MS	32888	6.1098	-	7471B	PPB	09 Mar 2022 17:32:36	U	SMPL
N1845-15MSD - 1	SAMPLE-15MSD	33145	6.1573	-	7471B	PPB	09 Mar 2022 17:34:52	U	SMPL
PB143213BL - 1	PBS	-68	0.0147	-	7471B	PPB	09 Mar 2022 17:37:09	U	SMPL
PB143213BS - 1	LCSS	20143	3.7526	-	7471B	PPB	09 Mar 2022 17:39:24	U	SMPL
N1843-01 - 1	FS-1	18092	3.3733	-	7471B	PPB	09 Mar 2022 17:41:41	U	SMPL
N1843-03 - 1	FS-2	4749	0.9055	-	7471B	PPB	09 Mar 2022 17:43:58	U	SMPL
N1851-01 - 1	DUM-34	629	0.1436	-	7471B	PPB	09 Mar 2022 17:46:15	U	SMPL
N1851-03 - 1	DUM-35	1065	0.2242	-	7471B	PPB	09 Mar 2022 17:48:32	U	SMPL
N1851-05 - 1	DUM-36	3319	0.6411	-	7471B	PPB	09 Mar 2022 17:50:48	U	SMPL
CCV32 - 1	CCV32	26817	4.9870	-	7471B	PPB	09 Mar 2022 17:53:04	U	SMPL
CCB32 - 1	CCB32	-45	0.0189	-	7471B	PPB	09 Mar 2022 17:55:19	U	SMPL
N1851-07 - 1	DUM-37	856	0.1855	-	7471B	PPB	09 Mar 2022 17:57:35	U	SMPL
N1851-09 - 1	DUM-38	705	0.1576	-	7471B	PPB	09 Mar 2022 17:59:51	U	SMPL
N1851-09DUP - 1	DUM-38DUP	702	0.1571	-	7471B	PPB	09 Mar 2022 18:02:07	U	SMPL
N1851-09MS - 1	DUM-38MS	21310	3.9685	-	7471B	PPB	09 Mar 2022 18:04:22	U	SMPL
N1851-09MSD - 1	DUM-38MSD	21376	3.9807	-	7471B	PPB	09 Mar 2022 18:06:38	U	SMPL
N1845-15LX5 - 1		1693	0.3403	-	7471B	PPB	09 Mar 2022 18:08:55	U	SMPL
N1845-15A - 1		32395	6.0186	-	7471B	PPB	09 Mar 2022 18:11:12	U	SMPL
N1851-09LX5 - 1		45	0.0355	-	7471B	PPB	09 Mar 2022 18:13:29	U	SMPL
N1851-09A - 1		21572	4.0169	-	7471B	PPB	09 Mar 2022 18:15:46	U	SMPL
CCV33 - 1	CCV33	26708	4.9668	-	7471B	PPB	09 Mar 2022 18:18:02	U	SMPL
CCB33 - 1	CCB33	5	0.0282	-	7471B	PPB	09 Mar 2022 18:20:18	U	SMPL