

LB120369


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs.:

54533



A= 0.0000e+000

B= 1.8370e-004

C= 4.4976e-002

Rho= 0.9999151

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.0	0.000	0.033	0.033	-65	0.000	-65					13
0.2	0.200	0.226	0.026	985	0.0 %	985					0
2.5	2.500	2.490	-0.010	13309	0.0 %	13309					-2
5.0	5.000	4.914	-0.086	26504	0.0 %	26504					0
7.5	7.500	7.475	-0.025	40445	0.0 %	40445					1
10.0	10.000	10.063	0.063	54533	0.0 %	54533					

LB120369 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 26 May 2022 13:20:36					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-65	-	0.0000	7471B	PPB	26 May 2022 13:27:21	S	Std
0.2 - 1	50.2	985	-	0.2000	7471B	PPB	26 May 2022 13:29:37	S	Std
2.5 - 1	52.5	13309	-	2.5000	7471B	PPB	26 May 2022 13:31:54	S	Std
5.0 - 1	55	26504	-	5.0000	7471B	PPB	26 May 2022 13:34:11	S	Std
7.5 - 1	57.5	40445	-	7.5000	7471B	PPB	26 May 2022 13:36:28	S	Std
10.0 - 1	510	54533	-	10.0000	7471B	PPB	26 May 2022 13:38:46	S	Std
ICV47 - 1	ICV47	21192	3.9380	-	7471B	PPB	26 May 2022 13:41:32	U	SMPL
ICB47 - 1	ICB47	-84	0.0295	-	7471B	PPB	26 May 2022 13:43:46	U	SMPL
CCV67 - 1	CCV67	27230	5.0471	-	7471B	PPB	26 May 2022 13:46:02	U	SMPL
CCB67 - 1	CCB67	-96	0.0273	-	7471B	PPB	26 May 2022 13:48:17	U	SMPL
CRA - 1	CRA	1002	0.2290	-	7471B	PPB	26 May 2022 13:50:33	U	SMPL
HighStd - 1	HighStd	54980	10.1448	-	7471B	PPB	26 May 2022 13:52:48	U	SMPL
ChkStd - 1	ChkStd	38487	7.1150	-	7471B	PPB	26 May 2022 13:55:03	U	SMPL
PB145114BL - 1	PBS	-37	0.0382	-	7471B	PPB	26 May 2022 14:00:05	U	SMPL
PB145114BS - 1	LCSS	22462	4.1713	-	7471B	PPB	26 May 2022 14:02:20	U	SMPL
N2930-01 - 1	P002-SS003-0006-01	3516	0.6909	-	7471B	PPB	26 May 2022 14:04:36	U	SMPL
N2930-02 - 1	P002-SS003-0612-01	4748	0.9172	-	7471B	PPB	26 May 2022 14:06:54	U	SMPL
N2930-03 - 1	P002-SS003-1218-01	6810	1.2960	-	7471B	PPB	26 May 2022 14:09:11	U	SMPL
N2930-04 - 1	P002-SS004-0006-01	10580	1.9885	-	7471B	PPB	26 May 2022 14:11:28	U	SMPL
N2930-07 - 1	P005-SS003-0006-01	15113	2.8212	-	7471B	PPB	26 May 2022 14:13:43	U	SMPL
CCV68 - 1	CCV68	27291	5.0583	-	7471B	PPB	26 May 2022 14:15:59	U	SMPL
CCB68 - 1	CCB68	-56	0.0347	-	7471B	PPB	26 May 2022 14:18:15	U	SMPL
N2930-08 - 1	P005-SS003-0612-01	20460	3.8035	-	7471B	PPB	26 May 2022 14:20:32	U	SMPL
N2930-09 - 1	P005-SS003-1218-01	12844	2.4044	-	7471B	PPB	26 May 2022 14:22:48	U	SMPL
N2930-10 - 1	P005-SS003-1824-01	8237	1.5581	-	7471B	PPB	26 May 2022 14:25:04	U	SMPL
N2930-11 - 1	P005-SS003-2430-01	4226	0.8213	-	7471B	PPB	26 May 2022 14:27:20	U	SMPL
N2930-12 - 1	P005-SS004-0006-01	5074	0.9771	-	7471B	PPB	26 May 2022 14:29:36	U	SMPL
N2930-17 - 1	P006-SS002-0006-01	3487	0.6855	-	7471B	PPB	26 May 2022 14:31:52	U	SMPL
N2930-18 - 1	P006-SS002-0612-01	4108	0.7996	-	7471B	PPB	26 May 2022 14:34:09	U	SMPL
N2930-19 - 1	P006-SS002-1218-01	6296	1.2016	-	7471B	PPB	26 May 2022 14:36:26	U	SMPL
N2930-20 - 1	P006-SS002-1824-01	6814	1.2967	-	7471B	PPB	26 May 2022 14:38:43	U	SMPL
N2930-21 - 1	P006-SS002-2430-01	5754	1.1020	-	7471B	PPB	26 May 2022 14:40:58	U	SMPL
CCV69 - 1	CCV69	26868	4.9806	-	7471B	PPB	26 May 2022 14:43:14	U	SMPL
CCB69 - 1	CCB69	-75	0.0312	-	7471B	PPB	26 May 2022 14:45:29	U	SMPL
N2930-21DUP - 1	P006-SS002-2430-01DUP	5910	1.1306	-	7471B	PPB	26 May 2022 14:47:46	U	SMPL
N2930-21MS - 1	P006-SS002-2430-01MS	28935	5.3603	-	7471B	PPB	26 May 2022 14:50:02	U	SMPL
N2930-21MSD - 1	P006-SS002-2430-01MSD	29427	5.4507	-	7471B	PPB	26 May 2022 14:52:17	U	SMPL
N2934-02 - 1	P025-SS001-0006-01	13941	2.6059	-	7471B	PPB	26 May 2022 14:54:34	U	SMPL
N2934-03 - 1	P025-SS001-0612-01	11120	2.0877	-	7471B	PPB	26 May 2022 14:56:52	U	SMPL
N2934-04 - 1	P025-SS001-1218-01	9782	1.8419	-	7471B	PPB	26 May 2022 14:59:09	U	SMPL
N2934-05 - 1	P025-SS001-1824-01	6173	1.1790	-	7471B	PPB	26 May 2022 15:01:26	U	SMPL
N2934-06 - 1	P025-SS001-2430-01	1842	0.3834	-	7471B	PPB	26 May 2022 15:03:43	U	SMPL
PB145115BL - 1	PBS	-60	0.0340	-	7471B	PPB	26 May 2022 15:06:01	U	SMPL
PB145115BS - 1	LCSS	22226	4.1279	-	7471B	PPB	26 May 2022 15:08:18	U	SMPL
CCV70 - 1	CCV70	27488	5.0945	-	7471B	PPB	26 May 2022 15:10:32	U	SMPL
CCB70 - 1	CCB70	-74	0.0314	-	7471B	PPB	26 May 2022 15:12:49	U	SMPL
N2656-11 - 1	PT-MET-SOIL	324046	59.5723	-	7471B	PPB	26 May 2022 15:15:07	U	SMPL
N2924-01 - 1	P024-SS009-0006-01	4102	0.7985	-	7471B	PPB	26 May 2022 15:17:22	U	SMPL
N2924-06 - 1	P024-SS010-0006-01	8443	1.5960	-	7471B	PPB	26 May 2022 15:19:52	U	SMPL
N2932-03 - 1	P016-SS003-0006-01	3319	0.6547	-	7471B	PPB	26 May 2022 15:22:08	U	SMPL
N2932-11 - 1	P016-SS004-0006-01	4228	0.8217	-	7471B	PPB	26 May 2022 15:24:23	U	SMPL
N2932-16 - 1	P016-SS005-0006-01	4851	0.9361	-	7471B	PPB	26 May 2022 15:26:40	U	SMPL
N2932-16DUP - 1	P016-SS005-0006-01DUP	5468	1.0494	-	7471B	PPB	26 May 2022 15:28:56	U	SMPL
N2932-17 - 1	N2932-16MS	28178	5.2213	-	7471B	PPB	26 May 2022 15:31:13	U	SMPL
N2932-18 - 1	N2932-16MSD	26733	4.9558	-	7471B	PPB	26 May 2022 15:33:30	U	SMPL
N2932-19 - 1	P016-SS005-0006-02	4744	0.9165	-	7471B	PPB	26 May 2022 15:35:47	U	SMPL
CCV71 - 1	CCV71	26638	4.9384	-	7471B	PPB	26 May 2022 15:38:04	U	SMPL
CCB71 - 1	CCB71	-73	0.0316	-	7471B	PPB	26 May 2022 15:40:20	U	SMPL
N2932-24 - 1	P023-SS001-0006-01	5976	1.1428	-	7471B	PPB	26 May 2022 15:42:37	U	SMPL
N2933-01 - 1	P023-SS001-0612-01	5611	1.0757	-	7471B	PPB	26 May 2022 15:44:53	U	SMPL
N2933-02 - 1	P023-SS001-1218-01	5109	0.9835	-	7471B	PPB	26 May 2022 15:47:08	U	SMPL
N2933-03 - 1	P023-SS001-1824-01	4571	0.8847	-	7471B	PPB	26 May 2022 15:49:24	U	SMPL
N2933-04 - 1	P023-SS001-2430-01	2438	0.4928	-	7471B	PPB	26 May 2022 15:51:40	U	SMPL
N2933-05 - 1	P023-SS002-0006-01	12511	2.3432	-	7471B	PPB	26 May 2022 15:53:56	U	SMPL
N2933-13 - 1	P023-SS003-0006-01	9230	1.7405	-	7471B	PPB	26 May 2022 15:56:13	U	SMPL
N2933-18 - 1	P023-SS004-0006-01	15570	2.9052	-	7471B	PPB	26 May 2022 15:58:30	U	SMPL
N2933-23 - 1	P023-SS005-0006-01	43746	8.0811	-	7471B	PPB	26 May 2022 16:00:47	U	SMPL
CCV72 - 1	CCV72	26783	4.9650	-	7471B	PPB	26 May 2022 16:03:04	U	SMPL
CCB72 - 1	CCB72	-151	0.0172	-	7471B	PPB	26 May 2022 16:05:21	U	SMPL
N2930-21LX5 - 1		1047	0.2373	-	7471B	PPB	26 May 2022 16:07:39	U	SMPL
N2930-21A - 1		29813	5.5216	-	7471B	PPB	26 May 2022 16:09:55	U	SMPL
N2932-16LX5 - 1		768	0.1861	-	7471B	PPB	26 May 2022 16:12:11	U	SMPL
N2932-16A - 1		27555	5.1068	-	7471B	PPB	26 May 2022 16:14:28	U	SMPL
N2656DLX10 - 1	PT-MET-SOIL	32835	6.0768	-	7471B	PPB	26 May 2022 16:23:51	U	SMPL
CCV73 - 1	CCV73	26816	4.9711	-	7471B	PPB	26 May 2022 16:33:18	U	SMPL
CCB73 - 1	CCB73	-122	0.0226	-	7471B	PPB	26 May 2022 16:35:36	U	SMPL