


7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.000	-0.000	112	0.000	112				
0.2	0.200	0.179	-0.021	1134	0.0 %	1134				
2.5	2.500	2.470	-0.030	14212	0.0 %	14212				
5.0	5.000	5.111	0.111	29283	0.0 %	29283				
7.5	7.500	7.460	-0.040	42692	0.0 %	42692				
10.0	10.000	9.981	-0.019	57082	0.0 %	57082				

LB120715 Instruent ID:CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Jun 2022 14:10:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	112	-	0.0000	7471B	PPB	16 Jun 2022 14:38:34	S	Std
0.2 - 1	50.2	1134	-	0.2000	7471B	PPB	16 Jun 2022 14:40:51	S	Std
2.5 - 1	52.5	14212	-	2.5000	7471B	PPB	16 Jun 2022 14:43:07	S	Std
5.0 - 1	55.0	29283	-	5.0000	7471B	PPB	16 Jun 2022 14:45:24	S	Std
7.5 - 1	57.5	42692	-	7.5000	7471B	PPB	16 Jun 2022 14:47:41	S	Std
10.0 - 1	50.0	57082	-	10.0000	7471B	PPB	16 Jun 2022 14:49:58	S	Std
ICV77 - 1	ICV77	23976	4.1808	-	7471B	PPB	16 Jun 2022 14:57:38	U	SMPL
ICB77 - 1	ICB77	34	-0.0140	-	7471B	PPB	16 Jun 2022 14:59:53	U	SMPL
CCV116 - 1	CCV116	30890	5.3921	-	7471B	PPB	16 Jun 2022 15:02:09	U	SMPL
CCB116 - 1	CCB116	34	-0.0140	-	7471B	PPB	16 Jun 2022 15:04:24	U	SMPL
CRA - 1	CRA	1119	0.1761	-	7471B	PPB	16 Jun 2022 15:06:40	U	SMPL
HighStd - 1	HighStd	57410	10.0386	-	7471B	PPB	16 Jun 2022 15:08:55	U	SMPL
ChkStd - 1	ChkStd	39306	6.8667	-	7471B	PPB	16 Jun 2022 15:11:10	U	SMPL
PB145584BL - 1	PB145584BL	-51	-0.0289	-	7471B	PPB	16 Jun 2022 15:13:28	U	SMPL
PB145584BS - 1	PB145584BS	19800	3.4491	-	7471B	PPB	16 Jun 2022 15:15:47	U	SMPL
N3336-01 - 1	HD-02-061322	2275	0.3786	-	7471B	PPB	16 Jun 2022 15:18:03	U	SMPL
N3337-01 - 1	SU-02-061322	3487	0.5910	-	7471B	PPB	16 Jun 2022 15:20:20	U	SMPL
N3340-01 - 1	CS-1	12410	2.1543	-	7471B	PPB	16 Jun 2022 15:22:37	U	SMPL
N3340-02 - 1	CS-2	2541	0.4252	-	7471B	PPB	16 Jun 2022 15:24:53	U	SMPL
N3340-03 - 1	CS-3	674	0.0981	-	7471B	PPB	16 Jun 2022 15:27:09	U	SMPL
CCV117 - 1	CCV117	29627	5.1708	-	7471B	PPB	16 Jun 2022 15:29:24	U	SMPL
CCB117 - 1	CCB117	-18	-0.0231	-	7471B	PPB	16 Jun 2022 15:31:39	U	SMPL
N3340-03DUP - 1	CS-3DUP	744	0.1104	-	7471B	PPB	16 Jun 2022 15:33:59	U	SMPL
N3340-03MS - 1	CS-3MS	19723	3.4356	-	7471B	PPB	16 Jun 2022 15:36:14	U	SMPL
N3340-03MSD - 1	CS-3MSD	19524	3.4007	-	7471B	PPB	16 Jun 2022 15:38:29	U	SMPL
N3347-01 - 1	CL-02-061422	7998	1.3813	-	7471B	PPB	16 Jun 2022 15:40:46	U	SMPL
N3348-01 - 1	TR-5-061422	5215	0.8937	-	7471B	PPB	16 Jun 2022 15:43:01	U	SMPL
N3351-05 - 1	40582-84	763	0.1137	-	7471B	PPB	16 Jun 2022 15:45:17	U	SMPL
N3351-06 - 1	40585-87	532	0.0732	-	7471B	PPB	16 Jun 2022 15:47:34	U	SMPL
N3351-07 - 1	40588-90	708	0.1041	-	7471B	PPB	16 Jun 2022 15:49:51	U	SMPL
N3351-08 - 1	40591-94	1167	0.1845	-	7471B	PPB	16 Jun 2022 15:52:08	U	SMPL
N3353-01 - 1	PL-01-061422	2522	0.4219	-	7471B	PPB	16 Jun 2022 15:54:24	U	SMPL
CCV118 - 1	CCV118	30615	5.3440	-	7471B	PPB	16 Jun 2022 15:56:39	U	SMPL
CCB118 - 1	CCB118	-15	-0.0226	-	7471B	PPB	16 Jun 2022 15:58:55	U	SMPL
N3364-01 - 1	OK-02-061522	1292	0.2064	-	7471B	PPB	16 Jun 2022 16:01:12	U	SMPL
N3340-03LX5 - 1		183	0.0121	-	7471B	PPB	16 Jun 2022 16:03:28	U	SMPL
N3340-03A - 1		21745	3.7899	-	7471B	PPB	16 Jun 2022 16:09:11	U	SMPL
CCV119 - 1	CCV119	30702	5.3592	-	7471B	PPB	16 Jun 2022 16:11:25	U	SMPL
CCB119 - 1	CCB119	-38	-0.0266	-	7471B	PPB	16 Jun 2022 16:13:42	U	SMPL