

Clear table


Instrument ID: IC-2										Analyst :		IZ		Method:		300.0 / 9056A	
ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst						
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-09-14_	9/14/22 16:03		IZ/MA						
STD2	0.4880	0.6710	0.7180	2.2970	0.5780	1.1500	3.5420	2022-09-14_	9/14/22 16:34		IZ/MA						
STD3	0.7620	1.2030	1.2020	4.0290	1.0030	1.9220	6.0530	2022-09-14_	9/14/22 17:06		IZ/MA						
STD4	1.0100	1.4730	1.4500	4.8790	1.2200	2.3670	7.3010	2022-09-14_	9/14/22 17:38		IZ/MA						
STD5	1.9310	2.9410	2.9000	9.7360	2.4380	5.0300	14.5000	2022-09-14_	9/14/22 18:09		IZ/MA						
STD6	3.9460	5.9540	5.9720	19.8380	4.9490	10.1030	29.6990	2022-09-14_	9/14/22 18:41		IZ/MA						
STD7	5.0630	7.5570	7.5590	25.2210	6.3120	12.4270	37.9070	2022-09-14_	9/14/22 19:13		IZ/MA						
ICV	1.8230	2.9100	2.9030	9.5920	2.4060	4.8250	14.2610	2022-09-14_	9/14/22 19:44		IZ/MA						
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-09-14_	9/14/22 20:16		IZ/MA						
CCV	1.9730	2.9170	2.9510	9.5380	2.3950	5.1290	14.3000	2022-09-30_	9/30/22 10:24		IZ/MA						
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-09-30_	9/30/22 11:02		IZ/MA						
N4894-03	0.5190	181.2060	0.0000	0.0000	2.8660	0.0000	16.0570	2022-09-30_	9/30/22 11:30		IZ/MA						
N4894-05	0.0000	509.1500	0.0000	0.0000	0.0000	0.0000	46.2000	2022-09-30_	9/30/22 12:01		IZ/MA						
N4894-07	0.0000	498.4100	0.0000	0.0000	0.0000	0.0000	45.3220	2022-09-30_	9/30/22 12:32		IZ/MA						
N4894-01	0.0000	179.7720	0.0000	0.0000	3.0620	0.0000	35.1870	2022-09-30_	9/30/22 13:02		IZ/MA						
N4894-01MS	3.3080	175.4430	2.0140	9.7430	5.5270	10.0470	49.0490	2022-09-30_	9/30/22 13:33		IZ/MA						
N4894-01MSD	3.1150	173.7580	2.1480	10.1550	5.6490	10.8670	49.2450	2022-09-30_	9/30/22 14:04		IZ/MA						
LB122311BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-09-30_	9/30/22 14:34		IZ/MA						
LB122311BSW	1.9230	2.9200	3.0770	9.5980	2.3930	6.2090	14.2500	2022-09-30_	9/30/22 15:05		IZ/MA						
CCV	1.8550	2.9380	3.0530	9.6390	2.4030	5.7870	14.3440	2022-09-30_	9/30/22 16:10		IZ/MA						
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-09-30_	9/30/22 16:41		IZ/MA						
N4894-01DLX50	0.0000	3.0500	0.0000	0.0000	0.0000	0.0000	1.4340	2022-09-30_	9/30/22 17:12		IZ/MA						
N4894-03DLX50	0.0000	2.8910	0.0000	0.0000	0.0000	0.0000	0.0000	2022-09-30_	9/30/22 17:42		IZ/MA						
N4894-05DLX5	0.0000	151.4170	0.0000	0.0000	0.0000	0.0000	8.6310	2022-09-30_	9/30/22 18:13		IZ/MA						
N4894-05DLX200	0.0000	3.0330	0.0000	0.0000	0.0000	0.0000	0.0000	2022-09-30_	9/30/22 18:44		IZ/MA						
N4894-07DLX5	0.0000	152.1810	0.0000	0.0000	0.0000	0.0000	8.5940	2022-09-30_	9/30/22 19:14		IZ/MA						
N4894-07DLX200	0.0000	3.0060	0.0000	0.0000	0.0000	0.0000	0.0000	2022-09-30_	9/30/22 19:45		IZ/MA						
CCV	1.9370	2.9960	3.0730	9.7750	2.4350	6.4050	14.5820	2022-09-30_	9/30/22 20:46		IZ/MA						
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-09-30_	9/30/22 21:17		IZ/MA						

Clear table

Instrument ID: IC-2									
Analyst : IZ				Method: 300.0 / 9056A					
ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.9510	2022-09-14_	9/14/22 16:03
STD2	0.4880	0.6710	0.7180	2.2970	0.5780	1.1500	3.5420	2022-09-14_	9/14/22 16:34
STD3	0.7620	1.2030	1.2020	4.0290	1.0030	1.9220	6.0630	2022-09-14_	9/14/22 17:06
STD4	1.0100	1.4730	1.4500	4.8790	1.2200	2.3670	7.3010	2022-09-14_	9/14/22 17:38
STD5	1.9310	2.9410	2.9000	9.7360	2.4380	5.0300	14.5000	2022-09-14_	9/14/22 18:09
STD6	3.9460	5.9540	5.9720	19.8380	4.9490	10.1030	29.6990	2022-09-14_	9/14/22 18:41
STD7	5.0630	7.5570	7.5590	25.2210	6.3120	12.4270	37.9070	2022-09-14_	9/14/22 19:13
ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4		
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000		
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000		
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000		
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000		
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000		
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000		
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.0000		
ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4	Initial wt/ Final Vol	Analyst
STD1									
STD2	22.0000	11.8333	19.6667	14.8500	15.6000	15.0000	18.0667		IZ/MA
STD3	-4.7500	0.2500	0.1667	0.7250	0.3000	-3.9000	0.8833		IZ/MA
STD4	1.0000	-1.8000	-3.3333	-2.4200	-2.4000	-5.3200	-2.6533		IZ/MA
STD5	-3.4500	-1.9667	-3.3333	-2.6400	-2.4800	0.6000	-3.3333		IZ/MA
STD6	-1.3500	-0.7667	-0.4667	-0.8100	-1.0200	1.0300	-1.0033		IZ/MA
STD7	1.2600	0.7600	0.7867	0.8840	0.9920	-0.5840	2.4514		IZ/MA

CALIBRATION OF COMPONENT F-

Method: AnionsIC2-091422.mtw
 Equation: $Q = 0.0683307 \cdot A + 3.59318$
 RSD: 3.280 %
 Correlation coefficient: 0.999413




K3 = 0 K2 = 0 K1 = 0.0683307 K0 = 3.59318
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	2.91	90.14	0.4	20	5.877	
3	6.268	170.4	0.8	20	5.877	
4	8.084	242.9	1	20	5.877	
5	17.91	512.7	2	20	5.877	
6	41.6	1103	4	20	5.877	
7	50.64	1429	5	20	5.877	

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-091422.mtw
 Equation: $Q = 0.120741 \cdot A + 2.87223$
 RSD: 1.828 %
 Correlation coefficient: 0.999818




K3 = 0 K2 = 0 K1 = 0.120741 K0 = 2.87223
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.226	87.36	0.6	20	9.035		
3	14.06	175.4	1.2	20	9.035		
4	17.75	220.3	1.5	20	9.035		
5	37.57	463.4	3	20	9.035		
6	79.02	962.5	6	20	9.035		
7	99.56	1228	7.5	20	9.035		

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-091422.mtw
 Equation: $Q = 0.0546844 \cdot A + 5.14338$
 RSD: 2.653 %
 Correlation coefficient: 0.999616




K3 = 0 K2 = 0 K1 = 0.0546844 K0 = 5.14338
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	10.35	168.4	0.6	20	10.91		
3	20.68	345.6	1.2	20	10.91		
4	26.65	436.1	1.5	20	10.91		
5	57.51	966.6	3	20	10.91		
6	117.1	2090	6	20	10.91		
7	147.6	2671	7.5	20	10.91		

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-091422.mtw
 Equation: $Q = 0.265411 \cdot A + 11.1115$
 RSD: 2.266 %
 Correlation coefficient: 0.999720




K3 = 0 K2 = 0 K1 = 0.265411 K0 = 11.1115
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	8.015	131.2	2	20	14.56		
3	15.95	261.8	4	20	14.56		
4	20.17	325.8	5	20	14.56		
5	43.67	691.8	10	20	14.56		
6	94.36	1453	20	20	14.56		
7	119.7	1859	25	20	14.56		

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-091422.mtw
 Equation: $Q = 0.0466297 \cdot A + 2.81687$
 RSD: 2.400 %
 Correlation coefficient: 0.999686




K3 = 0 K2 = 0 K1 = 0.0466297 K0 = 2.81687
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	9.456	187.7	0.5	20	16.64		
3	18.73	369.8	1	20	16.64		
4	23.66	462.7	1.25	20	16.64		
5	50.7	985.3	2.5	20	16.64		
6	108.8	2062	5	20	16.64		
7	138	2647	6.25	20	16.64		

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-091422.mtw
 Equation: $Q = 0.0575109 \cdot A + 7.83792$
 RSD: 2.284 %
 Correlation coefficient: 0.999716




K3 = 0 K2 = 0 K1 = 0.0575109 K0 = 7.83792
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	6.503	263.6	1	20	21.97		
3	13.52	532.2	2	20	21.97		
4	17.75	686.9	2.5	20	21.97		
5	43.29	1613	5	20	21.97		
6	96.49	3377	10	20	21.97		
7	121.5	4185	12.5	20	21.97		

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-091422.mtw
 Equation: $Q = 0.152891 \cdot A + 16.4169$
 RSD: 2.781 %
 Correlation coefficient: 0.999578



K3 = 0 K2 = 0 K1 = 0.152891 K0 = 16.4169
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0.5192	17.02		0	20		26.93
2	12.77	355.9		3	20		26.93
3	24.66	684.4		6	20		26.93
4	30.87	847.6		7.5	20		26.93
5	66.06	1789		15	20		26.93
6	142.9	3778		30	20		26.93
7	182.2	4851		37.5	20		26.93

Report date: 9/16/2022 12:48:59 PM
Printed by: wet


Ident: STD1
Analysis from: 9/14/2022 4:03:21 PM
File: _2022-09-14_

Last save: 9/16/2022 12:48:53 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 31861

Last save: 9/14/2022 3:48:

SAMPLE:
:
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 9/16/2022 12:50:50 PM
Printed by: wet


Ident: STD2
Analysis from: 9/14/2022 4:34:59 PM
File: _2022-09-14_

Last save: 9/15/2022 9:28:59 AM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 31862

Last save: 9/14/2022 3:48:

SAMPLE:
: IZ/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.353	2.91	5.08	90.141	7.02
2	8.93	0.171	7.23	12.63	87.363	6.80
3	10.81	0.209	10.35	18.08	168.448	13.12
4	13.96	0.239	8.01	14.00	131.211	10.22
5	16.07	0.284	9.46	16.53	187.685	14.61
6	22.68	0.485	6.50	11.36	263.587	20.52
7	26.40	0.417	12.77	22.31	355.910	27.71
7	29.00	0.308	57.22	99.99	1284.344	100.00

This report has been created by IC Net
METROHM LTD

Report date: 9/16/2022 12:50:42 PM
Printed by: wet


Ident: STD3
Analysis from: 9/14/2022 5:06:37 PM
File: _2022-09-14_

Last save: 9/15/2022 9:28:59 AM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 31863

Last save: 9/14/2022 3:48:

SAMPLE:
: IZ/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.313	6.27	5.50	170.351	6.71
2	8.94	0.173	14.06	12.34	175.399	6.91
3	10.83	0.214	20.68	18.16	345.569	13.61
4	13.96	0.237	15.95	14.01	261.769	10.31
5	16.05	0.282	18.73	16.45	369.754	14.56
6	22.72	0.474	13.52	11.87	532.157	20.96
7	26.45	0.415	24.66	21.66	684.366	26.95
7	29.00	0.301	113.87	99.99	2539.364	100.00

This report has been created by IC Net
METROHM LTD

Report date: 9/16/2022 12:50:34 PM
Printed by: wet


Ident: STD4
Analysis from: 9/14/2022 5:38:14 PM
File: _2022-09-14_

Last save: 9/15/2022 9:28:59 AM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 31864

Last save: 9/14/2022 3:48:

SAMPLE:
: IZ/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.84	0.335	8.08	5.58	242.893	7.54
2	8.98	0.173	17.75	12.24	220.281	6.84
3	10.87	0.213	26.65	18.39	436.108	13.53
4	14.01	0.235	20.17	13.92	325.768	10.11
5	16.10	0.280	23.66	16.33	462.665	14.36
6	22.79	0.466	17.75	12.25	686.901	21.32
7	26.54	0.411	30.87	21.30	847.630	26.31
7	29.00	0.302	144.94	100.00	3222.244	100.00

This report has been created by IC Net
METROHM LTD

Report date: 9/16/2022 12:50:25 PM
Printed by: wet


Ident: STD5
Analysis from: 9/14/2022 6:09:52 PM
File: _2022-09-14_

Last save: 9/15/2022 9:28:59 AM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 31865

Last save: 9/14/2022 3:48:

SAMPLE:
: IZ/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.81	0.332	17.91	5.65	512.733	7.30
2	8.92	0.168	37.57	11.86	463.447	6.60
3	10.78	0.212	57.51	18.16	966.633	13.77
4	13.92	0.228	43.67	13.79	691.797	9.85
5	15.96	0.274	50.70	16.01	985.257	14.03
6	22.64	0.452	43.29	13.67	1613.076	22.97
7	26.38	0.404	66.06	20.86	1789.345	25.48
7	29.00	0.296	316.71	100.00	7022.286	100.00

This report has been created by IC Net
METROHM LTD

Report date: 9/16/2022 12:49:41 PM
Printed by: wet


Ident: STD6
Analysis from: 9/14/2022 6:41:30 PM
File: _2022-09-14_

Last save: 9/15/2022 9:28:59 AM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 31866

Last save: 9/14/2022 3:48:

SAMPLE:
: IZ/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.313	41.60	6.11	1102.501	7.44
2	8.94	0.168	79.02	11.61	962.509	6.49
3	10.79	0.227	117.15	17.22	2089.931	14.10
4	13.96	0.223	94.36	13.87	1453.009	9.80
5	15.97	0.267	108.83	16.00	2062.218	13.91
6	22.65	0.428	96.49	14.18	3377.283	22.78
7	26.42	0.395	142.91	21.01	3777.591	25.48
7	29.00	0.289	680.35	100.00	14825.041	100.00

This report has been created by IC Net
METROHM LTD

Report date: 9/16/2022 12:49:48 PM
Printed by: wet


Ident: STD7
Analysis from: 9/14/2022 7:13:08 PM
File: _2022-09-14_

Last save: 9/15/2022 9:28:59 AM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 31867

Last save: 9/14/2022 3:48:

SAMPLE:
: IZ/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.86	0.353	50.64	5.89	1429.434	7.58
2	8.97	0.168	99.56	11.59	1228.028	6.51
3	10.81	0.227	147.60	17.18	2670.529	14.15
4	14.01	0.224	119.70	13.93	1858.672	9.85
5	16.00	0.268	137.96	16.06	2646.994	14.03
6	22.68	0.425	121.48	14.14	4185.371	22.18
7	26.47	0.397	182.19	21.21	4851.270	25.71
7	29.00	0.295	859.13	100.00	18870.297	100.00

This report has been created by IC Net
METROHM LTD

Report date: 9/16/2022 12:49:56 PM
Printed by: wet


Ident: ICV
Analysis from: 9/14/2022 7:44:45 PM
File: _2022-09-14_

Last save: 9/14/2022 8:13:46 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 31868

Last save: 9/14/2022 7:42:

SAMPLE:
: IZ/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.81	0.305	17.95	5.58	481.037	7.01
2	8.94	0.164	39.39	12.24	458.193	6.68
3	10.80	0.209	61.20	19.02	967.663	14.11
4	13.96	0.225	44.05	13.69	680.951	9.93
5	16.01	0.269	51.35	15.96	971.727	14.17
6	22.65	0.445	41.82	13.00	1541.547	22.47
7	26.39	0.399	66.03	20.52	1758.086	25.63
7	29.00	0.288	321.79	100.00	6859.204	100.00

This report has been created by IC Net
METROHM LTD

Report date: 9/16/2022 12:50:15 PM
Printed by: wet


Ident: ICB
Analysis from: 9/14/2022 8:16:23 PM
File: _2022-09-14_

Last save: 9/16/2022 12:50:12 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 31869

Last save: 9/14/2022 7:42:

SAMPLE:
: IZ/AP
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:49:03 AM
Printed by: wet


Ident: CCV
Analysis from: 9/30/2022 10:24:19 AM
File: _2022-09-30_

Last save: 9/30/2022 10:52:20 AM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32042

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 35
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.94	0.336	18.32	5.94	524.950	7.47
2	9.19	0.180	35.97	11.66	459.379	6.54
3	11.16	0.223	59.24	19.20	985.347	14.03
4	14.54	0.249	39.78	12.89	676.844	9.64
5	16.71	0.302	45.86	14.87	966.628	13.76
6	22.61	0.461	44.05	14.28	1647.353	23.45
7	26.13	0.409	65.27	21.16	1763.186	25.10
7	28.00	0.309	308.49	99.99	7023.688	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:49:14 AM
Printed by: wet


Ident: CCB
Analysis from: 9/30/2022 11:02:04 AM
File: _2022-09-30_

Last save: 9/30/2022 11:28:05 AM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32043

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:49:26 AM
Printed by: wet


Ident: N4894-03
Analysis from: 9/30/2022 11:30:44 AM
File: _2022-09-30_

Last save: 9/30/2022 11:58:44 AM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32044

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 15
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.95	0.660	2.02	0.13	99.253	0.30
2	9.21	0.303	1471.61	91.62	29992.015	90.19
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	16.50	0.280	59.20	3.69	1168.645	3.51
6	0.00	0.000	0.00	0.00	0.000	0.00
7	26.04	0.406	73.35	4.57	1993.124	5.99
7	28.00	0.236	1606.17	100.00	33253.038	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:49:36 AM

Printed by: wet

Ident: N4894-05

Analysis from: 9/30/2022 12:01:23 PM

File: _2022-09-30_

Last save: 9/30/2022 12:57:03 PM

Method: AnionsIC2-091422.mtw

Last save: 9/28/2022 4:40:

Run operator: wet

Analysis number: 32045

SAMPLE:


: IZ/MA

Vial number: 16

Volume: 20.0 µL

Dilution: 1.00

Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	9.55	0.809	1649.82	87.89	84314.072	93.42
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	26.14	0.391	227.34	12.11	5936.085	6.58
7	28.00	0.171	1877.16	100.00	90250.157	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:49:44 AM
Printed by: wet


Ident: N4894-07
Analysis from: 9/30/2022 12:32:01 PM
File: _2022-09-30_

Last save: 9/30/2022 1:01:09 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32046

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 17
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	9.49	0.790	1651.83	88.06	82534.943	93.41
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.97	0.388	223.98	11.94	5821.302	6.59
7	28.00	0.168	1875.80	100.00	88356.245	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:49:52 AM
Printed by: wet


Ident: N4894-01
Analysis from: 9/30/2022 1:02:38 PM
File: _2022-09-30_

Last save: 9/30/2022 1:30:38 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32047

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 18
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	9.21	0.308	1499.21	86.41	29754.452	83.81
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	16.41	0.277	64.57	3.72	1253.020	3.53
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.97	0.392	171.32	9.87	4495.469	12.66
7	28.00	0.140	1735.10	100.00	35502.942	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:50:14 AM
Printed by: wet


Ident: N4894-01MS
Analysis from: 9/30/2022 1:33:22 PM
File: _2022-09-30_

Last save: 10/3/2022 11:50:10 AM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32048

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 19
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.94	0.492	23.59	1.15	915.673	2.12
2	9.23	0.303	1491.58	72.98	29037.413	67.12
3	11.02	0.220	42.59	2.08	642.670	1.49
4	14.38	0.228	44.54	2.18	692.322	1.60
5	16.38	0.273	120.03	5.87	2310.359	5.34
6	22.58	0.499	80.27	3.93	3357.666	7.76
7	26.05	0.393	241.41	11.81	6308.802	14.58
7	28.00	0.344	2044.00	100.00	43264.906	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:50:37 AM
Printed by: wet


Ident: N4894-01MSD
Analysis from: 9/30/2022 2:04:00 PM
File: _2022-09-30_

Last save: 10/3/2022 11:50:32 AM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32049

Last save: 9/28/2022 4:40:

SAMPLE:
:
Vial number: 20
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.89	0.476	22.51	1.09	859.249	1.98
2	9.17	0.301	1496.02	72.28	28758.175	66.31
3	10.92	0.215	47.26	2.28	691.521	1.59
4	14.24	0.225	47.16	2.28	723.356	1.67
5	16.23	0.270	124.25	6.00	2362.702	5.45
6	22.42	0.493	88.07	4.25	3642.941	8.40
7	25.88	0.390	244.54	11.81	6334.451	14.60
7	28.00	0.339	2069.81	100.00	43372.395	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:50:46 AM
Printed by: wet


Ident: LB122311BLW
Analysis from: 9/30/2022 2:34:41 PM
File: _2022-09-30_

Last save: 10/3/2022 11:48:54 AM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32050

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:50:53 AM
Printed by: wet


Ident: LB122311BSW
Analysis from: 9/30/2022 3:05:20 PM
File: _2022-09-30_

Last save: 10/3/2022 11:48:54 AM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32051

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.89	0.333	18.02	5.48	510.160	6.87
2	9.05	0.173	37.51	11.40	459.943	6.19
3	10.95	0.215	64.97	19.75	1031.437	13.88
4	14.18	0.238	41.93	12.75	681.412	9.17
5	16.27	0.286	48.59	14.77	966.054	13.00
6	22.50	0.473	51.76	15.73	2022.803	27.23
7	26.01	0.403	66.20	20.12	1756.714	23.65
7	28.00	0.303	328.97	100.00	7428.523	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:51:12 AM
Printed by: wet


Ident: CCV
Analysis from: 9/30/2022 4:10:27 PM
File: _2022-09-30_

Last save: 9/30/2022 4:38:27 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32053

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.85	0.313	18.00	5.53	490.453	6.74
2	8.98	0.174	37.75	11.59	462.818	6.36
3	10.87	0.217	63.03	19.36	1022.408	14.05
4	14.07	0.235	42.77	13.14	684.497	9.41
5	16.12	0.282	49.59	15.23	970.334	13.34
6	22.40	0.480	47.30	14.53	1876.251	25.79
7	25.89	0.400	67.14	20.62	1769.028	24.31
7	28.00	0.300	325.58	100.00	7275.787	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:51:19 AM
Printed by: wet


Ident: CCB
Analysis from: 9/30/2022 4:41:23 PM
File: _2022-09-30_

Last save: 9/30/2022 5:09:23 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32054

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:51:27 AM
Printed by: wet


Ident: N4894-01DLX50
Analysis from: 9/30/2022 5:12:09 PM
File: _2022-09-30_

Last save: 9/30/2022 5:40:10 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32055

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 22
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.93	0.171	39.34	93.34	481.433	85.72
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.70	0.429	2.80	6.65	80.223	14.28
7	28.00	0.086	42.15	100.00	561.656	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:51:34 AM
Printed by: wet


Ident: N4894-03DLX50
Analysis from: 9/30/2022 5:42:47 PM
File: _2022-09-30_

Last save: 9/30/2022 6:10:47 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32056

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 23
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.95	0.172	37.58	99.99	455.137	100.00
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	28.00	0.025	37.58	99.99	455.137	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:51:47 AM
Printed by: wet


Ident: N4894-05DLX5
Analysis from: 9/30/2022 6:13:25 PM
File: _2022-09-30_

Last save: 9/30/2022 6:41:24 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32057

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 24
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	9.12	0.269	1451.81	97.46	25057.540	96.08
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.84	0.406	37.89	2.54	1021.675	3.92
7	28.00	0.097	1489.70	100.00	26079.214	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:51:54 AM
Printed by: wet


Ident: N4894-05DLX200
Analysis from: 9/30/2022 6:44:01 PM
File: _2022-09-30_

Last save: 9/30/2022 7:12:01 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32058

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 25
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.97	0.169	39.98	100.00	478.675	100.00
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	28.00	0.024	39.98	100.00	478.675	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:52:05 AM
Printed by: wet


Ident: N4894-07DLX5
Analysis from: 9/30/2022 7:14:39 PM
File: _2022-09-30_

Last save: 9/30/2022 7:42:40 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32059

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 26
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	9.11	0.270	1455.05	97.47	25184.074	96.12
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.73	0.406	37.70	2.53	1016.786	3.88
7	28.00	0.097	1492.74	100.00	26200.860	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:52:12 AM
Printed by: wet


Ident: N4894-07DLX200
Analysis from: 9/30/2022 7:45:17 PM
File: _2022-09-30_

Last save: 9/30/2022 8:13:17 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32060

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 27
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	9.06	0.175	38.24	100.00	474.141	100.00
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	28.00	0.025	38.24	100.00	474.141	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:52:29 AM
Printed by: wet


Ident: CCV
Analysis from: 9/30/2022 8:46:39 PM
File: _2022-09-30_

Last save: 9/30/2022 9:14:39 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32062

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.86	0.321	18.45	5.56	514.451	6.78
2	9.03	0.177	38.08	11.48	472.453	6.23
3	10.95	0.221	62.71	18.90	1030.027	13.58
4	14.21	0.241	42.33	12.76	694.755	9.16
5	16.31	0.291	48.81	14.71	984.132	12.97
6	22.35	0.478	53.27	16.05	2091.144	27.56
7	25.78	0.402	68.21	20.55	1800.142	23.73
7	28.00	0.304	331.86	100.00	7587.103	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/3/2022 11:52:37 AM
Printed by: wet


Ident: CCB
Analysis from: 9/30/2022 9:17:17 PM
File: _2022-09-30_

Last save: 9/30/2022 9:45:17 PM

Method: AnionsIC2-091422.mtw
Run operator: wet
Analysis number: 32063

Last save: 9/28/2022 4:40:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD