

LB123305


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

45310



A= 0.0000e+000

B= 2.2068e-004 slope

C= -3.0908e-002 intercept

Rho= 0.9999618

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.002	-0.002	130	0.000	130					
0.05	0.050					185					
0.20	0.200	0.173	-0.027	925	0.0 %	925	0				-14 14
2.50	2.500	2.494	-0.006	11443	0.0 %	11443					0
5.00	5.000	5.067	0.067	23100	0.0 %	23100					1
7.50	7.500	7.500	0.000	34127	0.0 %	34127					0
10.0	10.000	9.968	-0.032	45310	0.0 %	45310					0

12/11/22

LB123305

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	130	-		0 SFAM01.1	PPB	12/16/2022 11:13	Std
	0.2 S01	925	-		0.2 SFAM01.1	PPB	12/16/2022 11:16	Std
	2.5 S02	11443	-		2.5 SFAM01.1	PPB	12/16/2022 11:18	Std
	5 S03	23100	-		5 SFAM01.1	PPB	12/16/2022 11:20	Std
	7.5 S04	34127	-		7.5 SFAM01.1	PPB	12/16/2022 11:22	Std
	10 S05	45310	-		10 SFAM01.1	PPB	12/16/2022 11:25	Std
ICV058	ICV058	18963	4.1538 -		SFAM01.1	PPB	12/16/2022 11:27	SMPL
ICB058	ICB058	40	-0.0221 -		SFAM01.1	PPB	12/16/2022 11:30	SMPL
CCV082	CCV082	23065	5.059 -		SFAM01.1	PPB	12/16/2022 11:32	SMPL
CCB082	CCB082	-3	-0.0316 -		SFAM01.1	PPB	12/16/2022 11:34	SMPL
PB149703BL	PBS703	120	-0.0044 -		SFAM01.1	PPB	12/16/2022 11:36	SMPL
N6078-01	MCOFD9	2435	0.5064 -		SFAM01.1	PPB	12/16/2022 11:39	SMPL
N6078-02	MCOFE0	2341	0.4857 -		SFAM01.1	PPB	12/16/2022 11:41	SMPL
N6078-03	MCOFE4	3894	0.8284 -		SFAM01.1	PPB	12/16/2022 11:43	SMPL
N6078-04	MCOFE5	6998	1.5134 -		SFAM01.1	PPB	12/16/2022 11:45	SMPL
N6078-05	MCOFE5D	6366	1.3739 -		SFAM01.1	PPB	12/16/2022 11:48	SMPL
N6078-06	MCOFE5S	18786	4.1147 -		SFAM01.1	PPB	12/16/2022 11:53	SMPL
N6078-07	MCOFE6	4256	0.9083 -		SFAM01.1	PPB	12/16/2022 11:55	SMPL
N6078-08	MCOFE7	11141	2.4276 -		SFAM01.1	PPB	12/16/2022 11:57	SMPL
N6078-09	MCOFE8	1811	0.3687 -		SFAM01.1	PPB	12/16/2022 12:00	SMPL
N6078-10	MCOFE9	4586	0.9811 -		SFAM01.1	PPB	12/16/2022 12:02	SMPL
N6078-11	MCOFF0	3542	0.7507 -		SFAM01.1	PPB	12/16/2022 12:04	SMPL
CCV083	CCV083	23139	5.0753 -		SFAM01.1	PPB	12/16/2022 12:06	SMPL
CCB083	CCB083	-9	-0.0329 -		SFAM01.1	PPB	12/16/2022 12:09	SMPL