

Clear table

Instrument ID: IC-2					Analyst :	IZ	Method:	300.0 / 9056A			
ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-11-17_	11/17/22 11:09		IZ/MA
STD2	0.4740	0.6670	0.6970	2.2420	0.5680	1.1120	3.3950	2022-11-17_	11/17/22 11:40		IZ/MA
STD3	0.7820	1.2000	1.2010	4.0470	1.0040	1.9330	6.0320	2022-11-17_	11/17/22 12:10		IZ/MA
STD4	0.9960	1.4670	1.4680	4.8950	1.2220	2.4470	7.2720	2022-11-17_	11/17/22 12:41		IZ/MA
STD5	1.8820	2.9690	2.9160	9.7890	2.4520	5.0120	14.7020	2022-11-17_	11/17/22 13:11		IZ/MA
STD6	3.9100	5.9260	5.9450	19.7890	4.9360	9.9460	29.9000	2022-11-17_	11/17/22 13:42		IZ/MA
STD7	5.1030	7.5700	7.5720	25.2390	6.3170	12.5480	37.6980	2022-11-17_	11/17/22 14:13		IZ/MA
ICV	1.9060	3.0640	3.0810	10.2210	2.5420	4.8410	15.6160	2022-11-17_	11/17/22 15:33		IZ/MA
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-11-17_	11/17/22 16:06		IZ/MA
CCV	1.8050	2.7860	2.7650	9.1800	2.3000	5.4010	13.5650	2022-12-16_	12/16/22 10:06		IZ/MA
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-12-16_	12/16/22 10:40		IZ/MA
LB123336BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-12-16_	12/16/22 11:12		IZ/MA
LB123336BSW	1.8540	2.7830	2.6870	9.1850	2.3140	5.0900	13.6400	2022-12-16_	12/16/22 11:43		IZ/MA
N6087-01	0.0000	73.5140	0.5620	0.0000	0.5330	0.0000	30.1630	2022-12-16_	12/16/22 12:52		IZ/MA
N6087-01MS	2.9990	73.7110	3.4360	10.0560	2.9790	8.9880	44.2520	2022-12-16_	12/16/22 13:24		IZ/MA
N6087-01MSD	2.8080	74.4780	3.3450	10.0930	2.9930	8.9520	44.4340	2022-12-16_	12/16/22 13:55		IZ/MA
CCV	1.8450	2.7950	2.8550	9.2800	2.3200	6.1930	13.5250	2022-12-16_	12/16/22 16:03		IZ/MA
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-12-16_	12/16/22 16:35		IZ/MA

Report date: 12/19/2022 2:26:09 PM
Printed by: wet


Ident: CCV
Analysis from: 12/16/2022 10:06:44 AM
File: _2022-12-16_

Last save: 12/16/2022 10:35:44 AM

Method: AnionsIC2-111722.mtw
Run operator: wet
Analysis number: 32846

Last save: 12/1/2022 4:34:

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.91	0.628	9.92	5.44	451.507	6.93
2	8.83	0.248	22.88	12.54	428.505	6.57
3	10.65	0.317	37.95	20.80	923.288	14.16
4	13.79	0.352	24.05	13.18	625.218	9.59
5	15.75	0.433	27.40	15.02	890.503	13.66
6	21.53	0.729	26.69	14.63	1637.863	25.12
7	25.03	0.543	33.53	18.38	1562.574	23.97
7	29.00	0.464	182.42	100.00	6519.460	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 2:26:22 PM
Printed by: wet


Ident: CCB
Analysis from: 12/16/2022 10:40:31 AM
File: _2022-12-16_

Last save: 12/16/2022 11:09:31 AM

Method: AnionsIC2-111722.mtw
Run operator: wet
Analysis number: 32847

Last save: 12/1/2022 4:34:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 2:26:37 PM
Printed by: wet


Ident: LB123336BLW
Analysis from: 12/16/2022 11:12:09 AM
File: _2022-12-16_

Last save: 12/19/2022 2:25:59 PM

Method: AnionsIC2-111722.mtw
Run operator: wet
Analysis number: 32848

Last save: 12/1/2022 4:34:

SAMPLE:
: IZ/MA
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 2:26:44 PM
Printed by: wet


Ident: LB123336BSW
Analysis from: 12/16/2022 11:43:47 AM
File: _2022-12-16_

Last save: 12/19/2022 2:25:59 PM

Method: AnionsIC2-111722.mtw
Run operator: wet
Analysis number: 32849

Last save: 12/1/2022 4:34:

SAMPLE:
: IZ/MA
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.89	0.609	10.27	5.69	465.376	7.25
2	8.82	0.247	22.92	12.69	427.952	6.67
3	10.64	0.315	36.96	20.47	894.718	13.94
4	13.76	0.352	24.15	13.37	625.562	9.75
5	15.71	0.432	27.61	15.29	895.983	13.96
6	21.52	0.726	25.13	13.92	1536.423	23.94
7	25.02	0.546	33.50	18.55	1571.537	24.49
7	29.00	0.461	180.54	100.00	6417.550	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 2:26:58 PM
Printed by: wet


Ident: N6087-01
Analysis from: 12/16/2022 12:52:26 PM
File: _2022-12-16_

Last save: 12/16/2022 1:50:10 PM

Method: AnionsIC2-111722.mtw
Run operator: wet
Analysis number: 32851

Last save: 12/1/2022 4:34:

SAMPLE:
: IZ/MA
Vial number: 16
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.79	0.274	607.45	87.85	11838.388	75.41
3	10.89	0.863	2.41	0.35	122.618	0.78
4	0.00	0.000	0.00	0.00	0.000	0.00
5	15.72	0.390	5.56	0.80	166.432	1.06
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.01	0.536	75.98	10.99	3570.362	22.74
7	29.00	0.295	691.40	100.00	15697.800	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 2:27:07 PM
 Printed by: wet


Ident: N6087-01MS
 Analysis from: 12/16/2022 1:24:04 PM
 File: _2022-12-16_

Last save: 12/16/2022 1:53:04 PM

Method: AnionsIC2-111722.mtw
 Run operator: wet
 Analysis number: 32852

Last save: 12/1/2022 4:34:

SAMPLE:
 : IZ/MA
 Vial number: 17
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.94	0.690	15.75	1.77	786.396	3.31
2	8.82	0.268	621.51	69.91	11870.172	49.95
3	10.61	0.559	29.11	3.28	1167.032	4.91
4	13.75	0.315	28.76	3.24	688.536	2.90
5	15.62	0.396	38.55	4.34	1168.839	4.92
6	21.62	0.908	40.33	4.54	2810.040	11.82
7	25.06	0.535	114.94	12.93	5274.726	22.19
7	29.00	0.524	888.96	100.00	23765.741	100.00

This report has been created by IC Net
 METROHM LTD

Report date: 12/19/2022 2:27:14 PM
Printed by: wet


Ident: N6087-01MSD
Analysis from: 12/16/2022 1:55:42 PM
File: _2022-12-16_

Last save: 12/16/2022 2:24:42 PM

Method: AnionsIC2-111722.mtw
Run operator: wet
Analysis number: 32853

Last save: 12/1/2022 4:34:

SAMPLE:
: IZ/MA
Vial number: 18
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.89	0.669	14.95	1.66	732.779	3.08
2	8.77	0.265	633.90	70.34	11993.890	50.35
3	10.55	0.548	28.36	3.15	1133.732	4.76
4	13.66	0.313	29.12	3.23	691.223	2.90
5	15.53	0.394	39.06	4.33	1174.371	4.93
6	21.52	0.924	39.84	4.42	2798.423	11.75
7	24.95	0.534	115.93	12.86	5296.648	22.24
7	29.00	0.521	901.17	100.00	23821.066	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 2:27:56 PM
Printed by: wet


Ident: CCV
Analysis from: 12/16/2022 4:03:20 PM
File: _2022-12-16_

Last save: 12/16/2022 4:32:20 PM

Method: AnionsIC2-111722.mtw
Run operator: wet
Analysis number: 32857

Last save: 12/1/2022 4:34:

SAMPLE:
: IZ/MA
Vial number: 22
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.84	0.570	10.65	5.55	462.944	6.77
2	8.77	0.246	23.63	12.31	429.960	6.29
3	10.57	0.323	40.07	20.88	955.999	13.99
4	13.63	0.344	24.90	12.97	632.436	9.25
5	15.57	0.418	28.60	14.90	898.529	13.15
6	21.44	0.754	30.27	15.77	1896.898	27.75
7	24.93	0.539	33.83	17.62	1557.680	22.79
7	29.00	0.456	191.94	100.00	6834.446	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 2:28:04 PM
Printed by: wet


Ident: CCB
Analysis from: 12/16/2022 4:35:01 PM
File: _2022-12-16_

Last save: 12/16/2022 5:04:02 PM

Method: AnionsIC2-111722.mtw
Run operator: wet
Analysis number: 32858

Last save: 12/1/2022 4:34:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD