

LB124463


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49792



A= 0.0000e+000

B= 2.0385e-004 *slope*C= -1.5540e-001 *intercept*

Rho= 0.9997216

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.00	0.000	0.011	0.011	815	0.000	815					
0.05	0.050					185	0				
0.20	0.200	0.202	0.002	1755	0.0 %	1755					
2.50	2.500	2.575	0.075	13394	0.0 %	13394					
5.00	5.000	4.824	-0.176	24426	0.0 %	24426					
7.50	7.500	7.593	0.093	38011	0.0 %	38011					
10.0	10.000	9.995	-0.005	49792	0.0 %	49792					

11
3
-4
1
0

LB124463

INSTRUMENT ID : CV1

Sample ID	Extended II μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
0 S0	815	-	0	SFAM01.1	PPB	3/13/2023 10:51	Std
0.2 S01	1755	-	0.2	SFAM01.1	PPB	3/13/2023 10:54	Std
2.5 S02	13394	-	2.5	SFAM01.1	PPB	3/13/2023 10:56	Std
5 S03	24426	-	5	SFAM01.1	PPB	3/13/2023 10:58	Std
7.5 S04	38011	-	7.5	SFAM01.1	PPB	3/13/2023 11:00	Std
10 S05	49792	-	10	SFAM01.1	PPB	3/13/2023 11:03	Std
ICV081	ICV081	20621	4.0482	-	SFAM01.1	PPB	3/13/2023 11:08 SMPL
ICB081	ICB081	747	-0.0031	-	SFAM01.1	PPB	3/13/2023 11:10 SMPL
CCV052	CCV052	24501	4.8392	-	SFAM01.1	PPB	3/13/2023 11:12 SMPL
CCB052	CCB052	756	-0.0013	-	SFAM01.1	PPB	3/13/2023 11:15 SMPL
PB151380BL	PBS380	792	0.0061	-	SFAM01.1	PPB	3/13/2023 11:17 SMPL
O1378-04	MDL-SOIL	941	0.0364	-	SFAM01.1	PPB	3/13/2023 11:52 SMPL
CCV053	CCV053	24152	4.768	-	SFAM01.1	PPB	3/13/2023 11:57 SMPL
CCB053	CCB053	734	-0.0058	-	SFAM01.1	PPB	3/13/2023 11:59 SMPL