

Clear table

Clear table

Report date: 4/26/2023 4:37:06 PM
Printed by: wet


Ident: CCV
Analysis from: 4/25/2023 9:21:50 AM
File: _2023-04-25_

Last save: 4/25/2023 9:51:51 AM

Method: AnionsIC2-042423.mtw
Run operator: wet
Analysis number: 34018

Last save: 4/24/2023 2:43:

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.55	0.989	6.98	4.21	478.895	7.09
2	9.64	0.319	21.91	13.22	458.348	6.78
3	11.66	0.386	38.81	23.42	981.141	14.52
4	15.13	0.446	23.54	14.20	668.045	9.89
5	17.27	0.544	27.52	16.60	963.998	14.27
6	23.24	1.152	19.71	11.89	1498.595	22.18
7	26.45	1.016	27.28	16.46	1706.472	25.26
7	30.00	0.693	165.74	100.00	6755.494	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/26/2023 4:37:13 PM
Printed by: wet


Ident: CCB
Analysis from: 4/25/2023 9:59:01 AM
File: _2023-04-25_

Last save: 4/25/2023 10:29:01 AM

Method: AnionsIC2-042423.mtw
Run operator: wet
Analysis number: 34019

Last save: 4/24/2023 2:43:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/26/2023 4:37:19 PM
Printed by: wet


Ident: LB125169BLW
Analysis from: 4/25/2023 10:31:38 AM
File: _2023-04-25_

Last save: 4/26/2023 4:36:59 PM

Method: AnionsIC2-042423.mtw
Run operator: wet
Analysis number: 34020

Last save: 4/24/2023 2:43:

SAMPLE:
: IZ/MA
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/26/2023 4:37:25 PM
Printed by: wet


Ident: LB125169BSW
Analysis from: 4/25/2023 11:04:16 AM
File: _2023-04-25_

Last save: 4/26/2023 4:36:59 PM

Method: AnionsIC2-042423.mtw
Run operator: wet
Analysis number: 34021

Last save: 4/24/2023 2:43:

SAMPLE:
: IZ/MA
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.48	0.918	6.98	4.17	447.602	6.67
2	9.52	0.316	22.21	13.27	463.850	6.91
3	11.49	0.383	39.23	23.45	993.984	14.81
4	14.87	0.439	24.08	14.39	671.116	10.00
5	16.94	0.536	28.17	16.84	970.533	14.46
6	23.04	1.148	19.19	11.47	1452.462	21.65
7	26.25	1.019	27.45	16.41	1710.780	25.49
7	30.00	0.680	167.30	100.00	6710.326	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/26/2023 4:37:32 PM
Printed by: wet

Ident: 02213-08
Analysis from: 4/25/2023 11:37:33 AM
File: _2023-04-25_


Last save: 4/25/2023 12:07:33 PM

Method: AnionsIC2-042423.mtw
Run operator: wet
Analysis number: 34022

Last save: 4/24/2023 2:43:

SAMPLE:

: IZ/MA
Vial number: 30
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.41	0.904	1.09	3.69	71.520	6.29
2	9.53	0.309	4.15	14.00	85.161	7.49
3	11.50	0.376	7.33	24.75	184.790	16.24
4	14.87	0.425	4.23	14.28	113.289	9.96
5	16.99	0.526	5.11	17.25	172.706	15.18
6	23.09	1.126	2.38	8.02	177.564	15.61
7	26.35	1.004	5.32	17.96	332.708	29.24
7	30.00	0.667	29.61	99.96	1137.738	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/26/2023 4:37:38 PM
Printed by: wet


Ident: 02213-07
Analysis from: 4/25/2023 12:11:01 PM
File: _2023-04-25_

Last save: 4/25/2023 1:10:24 PM

Method: AnionsIC2-042423.mtw
Run operator: wet
Analysis number: 34023

Last save: 4/24/2023 2:43:

SAMPLE:
: IZ/MA
Vial number: 31
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.45	1.007	0.47	3.43	32.561	6.15
2	9.56	0.311	2.02	14.66	41.854	7.90
3	11.54	0.376	3.42	24.79	85.266	16.10
4	14.92	0.422	1.93	13.99	50.694	9.57
5	17.06	0.533	2.43	17.58	82.953	15.67
6	23.19	1.145	0.89	6.48	68.247	12.89
7	26.47	1.031	2.62	18.96	167.934	31.72
7	30.00	0.689	13.78	99.88	529.509	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/26/2023 4:37:45 PM
Printed by: wet


Ident: O2213-07RE 0.25ML
Analysis from: 4/25/2023 12:43:39 PM
File: _2023-04-25_

Last save: 4/25/2023 1:14:58 PM

Method: AnionsIC2-042423.mtw
Run operator: wet
Analysis number: 34024

Last save: 4/24/2023 2:43:

SAMPLE:
: IZ/MA
Vial number: 32
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.40	1.004	0.27	3.86	17.114	6.34
2	9.48	0.306	0.99	14.12	19.498	7.23
3	11.44	0.375	1.75	25.05	43.651	16.18
4	14.80	0.404	0.90	12.88	22.377	8.30
5	16.94	0.529	1.20	17.15	40.996	15.20
6	23.01	1.138	0.42	6.05	32.081	11.89
7	26.30	1.040	1.45	20.73	94.040	34.86
7	30.00	0.685	6.97	99.83	269.759	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/26/2023 4:37:51 PM
Printed by: wet


Ident: CCV
Analysis from: 4/25/2023 1:16:24 PM
File: _2023-04-25_

Last save: 4/25/2023 1:46:25 PM

Method: AnionsIC2-042423.mtw
Run operator: wet
Analysis number: 34025

Last save: 4/24/2023 2:43:

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.42	0.831	7.42	4.30	444.247	6.54
2	9.51	0.307	22.65	13.11	458.704	6.75
3	11.47	0.374	40.22	23.29	1002.786	14.75
4	14.84	0.425	24.73	14.32	670.264	9.86
5	16.90	0.518	28.91	16.74	967.741	14.24
6	23.10	1.108	20.92	12.11	1534.390	22.58
7	26.33	0.999	27.85	16.13	1718.423	25.28
7	30.00	0.652	172.70	100.00	6796.554	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/26/2023 4:37:58 PM
Printed by: wet


Ident: CCB
Analysis from: 4/25/2023 1:49:33 PM
File: _2023-04-25_

Last save: 4/25/2023 2:19:34 PM

Method: AnionsIC2-042423.mtw
Run operator: wet
Analysis number: 34026

Last save: 4/24/2023 2:43:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD