


LB125266

7471B

INSTRUMENT ID: CV1

 μ Abs.:

64451



Linear

A= 0.0000e+000

B= 1.5611e-004

C= 6.3472e-002

Rho= 0.9997262

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.071	0.071	48	0.000	48					1105
0.2	0.200	0.199	-0.001	870	0.0 %	870					-1
2.5	2.500	2.475	-0.025	15447	0.0 %	15447					-1
5.0	5.000	4.982	-0.018	31508	0.0 %	31508					0
7.5	7.500	7.348	-0.152	46662	0.0 %	46662					-2
10.0	10.000	10.125	0.125	64451	0.0 %	64451					1

LB125266 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 May 2023 14:03:23

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	48	-	0.0000	7471B	PPB	03 May 2023 14:06:21	S	Std
0.2 - 1	50.2	870	-	0.2000	7471B	PPB	03 May 2023 14:08:39	S	Std
2.5 - 1	52.5	15447	-	2.5000	7471B	PPB	03 May 2023 14:10:56	S	Std
5.0 - 1	55	31508	-	5.0000	7471B	PPB	03 May 2023 14:13:14	S	Std
7.5 - 1	57.5	46662	-	7.5000	7471B	PPB	03 May 2023 14:15:31	S	Std
10.0 - 1	510	64451	-	10.0000	7471B	PPB	03 May 2023 14:17:49	S	Std
ICV60 - 1	ICV60	23385	3.7141	-	7471B	PPB	03 May 2023 14:22:56	U	SMPL
ICB60 - 1	ICB60	-108	0.0466	-	7471B	PPB	03 May 2023 14:25:12	U	SMPL
CCV04 - 1	CCV04	31493	4.9798	-	7471B	PPB	03 May 2023 14:27:28	U	SMPL
CCB04 - 1	CCB04	-135	0.0424	-	7471B	PPB	03 May 2023 14:29:44	U	SMPL
CRA - 1	CRA	829	0.1929	-	7471B	PPB	03 May 2023 14:32:02	U	SMPL
HighStd - 1	HighStd	63852	10.0314	-	7471B	PPB	03 May 2023 14:34:17	U	SMPL
ChkStd - 1	ChkStd	43524	6.8580	-	7471B	PPB	03 May 2023 14:36:33	U	SMPL
PB152558BL - 1	PBS	-265	0.0221	-	7471B	PPB	03 May 2023 14:38:51	U	SMPL
PB152558BS - 1	LCSS	22722	3.6106	-	7471B	PPB	03 May 2023 14:41:10	U	SMPL
O2213-02 - 1	LOQ-SOIL-02-QT2-2023	851	0.1963	-	7471B	PPB	03 May 2023 14:53:22	U	SMPL
O2573-01 - 1	VNJ-240	5940	0.9908	-	7471B	PPB	03 May 2023 14:55:40	U	SMPL
O2573-01DUP - 1	VNJ-240DUP	5099	0.8595	-	7471B	PPB	03 May 2023 14:57:56	U	SMPL
O2573-01MS - 1	VNJ-240MS	31065	4.9130	-	7471B	PPB	03 May 2023 15:00:13	U	SMPL
CCV05 - 1	CCV05	31489	4.9792	-	7471B	PPB	03 May 2023 15:02:29	U	SMPL
CCB05 - 1	CCB05	-183	0.0349	-	7471B	PPB	03 May 2023 15:04:47	U	SMPL
O2573-01MSD - 1	VNJ-240MSD	31885	5.0410	-	7471B	PPB	03 May 2023 15:07:04	U	SMPL
O2582-01 - 1	HD-02-050223	2516	0.4562	-	7471B	PPB	03 May 2023 15:09:21	U	SMPL
O2583-01 - 1	CF-21	2701	0.4851	-	7471B	PPB	03 May 2023 15:11:39	U	SMPL
O2583-02 - 1	CF-22	3073	0.5432	-	7471B	PPB	03 May 2023 15:13:55	U	SMPL
O2583-03 - 1	CF-23	3416	0.5967	-	7471B	PPB	03 May 2023 15:16:11	U	SMPL
O2583-04 - 1	CF-24	2752	0.4931	-	7471B	PPB	03 May 2023 15:18:28	U	SMPL
O2583-05 - 1	CF-25	2550	0.4616	-	7471B	PPB	03 May 2023 15:20:44	U	SMPL
O2583-06 - 1	CF-26	2980	0.5287	-	7471B	PPB	03 May 2023 15:23:01	U	SMPL
O2583-07 - 1	CF-27	2736	0.4906	-	7471B	PPB	03 May 2023 15:25:18	U	SMPL
O2584-01 - 1	TR-05-050123	943	0.2107	-	7471B	PPB	03 May 2023 15:27:34	U	SMPL
CCV06 - 1	CCV06	31732	5.0171	-	7471B	PPB	03 May 2023 15:29:50	U	SMPL
CCB06 - 1	CCB06	-129	0.0433	-	7471B	PPB	03 May 2023 15:32:06	U	SMPL
O2585-01 - 1	OK-01-050223	2197	0.4064	-	7471B	PPB	03 May 2023 15:34:23	U	SMPL
O2213-01 - 1	LOD-MDL-SOIL-01-QT2-2023	-58	0.0544	-	7471B	PPB	03 May 2023 15:36:39	U	SMPL
O2573-01LX5 - 1		973	0.2154	-	7471B	PPB	03 May 2023 15:38:55	U	SMPL
O2573-01A - 1		29421	4.6564	-	7471B	PPB	03 May 2023 16:14:21	U	SMPL
CCV07 - 1	CCV07	30079	4.7591	-	7471B	PPB	03 May 2023 16:16:36	U	SMPL
CCB07 - 1	CCB07	-138	0.0419	-	7471B	PPB	03 May 2023 16:18:55	U	SMPL