


LB 126372

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.7289e-004 slope

C= -1.8142e-002 y-intercept

Rho= 0.9999213

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.010	0.010	164	0.000	164					
0.05	0.050					302	0				
0.20	0.200	0.171	-0.029	1092	0.0 %	1092					
2.50	2.500	2.544	0.044	14822	0.0 %	14822					
5.00	5.000	4.938	-0.062	28665	0.0 %	28665					
7.50	7.500	7.570	0.070	43891	0.0 %	43891					
10.0	10.000	9.967	-0.033	57751	0.0 %	57751					

0/00
= 15
2
1
1
0

LB123372

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	164	-		0 SFAM01.1	PPB	7/10/2023 14:34	Std
	0.2 S01	1092	-		0.2 SFAM01.1	PPB	7/10/2023 14:36	Std
	2.5 S02	14822	-		2.5 SFAM01.1	PPB	7/10/2023 14:39	Std
	5 S03	28665	-		5 SFAM01.1	PPB	7/10/2023 14:41	Std
	7.5 S04	43891	-		7.5 SFAM01.1	PPB	7/10/2023 14:48	Std
	10 S05	57751	-		10 SFAM01.1	PPB	7/10/2023 14:50	Std
ICV062	ICV062	21182	3.6441 -		SFAM01.1	PPB	7/10/2023 14:53	SMPL
ICB062	ICB062	-359	-0.0802 -		SFAM01.1	PPB	7/10/2023 14:55	SMPL
CCV008	CCV008	28865	4.9724 -		SFAM01.1	PPB	7/10/2023 15:00	SMPL
CCB008	CCB008	-82	-0.0323 -		SFAM01.1	PPB	7/10/2023 15:02	SMPL
PB154107BL	PBW107	29	-0.0131 -		SFAM01.1	PPB	7/10/2023 15:04	SMPL
O3427-01	A46J0	40	-0.0112 -		SFAM01.1	PPB	7/10/2023 15:06	SMPL
O3427-02	A46T2	32	-0.0126 -		SFAM01.1	PPB	7/10/2023 15:09	SMPL
O3427-03	A46T3	48	-0.0098 -		SFAM01.1	PPB	7/10/2023 15:11	SMPL
O3427-04	A46X9	29	-0.0131 -		SFAM01.1	PPB	7/10/2023 15:13	SMPL
O3427-05	A46R9	63	-0.0072 -		SFAM01.1	PPB	7/10/2023 15:16	SMPL
O3427-06	A46R9D	63	-0.0072 -		SFAM01.1	PPB	7/10/2023 15:18	SMPL
O3427-07	A46R9S	5439	0.9222 -		SFAM01.1	PPB	7/10/2023 15:20	SMPL
PB154108BL	PBW108	103	-0.0003 -		SFAM01.1	PPB	7/10/2023 15:22	SMPL
O3524-01	MH0F96	172	0.0116 -		SFAM01.1	PPB	7/10/2023 15:25	SMPL
O3524-02	MH0F97	195	0.0156 -		SFAM01.1	PPB	7/10/2023 15:27	SMPL
O3524-03	MH0F97D	261	0.027 -		SFAM01.1	PPB	7/10/2023 15:29	SMPL
O3524-04	MH0F97S	5862	0.9954 -		SFAM01.1	PPB	7/10/2023 15:31	SMPL
O3524-05	MH0F98	477	0.0643 -		SFAM01.1	PPB	7/10/2023 15:34	SMPL
O3524-06	MH0F99	299	0.0336 -		SFAM01.1	PPB	7/10/2023 15:36	SMPL
O3524-07	MH0FA0	165	0.0104 -		SFAM01.1	PPB	7/10/2023 15:38	SMPL
CCV009	CCV009	28742	4.9511 -		SFAM01.1	PPB	7/10/2023 15:40	SMPL
CCB009	CCB009	-39	-0.0249 -		SFAM01.1	PPB	7/10/2023 15:43	SMPL